Medical Mask

(Class I sterile)

1、富美企业营业执照



国家企业信用信息公示系统网址: http://www.gsxt.gov.cn

市场主体应当于每年1月1日至6月30日通过国家企业信用信息公示系统报送公示

国家市场监督管理总局监制

2、富美医疗器械生产许可证



0

0

0

9

Ø

a

Ø



消毒产品生产企业卫生许可证

皖卫消证字[2018]第 N0006号

0

0

0

单 位 名 称:安徽富美医疗科技有限公司

法定代表人: 戴文韬

注册地址: 六安市裕安区城南工业园区 (润南路)

生产地址: 六安市裕安区城南工业园区

生产方式:生产

生产项目:消毒器械类

生产类别:用于压力蒸汽灭菌且带有灭菌标识的包装物,用于环氧乙烷灭菌且带有灭菌标识的包装物

注:本许可证只对许可批准时的生产条件负责,不是对企业所生产产品的许可,不代表对企业

生产产品卫生质量的认可。应在卫生许可证有效期届满前30个工作日之前提出延续申请。

发证机关二〇一八年三月七日

有效期限: 2018年03月07日至2022年03月06日







Certificate

No. Q6 096662 0003 Rev. 01

Holder of Certificate: 3A Medical Products Co., Ltd.

Yu An Industrial Park 230001 Liu An

PEOPLE'S REPUBLIC OF CHINA

Facility(ies): 3A Medical Products Co., Ltd.

Yu An Industrial Park, 230001 Liu An, PEOPLE'S REPUBLIC OF

CHINA

Certification Mark:



Scope of Certificate: Production and Distribution of

Surgical Gown, Surgical Drape, Surgical Pack, Coverall, and

Isolation Gown

Applied Standard(s): EN ISO 13485:2016

Medical devices - Quality management systems -

Requirements for regulatory purposes

(ISO 13485:2016) DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system (excluding subclause 7.3), which meets the requirements of the listed standard(s). See also notes overleaf.

Report No.: SH19111104

 Valid from:
 2019-10-17

 Valid until:
 2022-10-16

Date. 2019-08-30

Stefan Preiß

Head of Certification/Notified Body



EC Certificate

Production Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex V (Devices in class I in sterile conditions, sterilised systems or procedure packs)

No. G2S 16 08 96662 002

Manufacturer:

3A Medical Products Co., Ltd.

Yu An Industrial Park 230001 Liu An

PEOPLE'S REPUBLIC OF CHINA



EC-Representative:

Shanghai International Holding

Corp. GmbH (Europe)

Eiffestraße 80 20537 Hamburg GERMANY

Product

Surgical Gown, Surgical Drape,

Category(ies):

Surgical Pack

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for manufacture in accordance with MDD Annex V. This quality assurance system covers those aspects of manufacture concerned with securing and maintaining sterile conditions of the respective devices / device categories and conforms to the requirements of this Directive. It is subject to periodical surveillance. See also notes overleaf.

Report No.:

SH16111101

Valid from:

2016-10-17

Valid until:

2021-10-16

Date, 2016-10-17

1. Paril





TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123

Page 1 of 2

3A Medical Product Co., Ltd

Declaration of Conformity(face mask)

1. Declaration of Conformity

Manufacturer: 3A Medical Products Co., Ltd..

Product name: medical mask

Type:16.5*11 cm

Classification: Class I Sterile

Rule: According to Rule I

We herewith declare that above mentioned products meet the requirement of the

MDD 93/42/EEC Medical device directive and the following harmonized

standards:

ENISO14971:2012

EN14683:2014

ENISO15223-1:2016

ENISO1041:2008A1:2013

ENISO10993-1:2009/AC:2010

ENISO10993-5:2009

ENISO10993-10:2013

Date CE mark was affixed: 2020-04-06

Signature:

Name: JiangJun

Position:General Manager

Date: 2020-04-06

86-564-3611700

Add: YuAn Industrial Park Lu An City, Anhui Province, China 237100

Fax:86-564-3611700





3A Medical Products Co. Ltd, Yu An Industrial Park, 230001 Liu An. PR China

Your notice of 17-03-2020

Your reference

Date 03-04-2020

Analysis Report

Required tests:

EN 14683 (2019) +

AC

(2019)+ AC (2019)

EN 14683 (2019) +

AC

(2019)

EN 14683 (2019) + AC

(2019)

EN 14683 (2019) +

AC (2019) EN 14683 - annex B

(2019)

ISO 22609 (2004)

EN 14683 - annex C

(2019)

+ AC (2019) EN 14683 - §5.2.5 (2019)

AC (2019)

Medical face masks - Splash Test

Bacterial filtration efficiency

Medical face masks -**Breathability (differential**

pressure)

Microbial cleanliness on masks

Identificatio Information given by the client Date of receipt n number T2006059 Cup mask code 17-03-2020 2002 Lot 202001202



Sylvie Niessen

Order

responsible

This report may be reproduced, as long as it is presented in its entire form, without written permission of Centexbel. The results of the analysis cover the received samples. Centexbel is not responsible for the representativeness of the samples. In assessing compliance with the specifications, we did not take into account the uncertainty on the test results.





CENTEXBEL • textile competence centre • www.centexbel.be • www.vkc.be

GENT • Technologiepark 70 • BE-9052 Zwijnaarde, Belgium • phone +32 9 220 41 51 • fax +32 9 220 49 55 • gent@centexbel.be

GRÂCE-HOLLOGNE • Rue du Travail 5 • BE-4460 Grâce-Hollogne, Belgium • phone +32 4 296 82 00 • g-h@centexbel.be

KORTRIJK • Etienne Sabbelaan 49 • BE-8500 Kortrijk, Belgium • phone +32 56 29 27 00 • fax +32 56 29 27 01 • info@vkc.be

VAT BE 0459.218.289 • IBAN BE44 2100 4729 6545 • BIC GEBABEBB





Analysis Report 20.01605.04 Date 03-04-2020 Page 2/9

Reference: T2006059 - Cup mask code

2002 Lot 202001202

Bacterial filtration efficiency

Date of ending the test 25-03-2020

Standard used EN 14683 - annex B (2019) + AC (2019)

Product standard EN 14683 (2019) + AC (2019)

Number of tested masks:

BFE Area tested: ± 49 cm²

Masks conditioning : 21 ± 5 °C and 85 ± 5 % RH

Side of the mask in contact with

the bacterial challenge:

Challenge bacterial strain used : Staphylococcus aureus ATCC6538

Bacterial challenge per test: 1700 - 3000 CFU

Total test time: 1 min. delivering challenge + 1 min. without challenge (air

flow continuing)

Inner side

Flow rate: 28.3 l/min.

Positive control Tests performed with no filter material in the air stream

Negative control Test performed without challenge





Analysis Report 20.01605.04 Date 03-04-2020 Page 3/9

Results

B = Bacterial filtration efficiency (%)

$$B = \frac{(C-T)}{C} X 100$$

With C = mean of the total plate counts for the positive control runs T = total count for the tested mask

# Mask	B (%)
1	> 99.9*
2	> 99.9*
3	> 99.9*
4	> 99.9*
5	> 99.9*

^{*} no detected colonie on any of the Andersen sampler plates

Mean particle size of the bacterial

2.9 µm

challenge aerosol:

Controls

Mean positive controls 2636 CFU Negative control < 1 CFU





Analysis Report 20.01605.04 Date 03-04-2020 Page 4/9

Reference: T2006059 - Cup mask code 2002 Lot 202001202

Medical face masks - Splash Test

Date of ending the test 26-03-2020 Standard used ISO 22609 (2004)

Product standard EN 14683 (2019) + AC (2019)

Number of tested masks: 32

Blood surface tension 42 ± 2 dynes/cm

Volume of the delivered blood 2 ml Distance "canula-mask" 30 ± 1 cm Side of the mask "impacted" Outer side

Masks conditioning : $21 \pm 5^{\circ}$ C and $85 \pm 5\%$ RH

Results

Blood pressure tested 16.0 kPa

Controls

Blood visualisation on the mask OK
Calibration procedure OK
Control of the blood volume delivered (2 ml)

before the test:after 16 masks:after 32 masks:OK







Analysis Report 20.01605.04 Date 03-04-2020 Page 5/9

Results obtained on the set of masks

<u># Mask</u>	<u>Results : pass / fail</u>			
1	Pass			
2	Pass			
3	Pass			
4	Pass			
5	Pass			
6	Pass			
7	Pass			
8	Pass			
9	Pass			
10	Pass			
11	Pass			
12	Pass			
13	Pass			
14	Pass			
15	Pass			
16	Pass			
17	Pass			
18	Pass			
19	Pass			
20	Pass			
21	Pass			
22	Pass			
23	Pass			
24	Pass			
25	Pass			
26	Pass			
27	Pass			
28	Pass			
29	Pass			
30	Pass			
31	Pass			
32	Pass			







Analysis Report 20.01605.04 Date 03-04-2020 Page 6/9

Summary

P = 16.0 kPa

Number of "Pass"	Number of "Fail"
masks	masks
32	0

Pass = no blood detected on the observed side Fail = blood detected on the observed side

In agreement with the customer the number of tested mask has been determined based on a single sampling plan providing an AQL of 4 % (acceptable quality limit).

If 29 masks or more over 32 obtain a "Pass" result the 4% AQL is reached.





Analysis Report 20.01605.04 Date 03-04-2020 Page 7/9

Reference: T2006059 - Cup mask code

2002 Lot 202001202

<u>Medical face masks - Breathability (differential pressure)</u>

Date of ending the test 25-03-2020

Standard used EN 14683 - annex C (2019) + AC (2019)

Product standard EN 14683 (2019) + AC (2019)

Number of tested masks: 5

Number of areas per mask 5 (see figure)


Dimension of the areas : Disc whose diameter is 2.5 cm

Surface areas : 4.9 cm² Flow rate : 8 l/min.

Direction of the air flow: From the inside of the mask to the outside

Masks conditioning : 21 ± 5 °C and 85 ± 5 % RH

Figure: Distribution of the areas in the mask







Analysis Report 20.01605.04 Date 03-04-2020 Page 8/9

Results \triangle **P**

	Mask 1	Mask 2	Mask 3	Mask 4	Mask 5
Area A	122.6	122.0	139.3	123.0	120.4
Area B	126.5	134.0	145.7	131.4	129.0
Area C	168.7	180.3	119.8	172.5	162.6
Area D	131.4	134.5	125.1	153.2	147.3
Area E	144.6	167.5	157.7	156.9	139.1
Average ∆P (Pa/cm²)	138.8	147.7	137.5	147.4	139.7





Analysis Report 20.01605.04 Date 03-04-2020 Page 9/9

Reference: T2006059 - Cup mask code

2002 Lot 202001202

Microbial cleanliness on masks

Date of ending the test 30-03-2020

Standard used EN 14683 - §5.2.5 (2019) AC (2019) Product standard EN 14683 (2019) + AC (2019)

Extraction liquid Peptone 1g/l, NaCl 5g/l & Tween 20 2g/l

Extraction volume 300 ml_e Extraction time 5 min._e

Culture media TSA (Tryptic Soy Agar)

SDA (Sabouraud Dextrose Agar with chloramphenicol)

Incubation conditions 3 days at 30°C (TSA)

7 days at 20-25°C (SDA)

Remark(s): For the fungi count of the mask #1 the result put in the

table corresponds to the reading after 3 days. After 7

days the reading was impossible due to an

overgrowing fungus.

Results_₽

# Mask <i>₄</i>	Mask weight⊬ (g)⊬	CFU*/mask <i></i>		Microbial cleanliness₽		
ţ.	₽ ³	Aerobic microbi al count (bacteria)∂	Fungi count (SDA)∉	Σ CFU/mask <i></i> -	Σ CFU/g⊹	
1₽	6.43₽	53₽	5₽	58₽	10₽	
2₽	6.48₽	114₽	308₽	422₽	66₽	
3₽	6.27₽	61₽	37₽	98₽	16₽	
4.₽	6.37₽	65₽	33₽	98₽	16₽	
5₽	6.48₽	49₽	<3₽	< 52₽	< 9₽	

3A Medical Product Co., Ltd

DECLARATION

To whom this may concern,

We 3A medical hereby confirm that the tested product REF:2001, LOT NO:202001202 correspond to the actual code REF291011 that are producing for Medicalwear, S.A.



Add: YuAn Industrial Park Lu An City, Anhui Province, China 237100
Tel: 86-564-3611700 Fax: 86-564-3611700

8、包装盒





9、实物图片







10. BOX Picture



Size	Qty	Weight
53cm x 38cm x 40 cm	1000 pcs	9.5kg

10. Offerte

	Name	Unit	Qty.	Price (\$)	Reamrk
			10,000,000	1.38\$	FOB
1	Medical Mask	PCS	30,000,000	1.28\$	ShangHai.
			50,000,000	1.18\$	China