

Number: GZHT02287761-S1

Report Ref:	GZHT02287761-S1	THIS IS TO SUPERSEDE REPORT NO.		
		GZHT02287761 DATED May 15, 2020		
Date Received:	May 06, 2020	Date Issued:	May 19, 2020	

Company Name: Address:	HUBEI WANLI PROTECTIVE PRODUCTS CO., LTD YUANSHI, GANHE, XIANTAO		
	CHINA		
Contact Name:	mosa		

The Following Sample Was Submitted And Identified By/On Behalf Of The Applicant As:			
End Uses :	Non-Sterile Medical Face Mask		
Ratings :	Type IIR		
Sample Name :	Face Mask disposable 3-ply		
Size :	17.5X9.5cm		
Colour :	Blue/White		
Standard :	EN 14683:2019+AC:2019		
Agent :	-		
Manufacturer :	HUBEI WANLI PROTECTIVE PRODUCTS CO., LTD		
Brand Name :	XIAN WANLI		
Date received/ Test Started :	May 06, 2020		
Ref :	WLM2002		
	鄂械注准 20182142554		

Prepared And Checked By:

For Intertek Testing Services Shenzhen Ltd. Guangzhou GDD Branch

Lin Lin

General Manager

Page 1 Of 7

Ula / hilaryxu

Intertek Testing Services Stienzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou China 中国广州经济技术开发区科学城彩频路 7 号之二第 1—8 层 02 房 91 房 101


E201、E301、E401、E501、E601、E701、E801

Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663

Hengyun Building, 235 Kaifa Ave., Guangzhou comic & Technological Development District, Guangzhou, China 中国广州经济技开发区开发大道 235 号恒运大厦 3 楼



Number: GZHT02287761-S1



Prepared And Checked By:

For Intertek Testing Services Shenzhen Ltd. Guangzhou GDD Branch

Lin Lin

General Manager

Ula / hilaryxu



Page 2 Of 7

Intertek Testing Services Stienzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E70/E601, No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China 中国广州经济技术开发区科学城彩频路 7 号之二第 1—8 层 02 房 91 房 101.

E201、E301、E401、E501、E601、E701、E801

Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663

Hengyun Building, 235 Kaifa Ave., Guangzhou comic & Technological Development District, Guangzhou, China 中国广州经济技开发区开发大道 235 号恒运大厦 3 楼



Number: GZHT02287761-S1

Summary of testing:

With reference to following standard:

EN 14683:2019+AC:2019 Medical face masks – Requirements and test methods Type IIR

Materials Used in The Submitted Sample Were Found To Comply With The Type IIR Requirements of EN 14683:2019+AC:2019 with respect to Microbial Cleanliness, Bacterial Filtration Efficiency, Differential Pressure and Splash Resistance Pressure tests.

Prepared And Checked By:

For Intertek Testing Services Shenzhen Ltd. Guangzhou GDD Branch

Lin Lin

General Manager

Page 3 Of 7

Ula / hilaryxu

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/F791/E601, No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou 中国广州经济技术开发区科学城彩频路7号之二第1-8层02房、 的馬

東国广州经济技开发区开发大道 235 号恒运大厦 3 楼

Hengyun Building, 235 Kaifa Ave., Guangzhou

E201、E301、E401、E501、E601、E701、E801

Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663

1: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730

pomic & Technological Development District, Guangzhou, China



Number: GZHT02287761-S1

Tests Conducted (As Requested By The Applicant)

1 Microbial Cleanliness

As Per EN 14683:2019+AC:2019 Medical face masks-Requirements and test methods Annex D.

Test Item	Result (cfu/g) (1)					<u>Limit</u>
	Specimen 1	Specimen 2	Specimen 3	Specimen 4	Specimen 5	<u>(cfu/g)</u>
Microbial cleanliness	5	<1#	5	<1#	<1#	Type IIR: ≤30

Remark:

cfu = Colony Forming Unit

 \leq = Not More Than

= No Colony Was Detected At The Extraction Liquid Of The Samples.

Sample Received Condition: Sample In Closed Plastic Bag.

Tested Component: (1) Blue/White Face Mask

Ula / hilaryxu

Page 4 Of 7

Hengyun Building, 235 Kaifa Ave., Guangzhou pomic & Technological Development District, Guangzhou, China 中国广州经济技开发区开发大道 235 号恒运大厦 3 楼



Number: GZHT02287761-S1

Tests Conducted (As Requested By The Applicant)

Bacterial Filtration Efficiency (BFE)
 As Per EN 14683:2019+AC:2019 Medical face masks – Requirements And Test Methods Annex B.

	Results (%)					<u>Performance</u>
						Requirement for
Test Item						<u>Medical Face</u>
	Specimen	Specimen	Specimen	Specimen	Specimen	<u>Mask</u>
	(1)	(2)	(3)	(4)	(5)	(%)
Bacterial						
Filtration	>99.9	99.9	>99.9	>99.9	99.9	Type IIR: ≥98
Efficiency (BFE)						

Remarks:

- 1. Biological Aerosol: Staphylococcus aureus (ATCC 6538).
- 2. Testing side: Outside of the test specimen was facing towards the challenge aerosol.
- 3. Test area: 78 cm²
- 4. Flow rate: 28.3 L/min
- 5. The average plate count results of the positive controls: 2200 CFU
- 6. The average plate count results of the negative controls: < 1 CFU
- 7. CFU = Colony Forming Unit

Ula / hilaryxu

Page 5 Of 7

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/C70/E801, No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China 中国广州经济技术开发区科学城彩频路 7 号之二第 1—8 层 02 房 01 房 101、

E201、E301、E401、E501、E601、E701、E801

Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663

Hengyun Building, 235 Kaifa Ave., Guangzhou Fronomic & Technological Development District, Guangzhou, China

中国广州经济技开发区开发大道 235 号恒运大厦 3 楼



Average

Number: GZHT02287761-S1

Tests Conducted (As Requested By The Applicant)

3 Differential Pressure (EN 14683:2019+AC:2019 Annex C): Air flow: 8L/min, Test area diameter 25 mm, Test area: 4.9 cm².

<u>Tested Sample</u>	Result (Pa/cm²)	Performance Requirement for Medical Face Mask (Pa/cm²)
Specimen (1) Specimen (2)	31.9 38.0	Type IIR: < 60
Specimen (3)	33.5	
Specimen (4)	28.3	
Specimen (5)	35.0	

33.3

Remark: Test was conducted by external provider

Ula / hilaryxu

Page 6 Of 7

Intertek Testing Services Strenzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

E201、E301、E401、E501、E601、E701、E801

Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663

Hengyun Building, 235 Kaifa Ave., Guangzhou Hengyun Building, 235 Kaifa Ave., Guangzhou Hengyun Building, 235 Kaifa Ave., Guangzhou, China 中国广州经济技开发区开发大道 235 号恒运大厦 3 楼



Number: GZHT02287761-S1

Tests Conducted (As Requested By The Applicant)

4 Splash Resistance Pressure (ISO 22609:2004):

Synthetic Blood Surface Tension: 0.040 N/m, Distance Between Blow Head Front End And Target Area: 300 mm, Artificial Blood Volumes: 2 mL, Test Pressure: 16.0 kPa, Velocity: 550 cm/s, Use A Fixed Target.

Tested Sample	<u>Result</u>	Performance Requirement for Medical Face Mask Type IIR: No penetration at 16.0 kPa
Specimen (1)	None seen	
Specimen (2)	None seen	
Specimen (3)	None seen	
Specimen (4)	None seen	
Specimen (5)	None seen	
Specimen (6)	None seen	
Specimen (7)	None seen	
Specimen (8)	None seen	
Specimen (9)	None seen	
Specimen (10)	None seen	
Specimen (11)	None seen	
Specimen (12)	None seen	
Specimen (13)	None seen	

Remark: Test was conducted by external provider.

End of Report

This report is made solely on the basis of your instructions and/or information and materials supplied by you. It is not intended to be a recommendation for any particular course of action. Intertek does not accept a duty of care or any other responsibility to any person other than the Client in respect of this report and only accepts liability to the Client insofar as is expressly contained in the terms and conditions governing Intertek's provision of services to you. Intertek makes no warranties or representations either express or implied with respect to this report save as provided for in those terms and conditions. We have aimed to conduct the Review on a diligent and careful basis and we do not accept any liability to you for any loss arising out of or in connection with this report, in contract, tort, by statute or otherwise, except in the event of our gross negligence or wilful misconduct. No copy of the test report(except for full text copy) shall be made without the written approval by Intertek.

Ula / hilaryxu

Page 7 Of 7

Intertek Testing Services Strenzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/F701/E801, No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China 中国广州经济技术开发区科学城彩频路7号之二第1-8层02房、 101、

E201、E301、E401、E501、E601、E701、E801

Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663

conomic & Technological Development District, Guangzhou, China 中国广州经济技开发区开发大道 235 号恒运大厦 3 楼 s: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730

Hengyun Building, 235 Kaifa Ave., Guangzhou



To: HUBEI WANLI PROTECTIVE PRODUCTS CO., LTD

Attention: mosa Date: May 19, 2020

Re: Report Revision Notification

Labtest Report Number GZHT02287761 date MAY 15, 2020

Please be informed that all the content recorded in the above captioned report will be void. This captioned report is now superseded by a revised Labtest Report, Number GZHT02287761-S1, issued on May 19, 2020.

Thank you for your attention

Prepared And Checked By:

For Intertek Testing Services Shenzhen Ltd. Guangzhou GDD Branch

Lin Lin

General Manager

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F & Room 01, E101/E201/E301/E401/E501/E601/E701/E301, No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China中国广州经济技术开发区科学城彩频路 7 号之二第 1—8 层 02 房 91/房 101、

E201、E301、E401、E501、E601、E701、E801 Tel: +86 20 8213 9001 Fax: +86 20 8208 9909 Postcode: 510663 Hengyun Building, 235 Kaifa Ave., Guangzhou Romomic & Technological Development District, Guangzhou, China 图广州经济技开发区开发大道 235 号恒运大厦 3 楼

:: +86 20 8396 6868 Fax: +86 20 8222 8169 Postcode: 510730

(10)