

- 1. Table of Contents (Page 1)**
- 2. Basic Information (Page 2)**
- 3. Packaging & Pictures (Page 3-8)**
- 4. European Test & Technical Documentation Review Report  
(Page 9-13)**
- 5. Registration on DIMDI & Declaration of Conformity (Page  
14-15)**
- 6. ISO Production Standards Report (Page 16-23)**
- 7. Manufacturer Qualification Certificate & Export License  
(Page 24-28)**

## **Disposable Medical Face Mask Inherent-Type I**

**Brand: Inherent**

**Type No.: YX001**

**Performance standard: EN 14683:2019 + AC:2019(E) Annex B  
/C/D & YY/T 0969-**

**2013, tested by TÜV Rheiland, Intertek & GTTC**

**Production standards: ISO 10933-10:2010 & ISO 10993-  
5:2009, tested by CCIC**

**Classifications: Type I (Non-Sterile), No latex ingredients**

**Material: Two-layer PP Nonwoven fabric and One-  
layer BFE95 Meltblown fabric (3 layers)**

**Mask specifications: Universal, 17,50 x 9,50 cm**

**Earloop design: Breathable and comfortable for prolonged w  
earing**

**Date of manufacture: From July 2020**

**P2**

**Expiration date: 2 years**

**Packing specifications: 10 pcs. / PE bag, 50 pcs. /  
box, 40 boxes 2000 pcs. / carton, 16 cartons / euro pallet**

**Carton measurements: 51.5 x 39 x 35.5 cm, G.W. 9.1 KGS.**

**Stock available in Rotterdam The Netherlands and Frankfurt  
am Main Germany**



**P3**







P6



Prüfbericht-Nr.: Test Report No.:	60373876 001	Auftrags-Nr.: Order No.:	168264746	Seite 1 von 12 Page 1 of 12												
Kunden-Referenz-Nr.: Client Reference No.:	N/A	Auftragsdatum: Order date:	May 13, 2020													
Auftraggeber: Client:	Hunan EEXI Technology & Service Co., Ltd. No.6, North of Pingluo road, Liuyang Hi-tech industrial development zone, Hunan, China															
Prüfgegenstand: Test item:	Disposable Medical Face Mask															
Bezeichnung / Typ-Nr.: Identification / Type No.:	YX001															
Auftragsinhalt: Order content:	Type test															
Prüfungsbasis: Test specification:	EN 14683:2019+AC:2019 except for clause 5.2.6															
Wareneingangdatum: Date of receipt:	May 14, 2020	See Attachment: Photo documentation for details.														
Prüfmuster-Nr.: Test sample No.:	20200504															
Prüfzeitraum: Testing period:	May 14, 2020 to May 28, 2020															
Ort der Prüfung: Place of testing:	See page 3															
Prüflaboratorium: Testing laboratory:	TÜV Rheinland (Shenzhen) Co., Ltd.															
Prüfergebnis: Test result:	Pass															
geprüft von / tested by:		kontrolliert von / reviewed by:														
Javen Ke Javen Ke / Assistant Project Engineer May 28, 2020 Lucy Jiang Lucy Jiang / Assistant Project Engineer May 28, 2020		Angela Chen Angela Chen / Department Manager May 28, 2020														
<table border="1"> <thead> <tr> <th>Datum</th> <th>Name / Stellung</th> <th>Unterschrift</th> <th>Datum</th> <th>Name / Stellung</th> <th>Unterschrift</th> </tr> </thead> <tbody> <tr> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> </tbody> </table>					Datum	Name / Stellung	Unterschrift	Datum	Name / Stellung	Unterschrift						
Datum	Name / Stellung	Unterschrift	Datum	Name / Stellung	Unterschrift											
Sonstiges / Other: - The test report consists of EN 14683 test report including this cover page (12 pages) and attachment: Photo documentation (6 pages). - The Biocompatibility (clause 5.2.6) is not evaluated in this test report.																
Zustand des Prüfgegenstandes bei Anlieferung: Prüfmuster vollständig und unbeschädigt Condition of the test item at delivery: Test item complete and undamaged																
*Legende: 1 = sehr gut 2 = gut 3 = befriedigend 4 = ausreichend 5 = mangelhaft (Pass) = entspricht o.g. Prüfungsregeln (Fail) = entspricht nicht o.g. Prüfungsregeln N/A = nicht anwendbar NT = nicht getestet Legend: 1 = very good 2 = good 3 = satisfactory 4 = sufficient 5 = poor (Pass) = passed a.m. test specification(s) (Fail) = failed a.m. test specification(s) N/A = not applicable NT = not tested																
Dieser Prüfbericht bezieht sich nur auf das o.g. Prüfmuster und darf ohne Genehmigung der Prüfstelle nicht auszugsweise vervielfältigt werden. Dieser Bericht berechtigt nicht zur Verwendung eines Prüfbildes. This test report only relates to the a. m. test sample. Without permission of the test center this test report is not permitted to be duplicated in extracts. This test report does not entitle to carry any test mark.																

EN 14683:2019+AC: 2019 Medical face masks Requirements and test methods	
Report Reference No.:	60373876 001
Date of issue:	See cover page
Total number of pages:	See cover page
Testing Laboratory:	TÜV Rheinland (Shenzhen) Co., Ltd.
Address:	1F East & 2-4F, Cybio Technology Building No.1, No.16 KejiBei 2nd Road, High-Tech Industrial Park North Nanshan District, 518057, Shenzhen, China
Applicant's name:	Hunan EEXI Technology & Service Co., Ltd.
Address:	No.6, North of Pingluo road, Liuyang Hi-tech industrial development zone, Hunan, China
Test specification:	EN 14683:2019+AC:2019
Standard:	Type test
Non-standard test method:	N/A
Test Report Form No.:	EN 14683:2019+AC:2019_A
Test Report Form Originator:	TÜV Rh (SZ)
Master TRF:	2020-03
Test item description:	Disposable Medical Face Mask
Trade Mark:	Inherent
Manufacturer:	Same as the applicant
Model/Type reference:	YX001
Classification:	Type I

List of Attachments (including a total number of pages in each attachment): Attachment – Photo Documentation (6 pages)	
Summary of testing: Tests performed (name of test and test clause): Construction check according to: Clause 5.1.1 Materials and construction Clause 5.1.2 Design	
Testing location: TÜV Rheinland (Shenzhen) Co., Ltd. 1F East & 2-4F, Cybio Technology Building No.1, No.16 KejiBei 2nd Road, High-Tech Industrial Park North Nanshan District, 518057, Shenzhen, China	Sichuan Testing Center of Medical Devices No. 4-28, Xinye Road, High tech west Area, Chengdu, Sichuan, 611731, P.R. China
Clause 5.2.2 Bacterial filtration efficiency (BFE) Clause 5.2.3 Breathability Clause 5.2.5 Microbial cleanliness (Bioburden)	

Copy of marking plate The artwork below may be only a draft. The use of certification marks on a product must be authorized by the respective NCBs that own these marks.																										
Article No.: YX001 Date of preparation of the manual: April 01, 2020 Manual version number: 1.1 Please refer to the instructions before use.																										
<b>Instruction</b> Product name: Disposable medical face mask Type and Specification: Flat earloop, 17.5x9.5cm Type: Type I Applicable standard: EN 14683:2019+AC:2019 YYT 0969-2013 Production license No.: Hunan Food/Drug Administration permission 202000203 Registration No.: Hunan medical device registration permission 20202140297																										
<b>Intended Use:</b> The medical face mask is single-use, disposable device, provided non-sterile, and intended to be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations.																										
<b>Structure and Components:</b> This mask is made by blank mask, nose piece and ear loops. The blank mask is made by three layers. The inner and outer layer are non-woven fabric, the middle layer is melt-blown polypropylene.																										
<b>Main performance:</b> 1) Bacterial filtration efficiency: ≥95% 2) Differential Pressure: < 40 Pa/cm <sup>2</sup> 3) Microbial cleanliness: ≤30 cfu/g																										
<b>Introduction for use:</b> 1) Open the package and remove a mask. 2) Hold the mask by the ear loops, confirm that the colored side is front and the nose piece is on the top of the mask. Hang the ear loops over your ears and mold the nose piece to the shape of your nose, pull the bottom of the mask over your mouth and chin. 3) After using the product, please dispose of it according to the requirements of the environmental protection agency or related authorities. 4) This product is a disposable device. It is not recommended to be clean or reuse it if you feel uncomfortable during use, please stop using it immediately or replace it with a new one.																										
<b>Storage:</b> Store in a dry, well-ventilated, non-corrosive gas place, avoid high temperature. Shelf life: 2 years.																										
<b>Marking of package symbols:</b> <table border="1"> <thead> <tr> <th>Symbol</th> <th>Description</th> <th>Yes</th> <th>No</th> <th>Meaning (check by manufacturer)</th> </tr> </thead> <tbody> <tr> <td>①</td> <td>Means "Do not reuse"</td> <td><input checked="" type="checkbox"/></td> <td><input type="checkbox"/></td> <td>Meaning "Do not reuse"</td> </tr> <tr> <td>②</td> <td>Means "Do not use after expiration date"</td> <td><input checked="" type="checkbox"/></td> <td><input type="checkbox"/></td> <td>Meaning "Do not use after expiration date"</td> </tr> <tr> <td>③</td> <td>Means "Keep away from sunlight"</td> <td><input checked="" type="checkbox"/></td> <td><input type="checkbox"/></td> <td>Meaning "Keep away from sunlight"</td> </tr> <tr> <td>④</td> <td>Means "Keep away from fire"</td> <td><input checked="" type="checkbox"/></td> <td><input type="checkbox"/></td> <td>Meaning "Keep away from fire"</td> </tr> </tbody> </table>		Symbol	Description	Yes	No	Meaning (check by manufacturer)	①	Means "Do not reuse"	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Meaning "Do not reuse"	②	Means "Do not use after expiration date"	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Meaning "Do not use after expiration date"	③	Means "Keep away from sunlight"	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Meaning "Keep away from sunlight"	④	Means "Keep away from fire"	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Meaning "Keep away from fire"
Symbol	Description	Yes	No	Meaning (check by manufacturer)																						
①	Means "Do not reuse"	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Meaning "Do not reuse"																						
②	Means "Do not use after expiration date"	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Meaning "Do not use after expiration date"																						
③	Means "Keep away from sunlight"	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Meaning "Keep away from sunlight"																						
④	Means "Keep away from fire"	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Meaning "Keep away from fire"																						
Batch code: refer to package Date of manufacture: refer to package Use by date: refer to package Authorized representative in the European Community: Shanghai International Holding Corp. GmbH (Europe) Address: Cellerstrasse 80, 20537 Hamburg, Germany Manufacturer: Hunan EEXI Technology & Service Co., Ltd. Address: No. 6 North of Pingluo Road, Liuyang Hi-tech Industrial Development Zone, Hunan, China. Hotline: +86-4000-333-088																										
See attachment for other information.																										



**TÜVRheinland®** Page 5 of 12 Report No. 60373876 001

**Testing**

Date of receipt of test item(s).....: See cover page

Dates of tests performed.....: See cover page

**Possible test case verdicts:**

- test case does not apply to the test object.....: N/A
- test object does meet the requirement.....: P (Pass)
- test object was not evaluated for the requirement.....: N/E (collateral standards only)
- test object does not meet the requirement.....: F (Fail)

**General remarks:**

"(See Attachment #)" refers to additional information appended to the report.  
 "(See appended table)" refers to a table appended to the report.  
 The tests results presented in this report relate only to the object tested.  
 This report shall not be reproduced except in full without the written approval of the testing laboratory.  
 List of test equipment must be kept on file and available for review.  
 Additional test data and/or information provided in the attachments to this report.

Throughout this report a  comma /  point is used as the decimal separator.

Name and address of factory (ies).....: Same as the applicant

**General product information:**

1. The tested medical mask classified as Type I.
2. The Biocompatibility (clause 5.2.6) is not evaluated in this test report.
3. The test results are for reference only. Relevant certification may be needed if the mask is intended to be sold in Europe.

QM/RT-3300SHG Revision number: 1.0 Effective date: 2020-03-12

**TÜVRheinland®** Page 6 of 12 Report No. 60373876 001

EN 14683:2019+AC:2019

Clause	Requirement + Test	Result - Remark	Verdict
<b>4</b>	<b>Classification</b>		<b>P</b>
	Medical face masks specified in this European Standard are classified into two types (Type I and Type II) according to bacterial filtration efficiency whereby Type II is further divided according to whether or not the mask is splash resistant. The 'R' signifies splash resistance.	Type I	<b>P</b>
<b>5</b>	<b>Requirements</b>		<b>P</b>
<b>5.1</b>	<b>General</b>		<b>P</b>
<b>5.1.1</b>	<b>Materials and construction</b>		<b>P</b>
	In the medical face mask is a medical device, generally composed of a filter layer that is placed, bonded or moulded between layers of fabric.	The Disposable Medical Face Masks are made of blank mask, nose clip and ear loops. The outer and inner layers of the mask are made of non woven fabrics, and the middle layer is made of melt-blown polypropylene.	<b>P</b>
	The medical face mask shall not disintegrate, split or tear during intended use.		<b>P</b>
	In the selection of the filter and layer materials, attention shall be paid to cleanliness.		<b>P</b>
<b>5.1.2</b>	<b>Design</b>		<b>P</b>
	The medical face mask shall have a means by which it can be fitted closely over the nose, mouth and chin of the wearer and which ensures that the mask fits closely at the sides.	With nose clip.	<b>P</b>
	Medical face masks may have different shapes and constructions as well as additional features such as a face shield (to protect the wearer against splashes and droplets) with or without anti-fog function, or a nose bridge (to enhance fit by conforming to the nose contours).		<b>P</b>
<b>5.2</b>	<b>Performance requirements</b>		<b>P</b>
<b>5.2.1</b>	<b>General</b>		<b>P</b>
	All tests shall be carried out on finished products or samples out from finished products.		<b>P</b>
<b>5.2.2</b>	<b>Bacterial filtration efficiency (BFE)</b>		<b>P</b>
	When tested in accordance with Annex B, the BFE of the medical face mask shall conform to the minimum value given for the relevant type in Table 1.	See appended table 5.2.2	<b>P</b>

QM/RT-3300SHG Revision number: 1.0 Effective date: 2020-03-12

**TÜVRheinland®** Page 7 of 12 Report No. 60373876 001

EN 14683:2019+AC:2019

Clause	Requirement + Test	Result - Remark	Verdict
	For thick and rigid masks such as rigid duckbill or cup masks the test method may not be suitable as a proper seal cannot be maintained in the cascade impactor. In these cases, another valid equivalent method shall be used to determine the BFE.	Not such mask.	<b>N/A</b>
	When a mask consists of two or more areas with different characteristics or different layer-composition, each panel or area shall be tested individually.	Same characteristics and same layer-composition declared by manufacturer.	<b>N/A</b>
	The lowest performing panel or area shall determine the BFE value of the complete mask.	See above	<b>N/A</b>
<b>5.2.3</b>	<b>Breathability</b>		<b>P</b>
	When tested in accordance with Annex C, the differential pressure of the medical face mask shall conform to the value given for the relevant type in Table 1.	See appended table 5.2.3	<b>P</b>
	If the use of a respiratory protective device as face mask is required in an operating theatre and/or other medical settings, it might not fulfil the performance requirements with regard to differential pressure as defined in this European Standard. In such case, the device should fulfil the requirement as specified in the relevant Personal Protective Equipment (PPE) standards).		<b>N/A</b>
<b>5.2.4</b>	<b>Splash resistance</b>		<b>N/A</b>
	When tested in accordance with ISO 22609:2004 the resistance of the medical face mask to penetration of splashes of liquid shall conform to the minimum value given for Type I/R in Table 1.	See appended table 5.2.4 Type I mask.	<b>N/A</b>
<b>5.2.5</b>	<b>Microbial cleanliness (Bioburden)</b>		<b>P</b>
	When tested according to EN ISO 11737-1:2018 the bioburden of the medical mask shall be < 30 CFU/g tested (see Table 1).	See appended table 5.2.5	<b>P</b>
<b>5.2.6</b>	<b>Biocompatibility</b>		<b>N/E</b>
	According to the definition and classification in EN ISO 10993-1:2009, a medical face mask is a surface device with limited contact.	The biocompatibility is not evaluated in this test report.	<b>N/E</b>
	The manufacturer shall complete the evaluation of the medical face mask according to EN ISO 10993-1:2009 and determine the applicable toxicology testing regime.		<b>N/E</b>
	The results of testing should be documented according to the applicable parts of the EN ISO 10993 series.		<b>N/E</b>
	The test results shall be available upon request.		<b>N/E</b>
<b>6</b>	<b>Marking, labelling and packaging</b>		<b>P</b>

QM/RT-3300SHG Revision number: 1.0 Effective date: 2020-03-12

**TÜVRheinland®** Page 8 of 12 Report No. 60373876 001

EN 14683:2019+AC:2019

Clause	Requirement + Test	Result - Remark	Verdict
	Annex I, §13, of the Medical Devices Directive (93/42/EEC) or Annex I, §23, of the Medical Device Regulation (EU) 2017/745 specifies the information that should be specified on the packaging in which the medical face mask is supplied.	See page 4 and attachment.	<b>P</b>
	The following information shall be supplied:		<b>P</b>
	a) number of this European Standard,		<b>P</b>
	b) type of mask (as indicated in Table 1).		<b>P</b>
	EN ISO 15223-1:2016 and EN 1041:2008+A1:2013 should be considered.		<b>P</b>

QM/RT-3300SHG Revision number: 1.0 Effective date: 2020-03-12

P9

EN 14683:2019+AC:2019								
Clause	Requirement + Test		Result - Remark			Verdict		
5.2.2	TABLE: Bacterial filtration efficiency (BFE)							
Batch/lot no.:	Test Specimen no.:	Dimension of the test specimen L x W (mm x mm)	test area (cm²)	Flow rate (l/min)	Mean of the total plate counts of the two positive controls	Mean plate count of the negative controls	BFE for each test specimen (%)	Remarks
20200504	1	161 x 152	63.6	28.3	1825	0	99.78%	--
	2	160 x 150	63.6	28.3			99.77%	--
	3	161 x 150	63.6	28.3			99.48%	--
	4	162 x 150	63.6	28.3			99.67%	--
	5	161 x 151	63.6	28.3			99.62%	--
Supplementary information:								
1. Each specimen was conditioned at 21 °C and 85 % relative humidity for 16h to bring them into equilibrium with atmosphere prior to testing.								
2. The side of the test specimen was facing towards the challenge aerosol: <u>the inside of the test specimen.</u>								

EN 14683:2019+AC:2019					
Clause	Requirement + Test		Result - Remark		Verdict
5.2.3	TABLE: Breathability (Differential pressure)				
Batch/lot no.:	Test Specimen number-Test area number	Differential pressure for each test area (Pa/cm²)	The averaged differential pressure for each test specimen (Pa/cm²)	Flow rate (l/min)	Remarks
20200504	1-1	15.5	22.5	8.0	--
	1-2	22.4		8.0	--
	1-3	26.9		8.0	--
	1-4	23.9		8.0	--
	1-5	23.7		8.0	--
	2-1	16.3	23.6	8.0	--
	2-2	23.4		8.0	--
	2-3	28.3		8.0	--
	2-4	23.7		8.0	--
	2-5	26.3		8.0	--
	3-1	13.7	21.0	8.0	--
	3-2	23.7		8.0	--
	3-3	24.7		8.0	--
	3-4	21.7		8.0	--
	3-5	21.4		8.0	--
	4-1	14.6	22.1	8.0	--
	4-2	23.1		8.0	--
	4-3	24.1		8.0	--
	4-4	25.7		8.0	--
	4-5	22.9		8.0	--
	5-1	20.0	22.4	8.0	--
	5-2	23.4		8.0	--
	5-3	21.3		8.0	--
	5-4	24.6		8.0	--
	5-5	22.8		8.0	--
Supplementary information:					
Each specimen was conditioned at 21 °C and 85 % relative humidity for 16h to bring them into equilibrium with atmosphere prior to testing.					

EN 14683:2019+AC:2019					
Clause	Requirement + Test		Result - Remark		Verdict
atmosphere prior to testing.					
5.2.4	TABLE: Splash resistance				N/A
Batch/lot no.:	Test mask no.:	The material of tested mask	Test result (Pass/fail)	Remarks	
20200504	1				
	2				
	3				
	4				
	5				
	6				
	7				
	8				
	9				
	10				
	11				
	12				
	13				
	14				
	15				
	16				
	17				
	18				
	19				
	20				
	21				
	22				
	23				
	24				
	25				
	26				
	27				
	28				

EN 14683:2019+AC:2019					
Clause	Requirement + Test		Result - Remark		Verdict
29					
Supplementary information:					
1. Each specimen was conditioned at ___ °C and ___ % relative humidity for ___ h to bring them into equilibrium with atmosphere prior to testing.					
2. The description of target area tested: _____					
3. Any technique used to enhance visual detection of synthetic blood: _____					
4. The temperature and relative humidity for testing: ___ °C and ___ %.					
5. Description of any pre-treatment techniques used: _____					
5.2.5	TABLE: Microbial cleanliness (Bioburden)				P
Batch/lot no.:	Mask/under test) no.:	Weight of each mask (g)	Total bioburden per individual mask (CFU/g)	Remarks	
20200504	1	3.0	<1	--	
	2	2.9	<1	--	
	3	2.9	<1	--	
	4	2.9	<1	--	
	5	3.0	<1	--	
Supplementary information:					

End of EN 14683 test report

P10

Product: Disposable medical face mask  
 Type Designation: YX001



Figure 1 Top view of Packaging box



Figure 2 Front / Back view of Packaging box

Product: Disposable medical face mask  
 Type Designation: YX001




Figure 4 Side view of packaging box




Figure 5 Bottom view of packaging box

Product: Disposable medical face mask  
 Type Designation: YX001



Figure 3 Side view of packaging box

Product: Disposable medical face mask  
 Type Designation: YX001



Figure 6 General view of packaging bag




Figure 7 View of medical mask




Figure 8 View of medical mask



Figure 9 View of medical mask (3 ply)



Figure 10 View of nose clip  
 END OF THE PHOTO DOCUMENTATION

Intertek  
 Total Quality Assured  
**TEST REPORT**

中国合格评定  
 国家认可  
 实验室  
 TESTING  
 CNAS L0220  
 Number: GZHT02293014

<b>Report Ref:</b>	GZHT02293014
<b>Date Received:</b>	May 18, 2020/May 19, 2020
<b>Date Issued:</b>	Jun 01, 2020
<b>Company Name:</b>	HUNAN EEXI TECHNOLOGY & SERVICE CO LTD
<b>Address:</b>	NO.6, NORTH OF PINGTOU ROAD LEIJIANG HI-TECH INDUSTRIAL DEVELOPMENT ZONE, HUNAN CHINA
<b>Contact Name:</b>	XU XIAO XIAO
The Following Sample Was Submitted And Identified By/On Behalf Of The Applicant As:	
End Uses	: Non-Sterile Medical Face Mask
Ratings	: Type I
Sample Name	: Disposable Medical Face Mask
Size	: 17.5*9.5cm
Colour	: Blue/White
Standard	: EN 14683:2019+AC:2019
Manufacturer	: HUNAN EEXI TECHNOLOGY & SERVICE CO.,LTD
Date received/ Test Started	: May 18, 2020/May 19, 2020
Ref	: Band Name: Inherent Type No.: YX001

Testing was conducted on specific items, at our client's request.

Approved by:

Sr. Manager



Assistant Supervisor




Page 1 Of 6

wen / abbyqzeng

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch  
 深圳天祥质量技术服务有限公司广州分公司  
 Room 02, 1&8F & Room 01, E101E201E301E401E501E601E701E801E901  
 No.72, Caishan Road, Guangzhou Science City, GUETD, Guangzhou, China  
 中国广东省广州市天河区科学城彩虹桥 72 号之二 1-8 楼 02 室, 01 室  
 001  
 E201, E301, E401, E501, E701, E801  
 Tel: +86 20 8215 9001 Fax: +86 20 8208 9609 Postcode: 510663

Intertek  
 Total Quality Assured  
**TEST REPORT**

中国合格评定  
 国家认可  
 实验室  
 TESTING  
 CNAS L0220  
 Number: GZHT02293014



Approved by:

Sr. Manager



Assistant Supervisor



Page 2 Of 6

wen / abbyqzeng

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch  
 深圳天祥质量技术服务有限公司广州分公司  
 Room 02, 1&8F & Room 01, E101E201E301E401E501E601E701E801E901  
 No.72, Caishan Road, Guangzhou Science City, GUETD, Guangzhou, China  
 中国广东省广州市天河区科学城彩虹桥 72 号之二 1-8 楼 02 室, 01 室  
 001  
 E201, E301, E401, E501, E701, E801  
 Tel: +86 20 8215 9001 Fax: +86 20 8208 9609 Postcode: 510663

Number: GZHT02293014

Summary of testing:

With reference to following standard:  
 EN 14683:2019+AC:2019 Medical face masks – Requirements and test methods Type I  
 Materials Used In The Submitted Sample Were Found To Comply With The Type I Requirements of EN 14683:2019+AC:2019 with respect to Microbial Cleanliness, Bacterial Filtration Efficiency and Differential Pressure tests.

Approved by:

*[Signature]*  
 Sr. Manager Assistant Supervisor

Page 3 Of 6

wen / abbyqiang

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch  
 深圳天祥质量技术服务有限公司广州分公司

Room 02, 1&F & Room 01, E11/E20/E30/E40/E50/E60/E70/E80/E90  
 No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China  
 中国广州科学城采品路7号之二第一-八层02室-01层  
 101  
 E201, E301, E401, E501, E701, E801, E901  
 Tel: +86 20 8213 9001 Fax: +86 20 8206 9999 Postcode: 510663

3/F, Hengyun Building, 235 Kaili Ave., Guangzhou  
 Economic & Technological Development District, Guangzhou, China  
 中国广州经济技术开发区云汇大厦235号恒运大厦3楼  
 Tel: +86 20 8396 6668 Fax: +86 20 8222 2169 Postcode: 510730

Number: GZHT02293014

Tests Conducted (As Requested By The Applicant)

1 Microbial Cleanliness

As Per EN 14683:2019+AC:2019 Medical face masks Requirements and test methods Annex D.

Test Item	Result (cfu/g)					Limit (cfu/g)
	Specimen (1)	Specimen (2)	Specimen (3)	Specimen (4)	Specimen (5)	
Microbial cleanliness	1	<1#	<1#	<1#	<1#	Type I: ≤30

Remark:  
 cfu = Colony Forming Unit  
 S = Not More Than  
 # = No Colony Was Detected At The Extraction Liquid Of The Samples.

Sample Received Condition: Sample In Closed Plastic Bag.

Remark: This Test Item Was Conducted In Caipin Road, Guangzhou Science City, GETDD, Guangzhou, Guangdong.

wen / abbyqiang

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch  
 深圳天祥质量技术服务有限公司广州分公司

Room 02, 1&F & Room 01, E11/E20/E30/E40/E50/E60/E70/E80/E90  
 No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China  
 中国广州科学城采品路7号之二第一-八层02室-01层  
 101  
 E201, E301, E401, E501, E701, E801, E901  
 Tel: +86 20 8213 9001 Fax: +86 20 8206 9999 Postcode: 510663

3/F, Hengyun Building, 235 Kaili Ave., Guangzhou  
 Economic & Technological Development District, Guangzhou, China  
 中国广州经济技术开发区云汇大厦235号恒运大厦3楼  
 Tel: +86 20 8396 6668 Fax: +86 20 8222 2169 Postcode: 510730

Page 4 Of 6

Tests Conducted (As Requested By The Applicant)

2 Bacterial Filtration Efficiency (BFE)

As Per EN 14683:2019+AC:2019 Medical face masks – Requirements And Test Methods Annex B.

Test Item	Results (%)					Performance Requirement for Medical Face Masks (%)
	Specimen (1)	Specimen (2)	Specimen (3)	Specimen (4)	Specimen (5)	
Bacterial Filtration Efficiency (BFE)	99.5	99.6	99.8	99.7	99.6	Type I: ≥95

Remarks:  
 1. Biological Aerosol: *Staphylococcus aureus* (ATCC 6538).  
 2. Testing side: Outside of the test specimen was facing towards the challenge aerosol.  
 3. Test area: 78 cm<sup>2</sup>  
 4. Flow rate: 28.3 L/min  
 5. The average plate count results of the positive controls: 2200 CFU  
 6. The average plate count results of the negative controls: < 1 CFU  
 7. CFU = Colony Forming Unit.

Remark: This Item Is Not Under The Testing Scope Of CNAS Accreditation.  
 This test item was conducted in Caipin Road, Guangzhou Science City, GETDD, Guangzhou, Guangdong.

wen / abbyqiang

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch  
 深圳天祥质量技术服务有限公司广州分公司

Room 02, 1&F & Room 01, E11/E20/E30/E40/E50/E60/E70/E80/E90  
 No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China  
 中国广州科学城采品路7号之二第一-八层02室-01层  
 101  
 E201, E301, E401, E501, E701, E801, E901  
 Tel: +86 20 8213 9001 Fax: +86 20 8206 9999 Postcode: 510663

3/F, Hengyun Building, 235 Kaili Ave., Guangzhou  
 Economic & Technological Development District, Guangzhou, China  
 中国广州经济技术开发区云汇大厦235号恒运大厦3楼  
 Tel: +86 20 8396 6668 Fax: +86 20 8222 2169 Postcode: 510730

Page 5 Of 6

Tests Conducted (As Requested By The Applicant)

3 Differential Pressure (EN 14683:2019+AC:2019 Annex C):

Air flow: 8 L/min, Test area diameter 25 mm, Test area: 4.9 cm<sup>2</sup>.

Tested Sample	Result (Pa/cm <sup>2</sup> )	Performance Requirement for Medical Face Mask (Pa/cm <sup>2</sup> )
Specimen (1)	32.7	Type I: < 40
Specimen (2)	32.8	
Specimen (3)	29.6	
Specimen (4)	33.6	
Specimen (5)	39.8	
Average	33.7	

Remark : 1. Test was conducted by external provider.  
 2. This item is not under the testing scope of CNAS accreditation.

End of Report

*This report is made solely on the basis of your instructions and/or information and materials supplied by you. It is not intended to be a recommendation for any particular course of action. Intertek does not accept a duty of care or any other responsibility to any person other than the client in respect of this report and only accepts liability to the client insofar as is expressly contained in the terms and conditions governing Intertek's provision of services to you. Intertek makes no warranties or representations either express or implied with respect to this report save as provided for in these terms and conditions. We have aimed to conduct the Review on a diligent and careful basis and we do not accept any liability to you for any loss arising out of or in connection with this report, in contract, tort, by statute or otherwise, except in the event of our gross negligence or willful misconduct. No copy of the test report (except for full test copy) shall be made without the written approval by Intertek.*

wen / abbyqiang

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch  
 深圳天祥质量技术服务有限公司广州分公司

Room 02, 1&F & Room 01, E11/E20/E30/E40/E50/E60/E70/E80/E90  
 No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, China  
 中国广州科学城采品路7号之二第一-八层02室-01层  
 101  
 E201, E301, E401, E501, E701, E801, E901  
 Tel: +86 20 8213 9001 Fax: +86 20 8206 9999 Postcode: 510663

3/F, Hengyun Building, 235 Kaili Ave., Guangzhou  
 Economic & Technological Development District, Guangzhou, China  
 中国广州经济技术开发区云汇大厦235号恒运大厦3楼  
 Tel: +86 20 8396 6668 Fax: +86 20 8222 2169 Postcode: 510730

Page 6 Of 6

Medizinprodukte - Informationssystem

**Suche** | **Suchergebnisse** | **Dokumentvergabe** | **Profilseite (13)**

**Aktuelle Medizinprodukte (MPA)**

**1. Wert & Einstell. MP Anzeigen (MPA) 3 | 121803**

Dokumentnummer: 95414390

**Anzeige**

Registrierdatum: 2020-04-09  
 Registrierungsnummer: DE/CA/05/MP-238321-2556-00  
 Typ der Anzeige: Störanzeige  
 Anzeiger nach § 25 MPG: Bevollmächtigter  
 Ort: Hamburg  
 Postleitzahl: 20537  
 Straße, Haus-Nr.: Effestraße 80  
 Land: Hamburg  
 Telefon: +49-40-2513175  
 Telefax: +49-40-255726  
 E-Mail: shholding@hotmail.com

**Zuständige Behörde**

Code: 002/CA/05  
 Bezeichnung: Behörde für Gesundheit und Verbraucherschutz  
 Zustufe: Referat V43  
 Staat: Deutschland  
 Land: Hamburg  
 Ort: Hamburg  
 Postleitzahl: 20539  
 Straße, Haus-Nr.: Effestraße 80  
 Telefon: +49-40-428290  
 Telefax: +49-40-427310017  
 E-Mail: medizinprodukte@bfgv.hamburg.de  
 Bearbeiter: Frau Seline Henning  
 Bearbeiter Telefon: 040 52822 2130

**Hersteller**

Bezeichnung: Human EXXI TechnologyService Co., Ltd.  
 Staat: China  
 Ort: Lianyungang  
 Postleitzahl: 410323  
 Straße, Haus-Nr.: No.6, North of Pingtuo road, Luyang Hi-tech Industrial Development Zone,  
 Telefon: +86-731-82371666  
 E-Mail: oversea01@idore.com.cn

**Medizinprodukt**

Produktort: Nichtaktives Medizinprodukt  
 Klasse: I  
 May Software auf mobilen Endgeräten: Ja  
 Nominaleinsache: EXXI  
 Allgemeine Produktbezeichnung: Disposable medical face mask  
 Nominaleinsache: 21642  
 Herstellerbezeichnung: Hiesia  
 Kategorie: EP  
 Produkte zum Berührungsbereich: Produkte zum Berührungsbereich  
 Kurzbeschreibung in Englisch: Disposable face mask made up with PP non-woven, meltblown non-woven, etc. Applied for ordinary medical health care and personal protection.

**Medizinprodukte (Medizinprodukte)**

**Suche** | **Suchergebnisse** | **Dokumentvergabe** | **Profilseite (13)**

**Service links**

- Widerrufen
- Informationen für Anzeiger von Stromführenden Profunoren

Anlage 1  
 (zu § 4 Abs. 1 Nr. 1 DRGMV)  
 Formulanummer 9203/1588

**Anzeiger / Reporting organisation (person)**

Code: DE/000040627  
 Bezeichnung / Name: Shanghai International Holding Corporation GmbH (Europe)  
 Staat / State: China | Land / Federal state: Deutschland | Hamburg  
 Ort / City: Hamburg | Postleitzahl / Postal code: 20537  
 Straße, Haus-Nr. / Street, house no.: Effestraße 80  
 Telefon / Phone: +49-40-2513175 | Telefax / Fax: +49-40-255726  
 E-Mail / E-mail: shholding@hotmail.com

**Hersteller / Manufacturer**

Bezeichnung / Name: Human EXXI Technology&Service Co., Ltd.  
 Staat / State: CN  
 Ort / City: Lianyungang | Postleitzahl / Postal code: 410323  
 Straße, Haus-Nr. / Street, house no.: No.6, North of Pingtuo road, Luyang Hi-tech Industrial Development Zone,  
 Telefon / Phone: +86-731-82371666 | Telefax / Fax:  
 E-Mail / E-mail: oversea01@idore.com.cn

**Sicherheitsbeauftragter für Medizinprodukte nach § 30 Abs. 2 MPG 9)  
 Safety officer for medical devices pursuant to § 30 (2) Medical Devices Act, MPG**

Bezeichnung / Name: Liang Jin  
 Staat / State: Deutschland | Land / Federal state: Hamburg  
 Ort / City: Hamburg | Postleitzahl / Postal code: 20537  
 Straße, Haus-Nr. / Street, house no.: Effestraße 80  
 Telefon / Phone: +49-40-2513175 | Telefax / Fax: +49-40-255726  
 E-Mail / E-mail: shholding@hotmail.com

- 2 -

**Allgemeine Anzeigepflicht nach §§ 25 und 30 Abs. 2 MPG  
 General Obligation to Notify pursuant to §§ 25 and 30 (2) Medical Devices Act, MPG**

**Formblatt für Medizinprodukte, außer In-vitro-Diagnostika  
 Form for Medical Devices except In Vitro Diagnostic Medical Devices**

Anlage 1  
 (zu § 4 Abs. 1 Nr. 1 DRGMV)  
 Formulanummer 9203/1588

**Zuständige Behörde / Competent authority**

Code: DE/CA/05  
 Bezeichnung / Name: Behörde für Gesundheit und Verbraucherschutz, Referat V43  
 Staat / State: Deutschland | Land / Federal state: Hamburg  
 Ort / City: Hamburg | Postleitzahl / Postal code: 20539  
 Straße, Haus-Nr. / Street, house no.: Billstraße 80  
 Telefon / Phone: +49-40-428290 | Telefax / Fax: +49-40-427310017  
 E-Mail / E-mail: medizinprodukte@bfgv.hamburg.de

**Anzeige / Notification**

Registrierdatum bei der zuständigen Behörde  
 Registration date at competent authority: DE/CA/05/MP-238321-2556-00  
 09.04.2020

Typ der Anzeige / Notification type  
 S Erstanzeige / Initial notification  
 E Änderungsanzeige / Notification of change  
 E Widerrufsanzeige / Notification of withdrawal

Frühere Registrierungsnummer bei Änderungs- und Widerrufsanzeige  
 Previous registration number if notification has been changed or withdrawn

Anzeigener nach § 25 MPG / Reporter pursuant to § 25 Medical Devices Act, MPG  
 E Hersteller / Manufacturer  
 S Bevollmächtigter / Authorised Representative  
 E Einführer / Importer  
 E Verantwortlicher für das Zusammensetzen von Systemen oder Behandlungseinheiten nach § 10 Abs. 1 und 2 MPG / Assembler of systems or procedure packs pursuant to § 10 (1) and (2) Medical Devices Act, MPG  
 E Betrieb oder Einrichtung (aufbereiten) nach § 25 Abs. 1 MPG i. V. m. § 4 Abs. 2 MPBetrieb/  
 Institution (processing) pursuant to § 25 (1) Medical Devices Act, MPG in connection with § 4 (2) MPBetrieb/  
 E Betrieb oder Einrichtung (sterilisieren) nach § 25 Abs. 2 i. V. m. § 10 Abs. 3 MPG  
 Institution (sterilizing) pursuant to § 25 (2) in connection with § 10 (3) Medical Devices Act, MPG

- 1 -

Anlage 1  
 (zu § 4 Abs. 1 Nr. 1 DRGMV)  
 Formulanummer 9203/1588

**Vertreter / Deputy (optional)**

Bezeichnung / Name:  
 Telefon / Phone:  
 Telefax / Fax:  
 E-Mail / E-mail:  
 E Erstanzeige / Initial notification  
 S Änderungsanzeige / Notification of change

- 3 -

P14

<b>Medizinprodukt (Erstmaliges Inverkehrbringen) / Medical device (First placing on the market)</b>	
Klasse / Class S I	
E I - steril / sterile	
E I - mit Messfunktion / with measuring function	
E I - steril und mit Messfunktion / sterile and with measuring function	
E IIa	
E IIb	
E III	
E III - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012	
E Aktives implantierbares Medizinprodukt / Active implantable medical device	
E Aktives implantierbares Medizinprodukt - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 Active implantable medical device - manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012	
App (Software auf mobilen Endgeräten)	E ja / yes S nein / no
Nummer(n) der Bescheinigung(en) / Certificate number(s)	
Handelsname des Produktes / Trade name of the device <b>EEXI</b>	
Produktbezeichnung / Name of device <b>Disposable medical face mask</b>	
Nomenklaturcode / Nomenclature code <b>12-447</b>	
Nomenklaturbezeichnung / Nomenclature term <b>Maske</b>	
Kategoriecode / Category code <b>10</b>	
Kategorie / Category <b>Produkte zum Einmalgebrauch</b>	
Kurzbeschreibung deutsch / German short description	
Kurzbeschreibung englisch / English short description <b>Disposable face mask made up with PP non-woven, meltblown non-woven, etc. Applied for ordinary medical health care and personal protection.</b>	

<b>Medizinprodukte (Aufbereiten) / Medical devices (Reprocessing)</b>	
E Semikritische Medizinprodukte / Semicritical medical devices E Gruppe A / Group A E Gruppe B / Group B	
E Kritische Medizinprodukte / Critical medical devices E Gruppe A / Group A E Gruppe B / Group B E Gruppe C / Group C Nummer der Bescheinigung / Certificate number	
Sterilisationsverfahren / Sterilisation procedures E Dampfsterilisation / Steam sterilisation E Gassterilisation / Gas sterilisation E Strahlsterilisation / Radiation sterilisation E andere / others Angewandtes Verfahren / Applied procedure	

Ich versichere, dass die Angaben nach bestem Wissen und Gewissen gemacht wurden.  
 I affirm that the information given above is correct to the best of my knowledge.

Ort City	<b>Hamburg</b>	Datum Date	<b>2020-04-07</b>
Name		<b>Liang Jin</b>	
		Unterschrift Signature	

<b>Bearbeitungsvermerke / Processing notes</b> Nur von der zuständigen Behörde auszufüllen / To be filled in only by the competent authority	
Bearbeiter / Person responsible <b>Frau Sylvia Frenzel</b>	Telefon / Phone <b>040 42837-2120</b>





# Skin Sensitization Test Guinea Pig Maximization Final Report

Article Name: Disposable medical face mask  
Report Number: CSTBB20030201  
Method Standard: ISO 10993-10: 2010

Sponsor	Test Facility
Human EEXI Technology&Service Co.,Ltd. No.6 north of Pingtoun road, Linyang Hi-tech industrial development zone, Hunan, China	CCIC Huatongwei international inspection (Suzhou) Co., Ltd Room 101, Building G, Raoshui Road 388, Suzhou, Jiangsu, China

CCIC Huatongwei international inspection (Suzhou) Co., Ltd  
Address: Room 101, Building G, Raoshui Road 388, Suzhou, Jiangsu, China, 512123 Tel: 0512-87657288 Fax: 0512-87657288  
Page 1 of 11

Report No.: CSTBB20030201

## Notices

1. Please apply for rechecking within 15 days of receiving the report if there is any objection.
2. Any erasure or without special testing seal renders the report null and void.
3. The report is only valid when signed by the persons who edited, checked and approved it.
4. The report is only responsible for the test results of the tested samples.
5. The report shall not be reproduced except in full without the written approval of the company.

Page 3 of 11

Report No.: CSTBB20030201

## CONTENTS

Notices.....	3
Abstract.....	4
Study Verification and Signature.....	5
1.0 Purpose.....	6
2.0 Reference.....	6
3.0 Test and control articles.....	6
4.0 Identification of test system.....	7
5.0 Animal Management.....	7
6.0 Equipment and reagents.....	7
7.0 Experiment design.....	8
8.0 The results observed.....	9
9.0 Evaluation criteria.....	9
10.0 Results of the test.....	9
11.0 Conclusion.....	9
12.0 Record.....	9
13.0 Confidentiality Agreement.....	9

Page 2 of 11

Report No.: CSTBB20030201

## Abstract

In this study, we took guinea pigs to observe the skin sensitization of the test article according to ISO 10993-10: 2010.

The test article were extracted in Constant Temperature Vibrator at 50 °C, 60 rpm for 72 h by 0.9 % Sodium Chloride Injection and Sesame Oil. Mix 50:50 (by volume) stable emulsion of Freund's complete adjuvant with selected solvent. Intradermal induction and topical induction were operated in the clipped intrascapular region of each animal. After the topical induction phase was completed on day 14, all test and control animals were challenged with the test sample. The erythema and edema of the challenge site were observed to test the sensitization response of the test article. According to the Magnusson and Klignan scales, the response to erythema and edema at each application site of the skin was described and scored 24 hours and 48 hours after the challenge phase.

The results showed that the guinea pigs in the negative control group (0.9 % Sodium Chloride Injection, Sesame Oil) retained a normal appearance throughout the test and showed no skin irritants. A severe skin reactions for erythema and oedema were shown in the positive control group (DNCB). While in test article group, the response of skin on testing side did not exceed that on the control side. The skin reactions for erythema and oedema were not observed in test article group. The data of each group met the acceptance criteria, and the results of this test were considered valid.

Based on the above results, it can be concluded that under the experimental conditions, the test article Disposable medical face mask has no potential skin sensitization on guinea pigs in the extraction method.

Page 4 of 11



Study Verification and Signature



Protocol Number SST2003006603BB  
 Protocol Effective Date 2020-03-13  
 Technical Initiation Date 2020-03-20  
 Technical Completion Date 2020-04-17  
 Final Report Completion Date 2020-04-18

Personnel: *[Signature]* Date Completed 2020-04-18  
 Approved: *[Signature]* Date Completed 2020-04-18  
 Study Director  
 Supervisory: *[Signature]*  
 Test Facility Manager  
 Huatongwei International Inspection (Seal)

1.0 Purpose

The test was designed to evaluate the potential of a test article to cause skin sensitization. The test is used as a procedure for screening of contact allergens in guinea pigs and extrapolating the results to humans, but it does not establish the actual risk of sensitization.

2.0 Reference

- Biological evaluation of medical devices Part 10: Tests for irritation and skin sensitization (ISO 10993-10:2010)
- Biological evaluation of medical devices-Part 12: Sample preparation and reference materials (ISO 10993-12:2012)
- Biological evaluation of medical devices-Part 2: Animal welfare requirements (ISO 10993-2:2006)

3.0 Test and control articles

Groups	Test article	Negative Control Article(Polar)	Negative Control Article(Non-Polar)	Positive Control
Name	Disposable medical face mask	Sodium Chloride Injection (SC)	Sesame Oil (SO)	2, 4-Dinitrochlorobenzene (DNCB)
Manufacture	Hunan EEXI Technology&Service Co.,Ltd.	Shijiazhuang No.4 Pharmaceutical	Jiangxi xinsen natural vegetable oil co., Ltd.	TOKYO CHEMICAL INDUSTRY CO., LTD
Sterilization state	No	/	/	/
Size	17.5cm*9.5cm	500 ml	25 kg	25 g
Model	YX001	/	/	/
Lot Batch#	Not provided	1912121907	181120	H2UKD-DM
Test Article Material	PP non-woven, Meltblown filtration layer, earloop and nose clip	/	/	/
Physical State	Solid	Liquid	Liquid	Solid
Color	Not provided	Colorless	Light yellow	Light yellow
Package material	PE poly bag and paper box	/	/	/
Concentration	/	0.9 %	/	Induction Concentration: 0.5 % Challenge Concentration: 1.0 % Dissolved in ethanol
Total	Not provided	/	/	/

Surface/Weight	Room Temp.	Room Temp.	Room Temp.	Room Temp.
Storage Condition				

The information about the test article was supplied by the sponsor wherever applicable.

4.0 Identification of test system

4.1 Test animal

Species: Hartley Guinea Pig (Cavia Porcellus)  
 Number: 30 (20 Test +10 Control)  
 Sex: 15 ♂, 15 ♀  
 Initial body weight: 301.5-313.5 g  
 Health status: Healthy, not previously used in other experimental procedures  
 Animal identification: Ear tag  
 Cages: Plastic cage  
 Acclimation Period: 7 days under the same conditions as for the actual test

4.2 Justification of test system

The albino guinea pig has been used historically for sensitization studies (Magnusson and Klignan, 1970). The guinea pig is believed to be the most sensitive animal model for this type of study. DNCB is the positive control article recommended in the test instructions. To ensure the sensitivity of the experimental system, the positive control article should be verified every three months.

5.0 Animal Management

Animal purchase: Wuxi hengtai experimental animal breeding co. LTD SCXK (SU) 2015-0004  
 Bedding: Corncob Jiangsu Xietong Pharmaceutical Bio-engineering Co., Ltd.  
 Feed: Guinea pigs were fed with full-price pellets Jiangsu Xietong Pharmaceutical Bio-engineering Co., Ltd.  
 Water: Drinking water met the Standards for Drinking Water Quality GB 5749-2006  
 Animal room temperature: 18-26 °C  
 Animal room relative humidity: 30 %-70 %  
 Lights: 12 hours light/dark cycle, full-spectrum lighting  
 Personnel: Associates involved were appropriately qualified and trained  
 Selection: Only healthy, previously unused animals were selected  
 There were no known contaminants present in the feed, water, or bedding expected to interfere with the test data.

6.0 Equipment and reagents

6.1 Instruments

Constant Temperature Vibrator (SHB007, calibration data: 2020/3/16), Autoclave (SHB026, calibration data: 2020/3/16), Electronic scale (SHB017, calibration data: 2020/3/16)

6.2 Reagents

Freund's adjuvant Complete liquid (SIGMA, Lot No: SLBR3877V), Sodium dodecyl sulfate (SDS SIGMA, Lot No: SLBL2304V)

7.0 Experiment design

7.1 Sample preparation

The extracts of test article will be prepared according to the following steps:

Sampling Manner	Aseptic Sampling		Extraction in sterile vessels				
	Actually sampling	Ratio	Reagent	Temperature	Time	pH	
Whole	570.2 cm <sup>2</sup>	6cm <sup>3</sup> : 1 ml	SC	95.0 ml	50 °C	72 h	5.5
	570.2 cm <sup>2</sup>		SO	95.0 ml			5.5

Both inducements and excitations were prepared by the number of times. The state of the leaching solution did not change visually after the leaching was advanced. After extraction, the samples were stored at room temperature for no more than 24 h. The extraction solution is clear, and the pH value has not been adjusted, filtered, centrifuged, diluted and other processes. The control solution was prepared under the same conditions.

7.2 Test method

7.2.1 Intradermal induction phase I

A pair of 0.1ml intradermal injections was made for each of the following, into each animal, at the injection sites (A, B and C) as shown in Figure 1 in the clipped intrascapular region.

- Site A: A 50:50 (volume ratio) stable emulsion of Freund's complete adjuvant mixed with the chosen solvent.
- Site B: The test sample (undiluted extract); the control animals were injected with the solvent alone.
- Site C: The test sample at the concentration used at site B, emulsified in a 50:50 volume ratio stable emulsion of Freund's complete adjuvant and the solvent (50 %); the control animals were injected with an emulsion of the blank liquid with adjuvant.




Figure 1 Location of intradermal injection sites

7.2.2 Topical induction phase II

The maximum concentration that can be achieved in Intradermal induction phase I did not produce irritation, animals are pretreated with 10% sodium dodecyl sulfate 24(±2) hours before the topical induction application.

At 7 d after completion of the intradermal induction phase, administer test article extract by topical application to the intrascapular region of each animal, using a patch of area approximately 8 cm<sup>2</sup> (absorbent gauze), so as to cover the intradermal injection sites. Secure the patches with an occlusive dressing. Remove the dressings and patches after (48±2) h.

Treat the control animals similarly, using the blank liquid alone.

7.2.3 Challenge phase

At 14d after completion of the topical induction phase, challenge all test and control animals with the test sample. Absorbent gauzes (2.5 cmx2.5 cm) were soaked respectively with test article and control article. Apply the test article extract and control article topically to two sites that were not treated during the induction stage. Secure with an occlusive dressing. Remove the dressings and patches after (24±2) h.

**8.0 The results observed**

The day after challenge exposure, the patch will be removed and the area cleaned gently with gauze if necessary. The site will be wiped gently with a 0.9 % saline soaked gauze sponge prior to each scoring period. The challenge sites will be observed for signs of irritation and sensitization reactions, as indicated by erythema and edema. If necessary, the fur will be shaved or clipped in advance for the convenience of dermal score.

Daily challenge observation scores will be recorded approximately 24, and 48 hours after patch removal in accordance with the following classification system for skin reactions:

**Table 1 Magnusson and Klignan scale**

Patch test reaction	Grading scale
No visible change	0
Discrete or patchy erythema	1
Moderate and confluent erythema	2
Intense erythema and/or swelling	3

**9.0 Evaluation criteria**

Magnusson and Klignan grades of 1 or greater in the test group generally indicate sensitization, provided grades of less than 1 are seen in control animals.

If grades of 1 or greater are noted in control animals, then the reactions of test animals which exceed the most severe reaction in control animals are presumed to be due to sensitization.

If the response is equivocal, rechallenge is recommended to confirm the results from the first challenge.

The outcome of the test is presented as the frequency of positive challenge results in test and control animals.

**10.0 Results of the test**

All animals were survived and no abnormal signs were observed during the study. Individual results of dermal scoring for the challenge appear in Table 2.

**11.0 Conclusion**

The test article showed no evidence of causing delayed dermal contact sensitization in the guinea pig. Results and conclusions apply only to the test article tested. Any extrapolation of these data to other articles is the sponsor's responsibility.

**12.0 Record**

All raw data pertaining to this study and a copy of the final report are retained in designated Huatongwei archive.

**13.0 Confidentiality Agreement**

Statements of confidentiality were as agreed upon prior to study initiation.

**Table 2 Guinea pig Sensitization Dermal Reactions**

Group	No.	Pretest weight(g)	Finished weight(g)	The Challenge patch was removed 24h later		The Challenge patch was removed 48h later		Positive rate	
				Erythema	Swelling	Erythema	Swelling		
SC	Test	1	312.8	350.1	0	0	0	0	0%
		2	310.8	349.8	0	0	0	0	
		3	311.7	346.8	0	0	0	0	
		4	311.8	346.5	0	0	0	0	
		5	313.5	350.2	0	0	0	0	
	Control	6	308.2	349.3	0	0	0	0	
		7	310.3	352.1	0	0	0	0	
		8	312.2	350.6	0	0	0	0	
		9	307.9	346.8	0	0	0	0	
		10	311.9	347.9	0	0	0	0	
		11	306.8	346.4	0	0	0	0	
		12	311.6	351.3	0	0	0	0	
		13	307.1	347.9	0	0	0	0	
		14	311.5	356.1	0	0	0	0	
		15	305.9	349.3	0	0	0	0	
SO	Test	16	301.5	346.6	0	0	0	0	0%
		17	311.9	351.6	0	0	0	0	
		18	307.3	348.6	0	0	0	0	
		19	306.0	344.4	0	0	0	0	
		20	302.2	340.5	0	0	0	0	
	Control	21	307.3	348.1	0	0	0	0	
		22	301.9	344.9	0	0	0	0	
		23	306.8	346.1	0	0	0	0	
		24	312.0	356.1	0	0	0	0	
		25	307.3	348.3	0	0	0	0	
		26	312.3	355.7	0	0	0	0	
		27	311.6	354.5	0	0	0	0	
		28	309.8	354.0	0	0	0	0	
		29	306.3	347.9	0	0	0	0	
		30	307.4	345.9	0	0	0	0	

**Table 3 Positive control**

Group	No.	Pretest weight(g)	Finished weight(g)	The Challenge patch was removed 24 h later		The Challenge patch was removed 48 h		Positive rate
				Erythema	Swelling	Erythema	Swelling	
Test	1	311.2	351.7	2	0	2	0	100%
	2	315.6	360.2	1	0	1	0	
	3	317.2	352.7	0	0	1	0	
	4	312.8	353.9	1	0	1	0	
	5	306.9	341.1	2	0	2	0	
	6	312.2	350.6	0	0	1	0	
	7	317.1	352.2	1	0	2	0	
	8	306.8	350.0	1	0	1	0	
	9	316.5	348.7	1	0	2	0	
	10	317.6	350.2	2	0	2	0	
Control	11	320.5	364.5	0	0	0	0	-
	12	307.6	350.2	0	0	0	0	
	13	306.9	345.1	0	0	0	0	
	14	310.0	352.3	0	0	0	0	
	15	315.8	346.6	0	0	0	0	

Note: The positive control was CSTBB20010001P1(Finish date: 2020-02-07).



**Skin Irritation Test**  
**Extraction Method**  
**Final Report**

Article Name: Disposable medical face mask

Report Number: CSTBB20030203

Method Standard: ISO 10993-10: 2010

**Sponsor**

Human EEXI Technology&Service Co.,Ltd.  
No.6 north of Pingtan road, Luyang Hi-tech industrial development zone, Hunan, China

**Test Facility**

CCIC Huatongwei international inspection (Suzhou) Co., Ltd.  
Room 101, Building G, Kuashui Road 388, Suzhou, Jiangsu, China

CONTENTS

Notices.....3
Abstract.....4
Study Verification and Signature.....5
1.0 Purpose.....6
2.0 Reference.....6
3.0 Test and control articles.....6
4.0 Identification of test system.....6
5.0 Animal Management.....7
6.0 Equipment and reagents.....7
7.0 Experiment design.....7
8.0 The results observed.....8
9.0 Evaluation criteria.....9
10.0 Results of the test.....9
11.0 Conclusion.....9
12.0 Record.....9
13.0 Confidentiality Agreement.....9

Notices

- 1. Please apply for rechecking within 15 days of receiving the report if there is any objection.
2. Any erasure or without special testing seal renders the report null and void.
3. The report is only valid when signed by the persons who edited, checked and approved it.
4. The report is only responsible for the test results of the tested samples.
5. The report shall not be reproduced except in full without the written approval of the company.



Abstract

In this study, we took New Zealand white Rabbits to observe the skin irritation of the test article according to ISO10993-10:2010.

The test article were extracted in Constant Temperature Vibrator at 50 °C, 60 rpm for 72 h by 0.9 % Sodium Chloride Injection and Sesame Oil. Apply 0.5 ml extracts of test article or control to 2.5 cm×2.5 cm absorbent gauze patches, and then apply the patch soaked with the extract of test article or control directly to the skin on each side of each rabbit, and then wrap the application sites with a bandage for a minimum of 4 h. At the end of the contact time, remove the dressing. The describe and score the skin reaction for erythema and oedema for each application site at each time interval. Record the appearance of each application site at (1±0.1) h, (24±2) h, (48±2) h and (72±2) h following removal of the patches.

The results showed that the rabbits in the negative control group (0.9 % Sodium Chloride Injection, Sesame Oil) retained a normal appearance throughout the test and showed no skin irritants. A severe skin reactions for erythema and oedema were shown in the positive control group (SDS). While in test article group, the response of skin on testing side did not exceed that on the control side. The skin reactions for erythema and oedema were not observed in test article group. The data of each group met the acceptance criteria, and the results of this test were considered valid.

Based on the above results, it can be concluded that under the experimental conditions, the test article Disposable medical face mask has no potential skin irritation on rabbit in the extraction method.

Study Verification and Signature



Protocol Number SST2003006601BB
Protocol Effective Date 2020-03-13
Technical Initiation Date 2020-03-20
Technical Completion Date 2020-03-27
Final Report Completion Date 2020-04-18

Personnel [Signature] 2020-03-14 Date Completed

Approved [Signature] 2020-03-19 Date Completed
Study Director

Supervisory [Signature]
Text Identify Manager



P19

**1.0 Purpose**

To evaluate the potential skin irritation caused by test article contact with the skin surface of rabbits and extrapolating the results to humans, but it does not establish the actual risk of irritation.

**2.0 Reference**

Biological evaluation of medical devices Part 10: Tests for irritation and skin sensitization (ISO 10993-10: 2010)

Biological evaluation of medical devices-Part 12: Sample preparation and reference materials (ISO 10993-12:2012)

Biological evaluation of medical devices-Part 2: Animal welfare requirements (ISO 10993-2:2006)

**3.0 Test and control articles**

Groups	Test article	Negative Control Article(Polar)	Negative Control Article(Non-Polar)	Positive Control
Name	Disposable medical face mask	Sodium Chloride Injection (SC)	Sesame Oil (SO)	10 % sodium dodecyl sulfate (SDS)
Manufacture	Hunan EEXI Technology&Service Co.,Ltd.	Shijiazhuang No.4 Pharmaceutical	Jiangxi xinsun natural vegetable oil co., Ltd.	SIGMA
Sterilization state	No	/	/	/
Size	17.5cm*9.5cm	500 ml	25 kg	25 g
Model	YX001	/	/	/
Lot Batch#	Not provided	1912121907	181120	SLBL2304V
Test Article Material	PP non-woven, Melblown filtration layer, earloop and nose clip	/	/	/
Physical State	Solid	Liquid	Liquid	Solid
Color	Not provided	Colorless	Light yellow	Colorless
Package material	PE poly bag and paper box	/	/	/
Concentration	/	0.9 %	/	10 %
Total Surface/Weight	Not provided	/	/	/
Storage Condition	Room Temp.	Room Temp.	Room Temp.	Room Temp.

The information about the test article was supplied by the sponsor wherever applicable.

**4.0 Identification of test system**

**4.1 Test animal**

Species: New Zealand white Rabbit

Number: 6

Sex: 3 ♀, 3 ♂

Weight: 2.04-2.17 kg

Health status: Healthy, not previously used in other experimental procedures

Animal identification: Ear tattoo

Cages: Stainless steel cage

Acclimation Period: 7 days under the same conditions as for the actual test

**4.2 Justification of test system**

The rabbit is specified as an appropriate animal model for evaluating potential skin irritants by the current testing standards. Positive control 10% sodium dodecyl sulfate has been substantiated at ITW with this method.

**5.0 Animal Management**

Animal purchase: Wuxi hengtai experimental animal breeding co. LTD SCXK (SU) 2015-0904

Bodding: /

Feeds: Experimental rabbits were fed a maintenance diet, Wuxi hengtai experimental animal breeding co. LTD

Water: Drinking water met the Standards for Drinking Water Quality GB 5749-2006

Animal room temperature: 18-26 °C

Animal room relative humidity: 30 %-70 %

Light: 12 hours light/dark cycle, full-spectrum lighting

Personnel: Associates involved were appropriately qualified and trained

Selection: Only healthy, previously unused animals were selected

There were no known contaminants present in the feed, water, or bedding expected to interfere with the test data

**6.0 Equipment and reagents**

**6.1 Instruments**

Constant Temperature Vibrator (SHB007, calibration data: 2020/3/16), Autoclave (SHB026, calibration date: 2020/3/16), Electronic scale (SHB017, calibration data: 2020/3/16)

**7.0 Experiment design**

**7.1 Sample preparation**

The extracts of test article will be prepared according to the following steps:


Sampling Manner	Aseptic Sampling		Extraction in sterile vessels				
	Actually sampling	Ratio	Reagent	Temperature	Time	pH	
Whole	570.2 cm <sup>3</sup>	6cm <sup>3</sup> : 1 ml	SC	95.0 ml	50 °C	72 h	5.5
	570.2 cm <sup>3</sup>		SO	95.0 ml			

The state of the leaching solution did not change visually after the leaching was advanced. The extractions were clear, and the pH value has not been adjusted, filtered, centrifuged, diluted and other processes, before dosing stored at room temperature no more than 24 h. The control solution was prepared under the same conditions

**7.2 Test method**

Use the rabbits with healthy intact skin. Fur was generally clipped within 24 h period before testing on the backs of the rabbits, a sufficient distance on both sides of the spine for application and observation of all test sites (approximately 10\*15 cm).

Apply 0.5 ml extract (s) of test article or control to 2.5 cm\*2.5 cm absorbent gauze patches, and then apply the patch soaked with the extract of test article or control directly to the skin on each side of each rabbit as shown in Figure 1, and then wrap the application sites with a bandage (semi-occlusive or occlusive) for a minimum of 4h. At the end of the contact time, remove the dressing.



1- Cranial end, 2- Test site, 3- Control site, 4- Clipped dorsal region, 5- Caudal end

**Figure1 Location of skin application sites**

**8.0 The results observed**

The Describe and score the skin reaction for erythema and oedema according to the scoring system given in Table 1 for each application site at each time interval. Record the appearance of each application site at (1±0.1) h, (24±2) h, (48±2) h and (72±2) h following removal of the patches.

**Table 1 Classification System for Skin Reaction**

Erythema and Eschar Formation:	Numerical Grading
No erythema	0
Very slight erythema (barely perceptible)	1
Well-defined erythema	2
Moderate erythema	3
Severe erythema (beet redness) to eschar formation preventing grading of erythema	4
Edema Formation:	
No edema	0
Very slight edema (barely perceptible)	1
Well-defined edema (edges of area well-defined by definite raising)	2
Moderate edema (raised approximately 1mm)	3
Severe edema (raised more than 1mm and extending beyond exposure area)	4
Maximal possible score for irritation	8

**Irritation Response Categories in the Rabbit**

Response Category	Mean score
Negligible	0 to 0.4
Slight	0.5 to 1.9
Moderate	2 to 4.9
Severe	5 to 8

**9.0 Evaluation criteria**

Use only (24±2) h, (48±2) h and (72±2) h observations for calculation.

After the 72 h grading, all erythema grades plus oedema grades (24±2) h, (48±2) h and (72±2) h were totaled separately for each test article and blank for each animal. The primary irritation score for an animal was calculated by dividing the sum of all the scores by 6 (two test/observation sites, three time points).

To obtain the primary irritation index for the test article, add all the primary irritation scores of the individual animals and divide by the number of animals.

When blank or negative control was used, calculate the primary irritation score for the controls and subtract that score from the score using the test material to obtain the primary irritation score.

**10.0 Results of the test**

All animals were survived and no abnormal signs were observed during the study. According to what observed, the response of skin on testing side did not exceed that on the control side. Thus, the primary irritation index for the test article was calculated to be 0. See table 2.

**11.0 Conclusion**

The test result showed that the response of the test article extract was categorized as negligible under the test condition.

**12.0 Record**

All raw data pertaining to this study and a copy of the final report are retained in designated Huatongwei archive.

**13.0 Confidentiality Agreement**

Statements of confidentiality were as agreed upon prior to study initiation.

P20

**Table 2 Skin irritation response observation**

Reagent	Rabbit No	Pretest weight(g)	Finished weight(g)	Group	Reaction	Interval (hours): score-left/right			
						1h	24h	48h	72h
SC	1	2.10	2.18	Test Article	Erythema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
				Test Article	Oedema	0/0	0/0	0/0	0/0
				Negative Control	Oedema	0/0	0/0	0/0	0/0
	2	2.09	2.16	Test Article	Erythema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
				Test Article	Oedema	0/0	0/0	0/0	0/0
				Negative Control	Oedema	0/0	0/0	0/0	0/0
	3	2.17	2.24	Test Article	Erythema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
				Test Article	Oedema	0/0	0/0	0/0	0/0
				Negative Control	Oedema	0/0	0/0	0/0	0/0
Primary irritation index						0			
SO	4	2.04	2.14	Test Article	Erythema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
				Test Article	Oedema	0/0	0/0	0/0	0/0
				Negative Control	Oedema	0/0	0/0	0/0	0/0
	5	2.11	2.20	Test Article	Erythema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
				Test Article	Oedema	0/0	0/0	0/0	0/0
				Negative Control	Oedema	0/0	0/0	0/0	0/0
	6	2.13	2.21	Test Article	Erythema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
				Test Article	Oedema	0/0	0/0	0/0	0/0
				Negative Control	Oedema	0/0	0/0	0/0	0/0
Primary irritation index						0			

**Table 3 Positive control**

Rabbit No	Group	Reaction	Interval (hours): score-left site/right site			
			1h	24h	48h	72h
1	Positive control	Erythema	0/0	1/2	2/3	3/3
		Oedema	0/0	2/1	2/2	3/3
	Negative Control	Erythema	0/0	0/0	0/0	0/0
		Oedema	0/0	0/0	0/0	0/0
2	Positive control	Erythema	0/1	2/1	3/3	4/5
		Oedema	1/0	2/2	3/3	3/4
	Negative Control	Erythema	0/0	0/0	0/0	0/0
		Oedema	0/0	0/0	0/0	0/0
3	Positive control	Erythema	1/0	1/2	3/3	4/5
		Oedema	0/1	2/1	3/4	3/4
	Negative Control	Erythema	0/0	0/0	0/0	0/0
		Oedema	0/0	0/0	0/0	0/0
Primary irritation index			5.2			

Positive control performed once every six months see CSTBB20020001P3(Finish date: 2020-Q2-21)



## In Vitro Cytotoxicity Test

### MTT Method

#### Final Report

Article Name: Disposable medical face mask  
 Report Number: CSTBB20030204  
 Method Standard: ISO 10993-5: 2009

**Sponsor**

Hunan EEXI Technology&Service Co.,Ltd.  
 No 6 north of Pingtuo road, Liuyang Hi-tech industrial development zone, Hunan, China

**Test Facility**

CCIC Huatongwei international inspection (Suzhou) Co., Ltd  
 Room 101, Building G, Raoshui Road 388, Suzhou, Jiangsu, China

### CONTENTS

Notices.....	3
Abstract.....	4
Study Verification and Signature.....	5
1.0 Purpose.....	6
2.0 Reference.....	6
3.0 Test and control articles.....	6
4.0 Identification and justification of test system.....	6
5.0 Equipment and reagents.....	7
6.0 Experiment design and dose.....	7
7.0 Statistical method.....	8
8.0 Evaluation criteria.....	8
9.0 Results of the test.....	8
10.0 Conclusion.....	9
11.0 Record.....	9
12.0 Confidentiality Agreement.....	9

**Notices**

1. Please apply for rechecking within 15 days of receiving the report if there is any objection.
2. Any erasure or without special testing seal renders the report null and void.
3. The report is only valid when signed by the persons who edited, checked and approved it.
4. The report is only responsible for the test results of the tested samples.
5. The report shall not be reproduced except in full without the written approval of the company.



**Abstract**

In this study, mammalian L-929 cells were cultured in vitro according to ISO 10993-5:2009 to test the potential cytotoxicity of the test article.

The test articles and the control material were separately placed in MEM medium containing 10% fetal bovine serum, and extracted in a 37 °C incubator for 24 hours. After the end of the extraction, the cell culture medium in the 96-well plate (10<sup>4</sup> cells/well) cultured for 24 hours was removed and replaced with the corresponding extract, cultured in 37 °C, 5% CO<sub>2</sub>, >90% humidity for 24 hours. After the culture, the morphology and cell lysis of the cells were observed under the microscope, and the cytotoxicity of the test samples was determined by MTT assay.

The results showed that the cells in the blank control group and the negative control group (high density polyethylene) were well-formed throughout the experiment and showed no cytotoxic reaction. A severe cytotoxic response was shown in the positive control group (ZDEC). The 100% concentration of the test extract retained a normal appearance after 24 hours of incubation, and the cell viability was 78.1%. The data of each group met the acceptance criteria, and the results of this test were valid.


Based on the above results, it can be concluded that under the experimental conditions, the test article Disposable medical face mask have no potential toxicity to L-929 in the MTT method.

**Study Verification and Signature**



Protocol Number SST2003006602BB  
 Protocol Effective Date 2020-03-23  
 Technical Initiation Date 2020-03-23  
 Technical Completion Date 2020-03-25  
 Final Report Completion Date 2020-04-18

Personnel  2020-04-18  
 Date Completed

Approved  2020-04-18  
 Study Director Date Completed

Supervisory    
 Test Facility Manager Date Completed

**1.0 Purpose**

The purpose of the test is to determine the potential cytotoxicity toxicity of a mammalian cell culture (mouse fibroblast L-929 cells) in response to the test article.

**2.0 Reference**

Biological evaluation of medical devices-Part 5: Tests for In Vitro Cytotoxicity (ISO 10993-5, 2009)  
 Biological evaluation of medical devices-Part 12: Sample preparation and reference materials (ISO 10993-12: 2012)

**3.0 Test and control articles**

Groups	Test article	Negative Control Article	Positive Control Article	Blank Control
Name	Disposable medical face mask	High Density Polyethylene Film	ZDEC	MEM medium, with addition 10% FBS
Manufacture	Huatao EEXI Technology&Service Co.,Ltd.	Hatano Research Institute, FDSC	Sigma-Aldrich.	Hyclone
Size	17.5cm*9.5cm	3 cm*10 cm (5 sheets)	25 g	500 ml
Model	YN001	/	/	/
Lot Batch#	Not provided	C-161	BCB06847V	AE29441978
Test Article Material	PP non-woven, Melblown filtration layer, earloop and nose clip	/	/	/
Physical State	Solid	Solid	Solid	Liquid
Color	Not provided	White	White	Pink
Packaging Material	PE poly bag and paper box	/	/	/
Sterilized or Not	No	No	No	Yes
Concentration	/	/	0.1%	/
Total Surface	Not provided	/	/	/
Storage Condition	Room Temp.	Room Temp.	Room Temp.	4°C

Note: The information about the test article was supplied by the sponsor wherever applicable.

**4.0 Identification and justification of test system**

L-929 mouse fibroblast cells obtained from American Type Culture Collection (ATCC). L-929 cells have been used for cytotoxicity studies because they demonstrate sensitivity to extractable cytotoxic articles. Also, the test article is extracted and administered in vitro to mouse fibroblast L929 cells through a solvent compatible with the test system, which is the optimal route of administration available in this test system as recommended in ISO 10993-5.

**5.0 Equipment and reagents**

**5.1 Instruments**

Vertical pressure steam sterilizer (SHB026), CO<sub>2</sub> Incubator (SHB002), Steel Straight Scale (SHB076), Electronic Balance (SHB016), Clean bench (SHB014), Multifan Spectrum Microplate Spectrophotometer (SHB003), Bench top low speed centrifuge (SHB022), Inverted microscope (SHB005)

**5.2 Reagents**

MEM (HyClone, AE2941978), FBS (Clark, JC5116), Penicillin-Streptomycin (Gibco, 2145453), Trypsin (Gibco, 204080), PBS (HyClone, AE29451445), MTT (3-(4,5-Dimethylthiazol-2-yl)-2,5-diphenyltetrazolium bromide) (Sigma, MKBG2038V), Isopropyl alcohol (Macklin, C10394867)

**6.0 Experiment design and dose**

**6.1 Sample preparation**

According to the table below, aseptic extraction of the test article sealed and incubated in MEM medium (10% FBS) at 37 °C, 5% CO<sub>2</sub> and 60 rpm for 24 hours.

Groups	Sampling		Sterilization Method	Aseptic Extraction In Inert Container				Final Extract Clear or Not
	Manner	Actually sampling		Ratio	Extracts	Condition	pH	
Test article	Whole	570.2 cm <sup>2</sup>	EO	6 cm <sup>2</sup> : 1 ml	95.0 ml	37 °C 24 h	7.4	Clear
Negative Control	Random	60 cm <sup>2</sup>	UV	3 cm <sup>2</sup> : 1 ml	20.0 ml	37 °C 24 h	7.4	Clear
Positive Control	Random	0.02 g	Filter	0.1 g: 100 ml	20.0 ml	37 °C 24 h	7.4	Clear
Blank Control	/	/	/	/	10.0 ml	37 °C 24 h	7.4	Clear

The changes of the leaching solution was observed after extraction. No particulates or color changes were observed in pre- and post-extraction, and immediately be used in the follow-up experiment. The color and pH of the extraction solution did not change before and after use, and the pH value was 7.4 after leaching.

**6.2 Test method**

Aseptic procedures were used for handling cell cultures. L-929 cells were cultured in MEM medium (10% FBS, 1% Penicillin-Streptomycin solution) at 37 °C in a humidified atmosphere of 5% CO<sub>2</sub>, then digested by 0.25% trypsin containing EDTA to get single cell suspension. 1 × 10<sup>6</sup> cell/ml suspension were obtained by centrifuging (1000 rpm, 5 min) and re-dispersing in MEM medium.

The suspended cells were dispensed at 100 µl per well in 96-well plate, and cultured in cell incubator (5% CO<sub>2</sub>, 37 °C, >90% humidity). Cell morphology was evaluated to verify that the monolayer was satisfactory.

After the cells grew to about 70% and form a monolayer, original culture medium was discarded. The 96-well plates were then treated with 100 µl of extract of test article (100%, 75%, 50%, 25%), control article, negative article and positive article respectively. The 96-well plate was incubated at 37 °C in cell incubator of 5% CO<sub>2</sub> for 24 h. Six replicates of each test were tested.

			cell lysis and slight growth inhibition.
25% Test article extract			The cells showed a round shape and a change in cell morphology occasionally, and there were particles in the cytoplasm, occasionally cell lysis and slight growth inhibition.

**9.2 Results of the cell vitality**

Table2 Results of the cell vitality

Group	OD value								Viab. (%)
	1	2	3	4	5	6	$\bar{x}$	s	
Blank control	0.601	0.663	0.638	0.655	0.606	0.629	0.632	0.025	100.0
Negative control	0.598	0.582	0.598	0.599	0.597	0.593	0.595	0.006	94.1
Positive control	0.054	0.062	0.061	0.059	0.068	0.064	0.061	0.005	9.7
100% test article extract	0.489	0.497	0.486	0.495	0.492	0.502	0.494	0.006	78.1
75% test article extract	0.523	0.525	0.535	0.524	0.519	0.508	0.522	0.009	82.6
50% test article extract	0.539	0.538	0.547	0.536	0.535	0.537	0.539	0.004	85.2
25% test article extract	0.549	0.555	0.567	0.589	0.585	0.565	0.568	0.016	89.9

**10.0 Conclusion**

Under the conditions of this study, the test article have no potential toxicity to L-929 cells.

**11.0 Record**

All raw data pertaining to this study and a copy of the final report are to be stored in the designated archive files at Histogwei.

**12.0 Confidentiality Agreement**

Statements of confidentiality were as agreed upon prior to study initiation.

After incubation, observe the cell morphology first and then discard the culture medium. Add 50 µl MTT (1mg/ml) to each well and then incubated at 37 °C in a humidified atmosphere of 5% CO<sub>2</sub> for 2 hours. The liquid in each well was tipped out and 100 µl isopropyl alcohol was added to each well to suspend the cell layer.

Evaluate the suspension above with a dual-wavelength spectrophotometer with the measurement wavelength at 570 nm.

**7.0 Statistical method**

Mean±standard deviation ( $\bar{x} \pm s$ )

The cell cytotoxicity ratio = OD<sub>750</sub> of test (or positive or negative) article group/ OD<sub>750</sub> of blank control group × 100%.

**8.0 Evaluation criteria**

8.1 The 50% extract of the test article should have at least the same or a higher viability than the 100% extract. Otherwise the test should be repeated.

8.2 The lower the Viab.% value, the higher the cytotoxic potential of the test article is.

8.3 If viability is reduced to < 70% of the blank, it has a cytotoxic potential.

8.4 The Viab.% of the 100% extract of the test article is the final result.

**9.0 Results of the test**

**9.1 Results of the cell morphology**

Table 1 Observation of the cell morphology

Group	Before inoculation	Before treated with extract	24 h after treatment
Blank control			Discrete intracytoplasmic granules, no cell lysis, no reduction of cell growth.
Negative control			Discrete intracytoplasmic granules, no cell lysis, no reduction of cell growth.
Positive control			Nearly complete or complete destruction of the cell layers.
100% Test article extract	Discrete intracytoplasmic granules, no cell lysis, no reduction of cell growth.	Discrete intracytoplasmic granules, no cell lysis, no reduction of cell growth.	The cells showed a round shape and a change in cell morphology occasionally, and there were particles in the cytoplasm, occasionally cell lysis and slight growth inhibition.
75% Test article extract			The cells showed a round shape and a change in cell morphology occasionally, and there were particles in the cytoplasm, occasionally cell lysis and slight growth inhibition.
50% Test article extract			The cells showed a round shape and a change in cell morphology occasionally, and there were particles in the cytoplasm, occasionally



### 中华人民共和国医疗器械注册证

注册证编号：湘械注准20202140297

注册人名称	湖南一喜科技发展有限公司
注册人住所	湖南浏阳高新技术产业开发区坪头北路6号3栋
生产地址	湖南浏阳高新技术产业开发区坪头北路6号3栋
代理人名称	不适用
代理人住所	不适用
产品名称	一次性使用医用口罩（非无菌型）
型号、规格	产品按照面罩形状分为平面形（PM），按照佩戴方式可以分为耳挂式（G）、带式（B）或头带（D）式，按照尺寸可以分为A、B、C、D、E、F、G七种规格。
结构及组成	本产品由罩布、鼻夹、口罩带组成。罩布内外层由无纺布，中间层由熔喷布制成。口罩带由非织造布或松紧带制成，鼻夹由可弯折的聚丙烯鼻夹材料制成。该产品非无菌供应。
适用范围	适用于佩戴者在无体液和飞溅风险的普通医疗环境下的卫生护理。
附件	产品技术要求
其他内容	
备注	1. 该产品为应急审批注册，有效期为六个月；2. 该产品在延续/变更注册时应按医疗器械注册管理要求完善相关资料。

审批部门：湖南省药品监督管理局

批准日期：2020年05月08日

有效期至：2020年11月07日



### 中华人民共和国 PEOPLE'S REPUBLIC OF CHINA 医疗器械产品出口销售证明 CERTIFICATE FOR EXPORTATION OF MEDICAL PRODUCTS

证书编号：湘药监械出 20200033 号  
Certificate NO.: HNMPA 20200033

产品名称：见附件  
Product(s): Attachment

规格型号：见附件  
Model: Attachment

产品注册或备案凭证号：见附件  
Registration certificate(s): Attachment

生产企业：湖南一喜科技发展有限公司  
Manufacturer: Hunan EEXI Technology & Service Co., Ltd.

生产企业住所：湖南浏阳高新技术产业开发区坪头北路6号3栋  
Address of manufacturer: No. 6, North of Pingtou road, Liuyang Hi-tech industrial development zone, Hunan, China

生产许可或备案凭证号：湘食药监械生产许 20200023 号  
Manufacturing License(s): Hunan CFDA Production Permit No. 20200023

兹证明上述产品已准许在中国生产和销售。  
This is to certify that the above products have been registered to be manufactured and sold in China.

证明有效日期至：2022年3月19日  
This certification valid until: Mar. 19, 2022

湖南省药品监督管理局  
HUNAN MEDICAL PRODUCTS ADMINISTRATION

2020年3月20日  
Mar. 20, 2020





附件  
(Attachment)

序号 No.	产品名称 Product (s)	规格型号 Model	注册证号 Registration certificate(s)
1	一次性使用医用口罩 (非无菌) Disposable medical mask (non-sterile)	规格型号: 平面挂耳型, A 11.5cm×7cm, B 12.5cm× 9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm×9.5cm, G 14.5cm×9cm Model: Flat mounting ear model, A 11.5cm× 7cm, B 12.5cm×9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm× 9.5cm, G 14.5cm×9cm	器械注册: 20202140297 Registration certificate(s): Hunan Machinery Registration: 20202140297
2	一次性使用医用口罩 (无菌型) Disposable medical face mask (Sterilized)	规格型号: 平面挂耳型, A 11.5cm×7cm, B 12.5cm× 9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm×9.5cm, G 14.5cm×9cm Model: Flat mounting ear model, A 11.5cm× 7cm, B 12.5cm×9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm× 9.5cm, G 14.5cm×9cm	器械注册: 20202140289 Registration certificate(s): Hunan Machinery Registration: 20202140289
3	医用外科口罩 (无菌型) Surgical mask (Sterilized)	规格型号: 平面挂耳型, A 11.5cm×7cm, B 12.5cm× 9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm×9.5cm, G 14.5cm×9cm Model: Flat mounting ear model, A 11.5cm× 7cm, B 12.5cm×9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm× 9.5cm, G 14.5cm×9cm	器械注册: 20202140298 Registration certificate(s): Hunan Machinery Registration: 20202140298
4	医用外科口罩 (非无菌) Surgical mask (non-sterile)	规格型号: 平面挂耳型, A 11.5cm×7cm, B 12.5cm× 9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm×9.5cm, G 14.5cm×9cm Model: Flat mounting ear model, A 11.5cm× 7cm, B 12.5cm×9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm× 9.5cm, G 14.5cm×9cm	器械注册: 20202140288 Registration certificate(s): Hunan Machinery Registration: 20202140288

**检测结论**

总结论: 湖南一喜科技服务有限公司的洁净生产车间、微生物室按《无菌医疗器具生产管理规范》YY 0033-2000 的规定对洁净生产车间符合 100000 级标准规定; 微生物室符合 10000 级标准规定。

评价人: 袁中 审核人: 刘露 批准人: 袁中

**认证证书**

证书号: USA20Q40924R0M

兹证明

**湖南一喜科技服务有限公司**

统一社会信用代码: 91430181MA4PHUE510

注册地址: 湖南浏阳高新技术产业开发区坪头北路 6 号  
生产地址: 湖南浏阳高新技术产业开发区坪头北路 6 号 3 栋一、二楼  
办公地址: 湖南浏阳高新技术产业开发区坪头北路 6 号办公楼

质量管理体系符合标准  
ISO 9001:2015

质量管理体系适用范围  
一次性使用医用口罩 (非无菌型)、一次性使用医用口罩 (无菌型)、医用外科口罩 (非无菌型)、医用外科口罩 (无菌型) 的生产 (该公司许可证范围内)

初次发证日期: 2020 年 04 月 20 日  
证书颁发日期: 2020 年 04 月 20 日  
证书有效期至: 2023 年 04 月 19 日

签发: 吴凤英

北京东方纵横认证中心有限公司

EACC ANAB IAF

**Certificate**

Certificate No.: USA20Q40924R0M

This is to certify that the Quality Management System of

**HUNAN EXEI TECHNOLOGY & SERVICE CO., LTD.**

Unified Social Credit Code: 91430181MA4PHUE510

Registered Office Address: No.6 North of Pingtuo Road, Liuyang Hi-Tech Industrial Development Zone, Hunan, China  
Production Address: Floor 1 and 2, Building 3, No.6 North of Pingtuo Road, Liuyang Hi-Tech Industrial Development Zone, Hunan, China

Has been audited to conform to the following Quality Management System standard  
ISO 9001:2015

This Quality Management System is valid for the  
The production of disposable medical masks (non-sterile), disposable medical masks (sterile), medical surgical masks (non-sterile) and medical surgical masks (sterile) (within the company's license scope)

Date of initial issuance: Apr. 20, 2020  
Date of issuance: Apr. 20, 2020  
Date of renewal: May 12, 2020  
Date of expiry: Apr. 19, 2023

Issued by: Wu Fengying

北京东方纵横认证中心有限公司  
Beijing East Allreach Certification Center Co., Ltd.

EACC ANAB IAF



Test Report

(Electronic version)

No: 20R000976

Bacterial filtration efficiency (BFE)

细菌过滤效率 (BFE)

Test method: EN 14683:2019+AC:2019 Annex B

测试方法: EN 14683:2019+AC:2019 附录B

Test principle:

测试原理:

A specimen of the mask material is clamped between a six-stage cascade impactor and an aerosol chamber. An aerosol of Staphylococcus aureus is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency (BFE) of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

把口罩材料的试样夹在六级冲击采样器和气溶胶室之间,将金黄色葡萄球菌的气溶胶引入到气溶胶室,并在真空状态下通过口罩和采样器,口罩的细菌过滤效率(BFE)等于通过应用口罩材料的菌落形成单位与气溶胶室中的菌落形成单位总数的比例,用百分数表示。

Test equipment:

测试设备:

Incubator  
恒温培养箱  
Electronic balance  
电子天平  
Autoclave  
压力蒸汽灭菌锅

Experimental system for bacterial filtration efficiency (BFE) of mask

口罩细菌过滤效率(BFE)实验系统

The environmental conditions of the laboratory and test condition:

实验室环境条件和测试条件:

Total bacteria: 0 CFU/plate

细菌总数: 0 CFU/皿

Total fungi: 0 CFU/plate

真菌总数: 0 CFU/皿

Blank experiment: Aseptic growth

空白实验: 无菌生长

Test environment temperature: 24.0°C, Relative humidity: 56.0%

测试环境温度: 24.0°C, 相对湿度: 56.0%

Culture medium: TSA agar medium

培养基名称: TSA琼脂培养基

Culture temperature: 37°C, Culture time: 48h

样品培养温度: 37°C, 培养时间: 48h

Test bacteria: staphylococcus aureus ATCC 6538

测试菌株: 金黄色葡萄球菌 ATCC 6538



Test Report

(Electronic version)

No: 20R000976

Results 测试结果:

Sample 样品	T 计数之和	BFE 细菌过滤效率 (%)	Requirement 技术要求 (%)	Classification 类别	Conclusion 单项结论
1	13	99.31	≥95	Type I	Pass 符合
2	22	98.84			
3	22	98.84			
4	15	99.21			
5	18	99.65			

Remarks 备注:

For each test specimen calculate the bacterial filtration efficiency B, as a percentage, using the following formula:

对于每个试样,按以下公式以百分形式计算细菌过滤效率:

$B = (C - T) / C \times 100$

where 式中

B is bacterial filtration efficiency (BFE), %;

C is positive control average;

T is the total plate count for the test specimen.

— 试验样品计数之和。



Test Report

(Electronic version)

No: 20R000976

Concentration of bacterium:  $5.0 \times 10^8$  CFU/ml

菌液浓度:  $5.0 \times 10^8$  CFU/ml

Positive control average (C):  $1.9 \times 10^8$  CFU

阳性质控平均值 (C):  $1.9 \times 10^8$  CFU

Negative monitor count: <1 CFU

阴性质控值: <1 CFU

Test area: 49 cm<sup>2</sup>

测试面积: 49 cm<sup>2</sup>

Flow rate: 28.3 l/min

气流流速: 28.3 l/min

Dimensions of the test specimens: 15cm\*15cm

试样尺寸: 15cm\*15cm

Pretreatment: Condition each specimen for 4 h by exposure to a temperature of (21±5)°C and a relative humidity of 85±5%

预处理方式: 温度 (21±5)°C, 相对湿度 (85±5)% 环境中预处理 4h

Mean particle size: 3.0 μm

平均粒径直径: 3.0 μm

The medical face mask in contact with the bacterial challenge: inside

口罩与细菌气溶胶接触点: 里层



Test Report

(Electronic version)

No: 20R000976

Results 测试结果:

Sample 样品	Bacteria 细菌 (CFU/g)	Fungi 真菌 (CFU/g)	Microbial cleanliness 微生物 (CFU/g)	Requirement 技术要求 (CFU/g)	Classification 类别	Conclusion 单项结论
1	3	5	8	≤20	Type I	Pass 符合
2	5	3	8			
3	2	4	6			
4	1	0	1			
5	4	3	7			



Test Report

(Electronic version)

No: 20R000976

Microbial cleanliness

洁净度-微生物

Test method: EN ISO 11737-1:2018, Membrane filtration  
测试方法: EN ISO 11737-1:2018 膜过滤法

Test principle:

Take the required samples from the original packaging. Weigh a certain amount of sample and placed in a sterile 500 ml bottle containing 300 ml of extraction liquid (1 g/l Peptone, 5 g/l NaCl and 2 g/l Tween 20). The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm. After this extraction step, 100 ml of the extraction liquid is filtered through a 0.45 µm filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 ml aliquot of the same extraction liquid is filtered in the same way and the filter plated on Sabouraud Dextrose agar (SDA) for fungi enumeration. The plates are incubated for 3 days at 30 °C and 7 days at (20 to 25) °C for TSA and SDA plates respectively. The total turbidity is expressed by addition of the TSA and SDA counts.  
从原始包装中取出所需样品, 在一定条件下称取一定量的样品装收到无菌瓶中, 其中含有300ml的萃取液(1g/l的蛋白胨, 5g/l NaCl和2g/l Tween 20), 在250rpm下振荡时间5min, 提取100ml萃取液, 用0.45µm的薄膜过滤器, 将滤液过滤到TSA平板上, 用于测定细菌菌落总数, 取100ml萃取液, 用0.45µm的薄膜过滤器, 将滤液过滤到SDA平板上, 用于测定真菌菌落总数, 这些TSA和SDA平板分别在30℃下培养3天和20-25℃培养7天, 总的微生物含量用TSA和SDA的计数和来表示。

Test equipment:

测试设备:  
Constant temperature incubator 恒温培养箱  
Electronic balance 电子天平  
Pressure steam sterilizer 压力蒸汽灭菌锅  
Biosafety cabinet 生物安全柜

The environmental conditions of the laboratory and test condition:

实验室环境条件和测试条件:  
Test environment temperature: 24.0 °C, relative humidity: 56.0%  
测试环境温度: 24.0 °C 相对湿度: 56.0%  
Test environment monitoring: total bacteria: 0 CFU/plate, total fungi: 0 CFU/plate, blank experiment: sterile growth  
测试环境监测: 细菌总数: 0 CFU/皿, 真菌总数: 0 CFU/皿, 空白实验: 无菌生长



Guangzhou Inspection Testing and Certification Group Co., Ltd.  
Add: No. 1, Zhong Road, Panyu District, Guangzhou, Guangdong, P.R.China

Tel: +86-20-61994599 61994599

Test Report

(Electronic version)

No: 20R000976

Results 测试结果:

Sample 样品	Differential pressure 压力差 (Pa/cm <sup>2</sup> )	Requirement 技术要求 (Pa/cm <sup>2</sup> )	Classification 级别	Conclusion 单项结论
1	24.5	<40 EN 14683:2019+AC:2019	Type I	Pass 符合
2	27.1			
3	23.3			
4	25.1			
5	27.3			



—本报告结束 (End of Report)—

Guangzhou Inspection Testing and Certification Group Co., Ltd.  
Add: No. 1, Zhong Road, Panyu District, Guangzhou, Guangdong, P.R.China

Tel: +86-20-61994599 61994599

Test Report

(Electronic version)

No: 20R000976

Differential pressure

压力差

Test method: EN 14683:2019+AC:2019 Annex C  
测试方法: EN 14683:2019+AC:2019 附录C

Test principle:

This procedure was performed to evaluate the differential pressure of the medical face mask material by measuring the air exchange pressure through a measured surface area at a constant air flow rate.  
通过口罩压力差测试仪, 抽取空气以恒定的流速通过已经测定表面面积的医用口罩材料, 从而测定空气交换的压力差。

Test equipment:

测试设备:  
GTTC-YLC-1 Apparatus for differential pressure  
GTTC-MLC-1口罩压力差测试仪

The environmental conditions of the laboratory and test condition:

实验室环境条件和测试条件:  
Air flow: 8 l/min  
气体流量: 8 l/min  
Test area: 4.9cm<sup>2</sup>  
试验面积: 4.9cm<sup>2</sup>  
Pre-treatment: Condition each specimen for a minimum of 4 h by exposure to a temperature of (21±5) °C and a relative humidity of (85±5)%  
预处理方式: 温度 (21±5) °C, 相对湿度 (85±5)% 环境中预处理大于4h  
General location of the areas of the mask the differential measurements: specimen center  
口罩压力差测试大概位置: 试样中心



Guangzhou Inspection Testing and Certification Group Co., Ltd.  
Add: No. 1, Zhong Road, Panyu District, Guangzhou, Guangdong, P.R.China

Tel: +86-20-61994599 61994599