

1. Table of Contents (Page 1)
2. Basic Information (Page 2)
3. Packaging & Pictures (Page 3-4)
4. European Test & Technical Documentation Review Report  
(Page 5-9)
5. Registration on DIMDI & Declaration of Conformity (Page  
10-11)
6. ISO Production Standards Report (Page 12-19)
7. Manufacturer Qualification Certificate & Export License  
(Page 20-22)

## **Disposable Medical Face Mask Inherent-Type IIR Strap-on**

**Brand: Inherent**

**Type No.: YX121 (Strap-on)**

**Performance standard: EN 14683:2019 + AC:2019(E) Annex B /C/D & YY/T 0969-2013, tested by TÜV Rheiland & Intertek**

**Production standards: ISO 10933-10:2010 & ISO 10993-5:2009, tested by CCIC**

**Classifications: Type IIR Strap-on (Non-Sterile), No latex ingredients**

**Material: Two-layer PP Nonwoven fabric and One-layer BFE95 Meltblown fabric (3 layers)**

**Mask specifications: Universal, 17,50 x 9,50 cm**

**Earloop design: Breathable and comfortable for prolonged wearing**

**Expiration date: 2 years from date of manufacturing**

**P2**


**Packing specifications: 50 pcs. / box, 40 boxes 2000 pcs. /  
carton, 16 cartons / euro pallet**

**Carton measurements: 57 x 39 x 42 cm, G.W. 9.2 KGS.**

**Stock available: Rotterdam The Netherlands, Frankfurt a/m  
Germany and Madrid Spain**



**P3**



Produkte Products TÜVRheinland®

Prüfbericht-Nr.: 60377902 002 Auftrags-Nr.: 168264746 Seite 1 von 6  
Test Report No.: Order No.: Page 1 of 6

Kunden-Referenz-Nr.: N/A Auftragsdatum: May 13, 2020  
Client Reference No.: Order date:

Auftraggeber: Hunan EEXI Technology & Service Co., Ltd.  
Client: No. 8, North of Pingzhou road, Liuyang Hi-tech industrial development zone, Hunan, China

Prüfgegenstand: Surgical Face Mask (non-sterile)  
Test item:

Bezeichnung / Typ-Nr.: YX011, YX121  
Identification / Type No.:

Auftrags-Inhalt: Type test  
Order content:

Prüfgrundlage: EN 14683:2019+AC:2019 except for clause 5.2.6  
Test specification:

Warnungseingangsdatum: May 25, 2020  
Date of receipt:

Prüfmuster-Nr.: 20200522  
Test sample No.:

Prüfzeitraum: N/A  
Testing period:

Ort der Prüfung: N/A  
Place of testing:

Prüflaboratorium: TÜV Rheinland (Shenzhen)  
Testing laboratory:

Prüfergebnis\*: Pass  
Test result\*:

geprüft von / tested by: Lucy Jiang kontrolliert von / reviewed by: Angela Chen / Department Manager

Jun. 02, 2020 Lucy Jiang / Assistant Project Engineer Jun. 02, 2020 Angela Chen / Department Manager

Unterschrift Name / Position Datum Name / Position Unterschrift  
Signature Date Signature

Sortiertes / Other:  
- The test report consists of EN 14683 test report including this cover page (6 pages) and attachment: Photo documentation (6 pages).  
- This report is only valid in conjunction with previous report No. 60377902 001.

Zustand des Prüfgegenstandes bei Anlieferung: Prüfmuster vollständig und unbeschädigt  
Condition of the test item at delivery: Test item complete and undamaged

Legende:  
1 = sehr gut 2 = gut 3 = befriedigend 4 = ausreichend 5 = mangelhaft  
Pass = entspricht o.g. Prüfgrundlage Fail = entspricht nicht o.g. Prüfgrundlage Not = nicht anwesend Not = not present  
Legend: 1 = very good 2 = good 3 = satisfactory 4 = sufficient 5 = poor  
Pass = correct a.m. test specification Fail = failed a.m. test specification Not = not applicable Not = not tested

Dieser Prüfbericht bezieht sich nur auf das o.g. Prüfmuster und darf ohne Genehmigung der Prüfstelle nicht auszusweise vervielfältigt werden. Dieser Bericht berechtigt nicht zur Verwendung eines Prüfzeichens.  
This test report only relates to the a.m. test sample. Without permission of the test center this test report is not permitted to be duplicated in extracts. This test report does not entitle to carry any test mark.

TÜV Rheinland (Shenzhen) Co., Ltd., East of F11, F12, F18, Building 1, Cybio Technology Building, No. 6 Langshan No. 2 Road, North Hi-tech Industry Park, Nanshan District, Shenzhen, P.R. China  
http://www.tuv.com

Produkte Products TÜVRheinland® Page 2 of 6 Report No. 60377902 002

EN 14683:2019+AC: 2019  
Medical face masks –  
Requirements and test methods

Report Reference No.: 60377902 002  
Date of issue: See cover page  
Total number of pages: See cover page

Testing Laboratory: TÜV Rheinland (Shenzhen) Co., Ltd.  
Address: 1F East & 2-4F, Cybio Technology Building No. 1, No. 16 Kejibel 2nd Road, High-Tech Industrial Park North Nanshan District, 518057, Shenzhen, China

Applicant's name: Hunan EEXI Technology & Service Co., Ltd.  
Address: No.8, North of Pingzhou road, Liuyang Hi-tech industrial development zone, Hunan, China

Test specification: Standard: EN 14683:2019+AC:2019  
Test procedure: Type test  
Non-standard test method: N/A  
Test Report Form No.: EN 14683:2019+AC:2019\_A  
Test Report Form Originator: TÜV Rh (SZ)  
Master TRF: 2020-03  
Test item description: Surgical Face Mask (non-sterile)  
Trade Mark:

Manufacturer: Same as the applicant  
Model/Type reference: YX011, YX121  
Classification: Type IIR

QM-FR-333085HG Revision number: 1.0 Effective date: 2020-03-12

Produkte Products TÜVRheinland® Page 3 of 6 Report No. 60377902 002

List of Attachments (including a total number of pages in each attachment):  
Attachment – Photo Documentation (6 pages)

Summary of testing:  
Tests performed (name of test and test clause):  
Construction check according to:  
Clause 5.1.1 Materials and construction  
Clause 5.1.2 Design

Testing location:  
TÜV Rheinland (Shenzhen) Co., Ltd.  
1F East & 2-4F, Cybio Technology Building  
No. 1, No. 16 Kejibel 2nd Road, High-Tech  
Industrial Park North Nanshan District, 518057,  
Shenzhen, China

QM-FR-333085HG Revision number: 1.0 Effective date: 2020-03-12

Produkte Products TÜVRheinland® Page 4 of 6 Report No. 60377902 002

Copy of marking plate

The artwork below may be only a draft. The use of certification marks on a product must be authorized by the respective NCBs that own these marks.

Article No. YX121

Product name: Surgical Face Mask(non-sterile) Date of preparation of the manual: April 10, 2020  
Manual version number: 01

Type: Type IIR

Type and Specification: Flat, six-on mask, 17.5x5.5cm

Production license No.: Human Food/Drug Administration permission: 20200523

Registration No.: Human medical device registration permission: 2020140208

Material: 3M H9500 non-woven, 75% non-woven fabric

Standard Applied: EN 14683:2019+AC:2019 YY0469-2011

Intended Use:  
The surgical face mask is single-use, disposable device, provided non-sterile, and intended to be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations.

Benefits and Components:  
The mask is made by three layers. The inner and outer layer are non-woven fabric, the middle layer is melt-blown polypropylene.

Main performance:  
1) Bacterial leakage efficiency: <0.05%  
2) Differential Pressure: <0.05 Pa/cm²  
3) Respirator clearance: <0.05 cm³  
4) Liquid resistance pressure: >16Pa

Instruction for six-on face mask:  
1. Clean a mask, with the dark side facing outwards, hold the mask by hand and insert it on your face, make sure that the nose clip is located on the middle of the nose.  
2. Pull the mask both above the ears and fast it appropriately on the top of the head.  
3. Put the rest of the mask both below the ears and fast it appropriately over the back of the head.  
4. Adjust the mask into fit the shape of the bridge of the nose, and then adjust the tightness of the mask both to make the mask fit your face.  
5. Do not touch the front of the mask after wearing a mask. You are supposed to untie the mask both first and then discard it into an specified container.

Precautions:  
1) Check the package is intact before use, confirm the marks and date of manufacture on the external package, and use it before the expiry date.  
2) Do not use the product, please dispose of it according to the requirements of the environmental protection agency or related authorities.  
3) This product is a disposable device. It is not recommended to the clean or reuse it, if you feel uncomfortable during use, please stop using it immediately or replace it with a new one.

Storage:  
Store in dry, well-ventilated, non-corrosive gas place, avoid high temperature.  
Shelf life: 2 years.

Warning of package symbols:  

Symbol	Description	Caution	Meaning
1	Do not use after the expiry date	Yes	Do not use after the expiry date
2	Do not use after the expiry date	Yes	Do not use after the expiry date
3	Do not use after the expiry date	Yes	Do not use after the expiry date
4	Do not use after the expiry date	Yes	Do not use after the expiry date

Batch code: refer to package  
Date of manufacture: refer to package  
Use by date: refer to package  
Authorized representative in the European Community: Shanghai International Holding Co., Ltd. (Shanghai)  
Address: Eilshausen 80, 20531 Hamburg, Germany  
Manufacturer: Hunan EEXI Technology & Service Co., Ltd.  
Address: No. 8 North of Pingzhou Road, Liuyang Hi-tech Industrial Development Zone, Hunan, China  
TEL: +86-400-933-686

See attachment for other information.

QM-FR-333085HG Revision number: 1.0 Effective date: 2020-03-12

P5

TÜVRheinland® Page 5 of 6 Report No. 60377902 002

Testing

Date of receipt of test item(s).....: See cover page

Dates of tests performed.....: See cover page

Possible test case verdicts:

- test case does not apply to the test object.....: N/A

- test object does meet the requirement.....: P (Pass)

- test object was not evaluated for the requirement.....: N/E (collateral standards only)

- test object does not meet the requirement.....: F (Fail)

General remarks:

\*(See Attachment #) refers to additional information appended to the report.

\*(See appended table) refers to a table appended to the report.

The tests results presented in this report relate only to the object tested.

This report shall not be reproduced except in full without the written approval of the testing laboratory.

List of test equipment must be kept on file and available for review.

Additional test data and/or information provided in the attachments to this report.

Throughout this report a ☐ comma / ☒ point is used as the decimal separator.

Name and address of factory (ies).....: Same as the applicant

General product information:

1. This report is only valid in conjunction with previous report No. 60377902 001. Refer to previous test report No. 60377902 001 for more details.

2. The change is to add one model YX121 which is same as previous model YX011 except for manner of wearing. YX121 uses tie-on belts and YX011 uses ear loops.

3. Clause 5.1.1 & 5.1.2 have been checked accordingly, no more test deem necessary after review.

QM-RT-33008SHG

Revision number: 1.0

Effective date: 2020-03-12

TÜVRheinland® Page 6 of 6 Report No. 60377902 002

EN 14683:2019+AC:2019			
Clause	Requirement + Test	Result - Remark	Verdict
5	Requirements		P
5.1	General		P
5.1.1	Materials and construction		P
	The medical face mask is a medical device, generally composed of a filter layer that is placed, bonded or moulded between layers of fabric.	The Surgical Face Masks are made of blank mask, nose clip and tie-on belts. The outer and inner layers of the mask are made of non-woven fabrics, and the middle layer is made of melt-blown polypropylene.	P
	The medical face mask shall not disintegrate, split or tear during intended use.		P
	In the selection of the filter and layer materials, attention shall be paid to cleanliness.		P
5.1.2	Design		P
	The medical face mask shall have a means by which it can be fitted closely over the nose, mouth and chin of the wearer and which ensures that the mask fits closely at the sides.		P
	Medical face masks may have different shapes and constructions as well as additional features such as a face shield (to protect the wearer against splashes and droplets) with or without anti-fog function, or a nose bridge (to enhance fit by conforming to the nose contour).	With nose clip	P
6	Marking, labelling and packaging		P
	Annex I, §13, of the Medical Devices Directive (93/42/EEC) or Annex I, §23, of the Medical Device Regulation (EU) 2017/745 specifies the information that should be specified on the packaging in which the medical face mask is supplied.	See page 4 and attachment.	P
	The following information shall be supplied:		P
	a) number of this European Standard;		P
	b) type of mask (as indicated in Table 1).		P
	EN ISO 15223-1:2016 and EN 1041:2008+A1:2013 should be considered.		P

End of EN 14683 test report

QM-RT-33008SHG

Revision number: 1.0

Effective date: 2020-03-12



Figure 1 Front / Back view of packaging box (Model: YX121)



Figure 2 Top view of packaging box (Model: YX121)




Figure 3 Side view of packaging box (Model: YX121)

P6



Figure 4 Side view of packaging box (Model: YX121)




Figure 5 Bottom view of packaging box (Model: YX121)



Figure 6 View of packaging bag (Model: YX121)




Figure 7 View of face mask (Model: YX121)



Figure 8 View of face mask (Model: YX121)



Figure 9 View of face mask (3-ply, Model: YX121)

END OF THE PHOTO DOCUMENTATION





Tests Conducted (As Requested By The Applicant)

Number: GZHT02294303-S1

## 2 Bacterial Filtration Efficiency (BFE)

As Per EN 14683:2019+AC:2019 Medical face masks – Requirements And Test Methods Annex B

Test Item	Results (%)					Performance Requirement for Medical Face Mask (%)
	Specimen (1)	Specimen (2)	Specimen (3)	Specimen (4)	Specimen (5)	
Bacterial Filtration Efficiency (BFE)	>99.9	99.9	99.8	99.9	99.9	Type III: ≥98

**Remarks:**

1. Biological Aerosol: *Staphylococcus aureus* (ATCC 6538).
2. Testing side: Outside of the test specimen was facing towards the challenge aerosol.
3. Test area: 78 cm<sup>2</sup>
4. Flow rate: 28.3 L/min
5. The average plate count results of the positive controls: 1900 CFU
6. The average plate count results of the negative controls: < 1 CFU
7. CFU = Colony-Forming Unit

This Item Is Not Under The Testing Scope Of CNAS Accreditation  
This Test Item Was Conducted In Caipin Road, Guangzhou Science City, GETDD, Guangzhou,  
Guangdong.

QIN / Hilaryxu

Page 5 of 7

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch  
 深圳天祥质量技术服务有限公司广州分公司

Room G2, 5/F & Room E1, E105/E201/E301/E405/E501/E601/E701/E801,  
 No.7-2, Cagier Road, Guangzhou Science City, GETCO, Guangzhou, China  
 中国广州经济技术开发区科学城彩虹桥7号2-1-8 02 02 02 01 01  
 001,  
 E01, E04, E05, E07, E08, E09, E10, E11, E12  
 01, 02, 03, 04, 05, 06, 07, 08, 09, 10, 11, 12, 13, 14, 15, 16, 17, 18, 19, 20, 21, 22, 23, 24, 25, 26, 27, 28, 29, 30, 31, 32, 33, 34, 35, 36, 37, 38, 39, 40, 41, 42, 43, 44, 45, 46, 47, 48, 49, 50, 51, 52, 53, 54, 55, 56, 57, 58, 59, 60, 61, 62, 63, 64, 65, 66, 67, 68, 69, 70, 71, 72, 73, 74, 75, 76, 77, 78, 79, 80, 81, 82, 83, 84, 85, 86, 87, 88, 89, 90, 91, 92, 93, 94, 95, 96, 97, 98, 99, 100, 101, 102, 103, 104, 105, 106, 107, 108, 109, 110, 111, 112, 113, 114, 115, 116, 117, 118, 119, 120, 121, 122, 123, 124, 125, 126, 127, 128, 129, 130, 131, 132, 133, 134, 135, 136, 137, 138, 139, 140, 141, 142, 143, 144, 145, 146, 147, 148, 149, 150, 151, 152, 153, 154, 155, 156, 157, 158, 159, 160, 161, 162, 163, 164, 165, 166, 167, 168, 169, 170, 171, 172, 173, 174, 175, 176, 177, 178, 179, 180, 181, 182, 183, 184, 185, 186, 187, 188, 189, 190, 191, 192, 193, 194, 195, 196, 197, 198, 199, 200, 201, 202, 203, 204, 205, 206, 207, 208, 209, 210, 211, 212, 213, 214, 215, 216, 217, 218, 219, 220, 221, 222, 223, 224, 225, 226, 227, 228, 229, 230, 231, 232, 233, 234, 235, 236, 237, 238, 239, 240, 241, 242, 243, 244, 245, 246, 247, 248, 249, 250, 251, 252, 253, 254, 255, 256, 257, 258, 259, 260, 261, 262, 263, 264, 265, 266, 267, 268, 269, 270, 271, 272, 273, 274, 275, 276, 277, 278, 279, 280, 281, 282, 283, 284, 285, 286, 287, 288, 289, 290, 291, 292, 293, 294, 295, 296, 297, 298, 299, 300, 301, 302, 303, 304, 305, 306, 307, 308, 309, 310, 311, 312, 313, 314, 315, 316, 317, 318, 319, 320, 321, 322, 323, 324, 325, 326, 327, 328, 329, 330, 331, 332, 333, 334, 335, 336, 337, 338, 339, 340, 341, 342, 343, 344, 345, 346, 347, 348, 349, 350, 351, 352, 353, 354, 355, 356, 357, 358, 359, 360, 361, 362, 363, 364, 365, 366, 367, 368, 369, 370, 371, 372, 373, 374, 375, 376, 377, 378, 379, 380, 381, 382, 383, 384, 385, 386, 387, 388, 389, 390, 391, 392, 393, 394, 395, 396, 397, 398, 399, 400, 401, 402, 403, 404, 405, 406, 407, 408, 409, 410, 411, 412, 413, 414, 415, 416, 417, 418, 419, 420, 421, 422, 423, 424, 425, 426, 427, 428, 429, 430, 431, 432, 433, 434, 435, 436, 437, 438, 439, 440, 441, 442, 443, 444, 445, 446, 447, 448, 449, 450, 451, 452, 453, 454, 455, 456, 457, 458, 459, 460, 461, 462, 463, 464, 465, 466, 467, 468, 469, 470, 471, 472, 473, 474, 475, 476, 477, 478, 479, 480, 481, 482, 483, 484, 485, 486, 487, 488, 489, 490, 491, 492, 493, 494, 495, 496, 497, 498, 499, 500, 501, 502, 503, 504, 505, 506, 507, 508, 509, 510, 511, 512, 513, 514, 515, 516, 517, 518, 519, 520, 521, 522, 523, 524, 525, 526, 527, 528, 529, 530, 531, 532, 533, 534, 535, 536, 537, 538, 539, 540, 541, 542, 543, 544, 545, 546, 547, 548, 549, 550, 551, 552, 553, 554, 555, 556, 557, 558, 559, 560, 561, 562, 563, 564, 565, 566, 567, 568, 569, 570, 571, 572, 573, 574, 575, 576, 577, 578, 579, 580, 581, 582, 583, 584, 585, 586, 587, 588, 589, 590, 591, 592, 593, 594, 595, 596, 597, 598, 599, 600, 601, 602, 603, 604, 605, 606, 607, 608, 609, 610, 611, 612, 613, 614, 615, 616, 617, 618, 619, 620, 621, 622, 623, 624, 625, 626, 627, 628, 629, 630, 631, 632, 633, 634, 635, 636, 637, 638, 639, 640, 641, 642, 643, 644, 645, 646, 647, 648, 649, 650, 651, 652, 653, 654, 655, 656, 657, 658, 659, 660, 661, 662, 663, 664, 665, 666, 667, 668, 669, 670, 671, 672, 673, 674, 675, 676, 677, 678, 679, 680, 681, 682, 683, 684, 685, 686, 687, 688, 689, 690, 691, 692, 693, 694, 695, 696, 697, 698, 699, 700, 701, 702, 703, 704, 705, 706, 707, 708, 709, 710, 711, 712, 713, 714, 715, 716, 717, 718, 719, 720, 721, 722, 723, 724, 725, 726, 727, 728, 729, 730, 731, 732, 733, 734, 735, 736, 737, 738, 739, 740, 741, 742, 743, 744, 745, 746, 747, 748, 749, 750, 751, 752, 753, 754, 755, 756, 757, 758, 759, 760, 761, 762, 763, 764, 765, 766, 767, 768, 769, 770, 771, 772, 773, 774, 775, 776, 777, 778, 779, 780, 781, 782, 783, 784, 785, 786, 787, 788, 789, 790, 791, 792, 793, 794, 795,

Tests Conducted (As Requested By The Applicant)

Manuscript: C2HTD3204303-S1

3. Splash Resistance Pressure (ISO 22609:2004):

Synthetic Blood Surface Tension: 0.040 N/m, Distance Between Blow Head Front End And Target Area: 300 mm, Artificial Blood Volumes: 2 mL, Test Pressure: 16.0 kPa, Velocity: 550 cm/s, Use A Fixed Target.

Tested Sample	Result	Performance Requirement for Medical Face Mask Type IIR: No penetration at 15.0 kPa
Specimen (1)	None seen	
Specimen (2)	None seen	
Specimen (3)	None seen	
Specimen (4)	None seen	
Specimen (5)	None seen	
Specimen (6)	None seen	
Specimen (7)	None seen	
Specimen (8)	None seen	
Specimen (9)	None seen	
Specimen (10)	None seen	
Specimen (11)	None seen	
Specimen (12)	None seen	
Specimen (13)	None seen	

Remark : 1. Test was conducted by CNTAC Testing Service Co.,Ltd.(Foshan)  
2.This Item Is Not Under The Testing Scope Of CNAS Accreditation

OTN / hilaryw

Page 6 Of 7

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch  
 深圳天祥质量技术服务有限公司广州分公司  
 Room 02, 1&F & Room 01, E101E2102103104105106107108109110111112113114115116117118119120121122123124125126127128129130131132133134135136137138139140141142143144145146147148149150151152, Caim Road, Guangzhou Science City, GUETCO, Guangzhou, China  
 中国广州城市技术开发区华南新城第七号之二—1&F 02 房, 01 房  
 0201, E301, E401, E501, E601, E701, E801  
 2/F, Hengbin Building, 235 Kailie Ave., Guangzhou  
 Economic & Technological Development District, Guangzhou, China  
 中国广州经济技术开发区开泰大道 235 号恒滨大厦 2 楼  
 +86 20 8206 8888 Fax +86 20 8222 8769 Postcode: 510726

Tests Conducted (As Requested By The Applicant)

Number: G24T02294303-S1

## 4 Differential Pressure (EN 14683:2019+AC:2019 Annex C)

Air flow: 8L/min, Test area diameter 25 mm, Test area: 4.9 cm<sup>2</sup>

Tested Sample	Result (Pa/cm <sup>2</sup> )	Performance Requirement for Medical Face Basis (Pa/cm <sup>2</sup> )
Specimen (1)	39.2	Type IIR < 60
Specimen (2)	41.8	
Specimen (3)	36.6	
Specimen (4)	37.2	
Specimen (5)	42.9	
Average	39.5	

Remark : 1. Test was conducted by CNTAC Testing Service Co.,Ltd.(Foshan)  
2. This Item Is Not Under The Testing Scope Of CNAS Accreditation

End of Report

*This report is made solely on the basis of your instructions and/or information and materials supplied by you. It is not intended to be a recommendation for any particular course of action. Intertek does not accept a duty of care or any other responsibility to any person other than the Client in respect of this report and only accepts liability to the Client insofar as is expressly contained in the terms and conditions governing Intertek's provision of services to you. Intertek makes no warranties or representations either express or implied with respect to this report as provided for in these terms and conditions. We have aimed to conduct the Review on a diligent and careful basis and we do not accept any liability to you for any loss arising out of or in connection with this report, in contract, tort, by statute or otherwise, in the event of a negligent or willful misconduct. No copy of this report (except for full test copy) shall be made without the written approval by Intertek.*

QCN / hilaryxu

Page 7 Of 7

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch  
 深圳天祥质量技术服务有限公司广州分公司

Room 02, 1&F & Room 01, E101E201E301E401E501E601E701E801E901E1001E1101E1201E1301E1401E1501E1601E1701E1801E1901E2001E2101E2201E2301E2401E2501E2601E2701E2801E2901E3001E3101E3201E3301E3401E3501E3601E3701E3801E3901E4001E4101E4201E4301E4401E4501E4601E4701E4801E4901E5001E5101E5201E5301E5401E5501E5601E5701E5801E5901E6001E6101E6201E6301E6401E6501E6601E6701E6801E6901E7001E7101E7201E7301E7401E7501E7601E7701E7801E7901E8001E8101E8201E8301E8401E8501E8601E8701E8801E8901E9001E9101E9201E9301E9401E9501E9601E9701E9801E9901E10001E1001E1002E1003E1004E1005E1006E1007E1008E1009E1010E1011E1012E1013E1014E1015E1016E1017E1018E1019E1020E1021E1022E1023E1024E1025E1026E1027E1028E1029E1030E1031E1032E1033E1034E1035E1036E1037E1038E1039E1040E1041E1042E1043E1044E1045E1046E1047E1048E1049E1050E1051E1052E1053E1054E1055E1056E1057E1058E1059E1060E1061E1062E1063E1064E1065E1066E1067E1068E1069E1070E1071E1072E1073E1074E1075E1076E1077E1078E1079E1080E1081E1082E1083E1084E1085E1086E1087E1088E1089E1090E1091E1092E1093E1094E1095E1096E1097E1098E1099E1100E1101E1102E1103E1104E1105E1106E1107E1108E1109E1110E1111E1112E1113E1114E1115E1116E1117E1118E1119E1120E1121E1122E1123E1124E1125E1126E1127E1128E1129E1130E1131E1132E1133E1134E1135E1136E1137E1138E1139E1140E1141E1142E1143E1144E1145E1146E1147E1148E1149E1150E1151E1152E1153E1154E1155E1156E1157E1158E1159E1160E1161E1162E1163E1164E1165E1166E1167E1168E1169E1170E1171E1172E1173E1174E1175E1176E1177E1178E1179E1180E1181E1182E1183E1184E1185E1186E1187E1188E1189E1190E1191E1192E1193E1194E1195E1196E1197E1198E1199E1200E1201E1202E1203E1204E1205E1206E1207E1208E1209E1210E1211E1212E1213E1214E1215E1216E1217E1218E1219E1220E1221E1222E1223E1224E1225E1226E1227E1228E1229E1230E1231E1232E1233E1234E1235E1236E1237E1238E1239E1240E1241E1242E1243E1244E1245E1246E1247E1248E1249E1250E1251E1252E1253E1254E1255E1256E1257E1258E1259E1260E1261E1262E1263E1264E1265E1266E1267E1268E1269E1270E1271E1272E1273E1274E1275E1276E1277E1278E1279E1280E1281E1282E1283E1284E1285E1286E1287E1288E1289E1290E1291E1292E1293E1294E1295E1296E1297E1298E1299E1300E1301E1302E1303E1304E1305E1306E1307E1308E1309E1310E1311E1312E1313E1314E1315E1316E1317E1318E1319E1320E1321E1322E1323E1324E1325E1326E1327E1328E1329E1330E1331E1332E1333E1334E1335E1336E1337E1338E1339E1340E1341E1342E1343E1344E1345E1346E1347E1348E1349E1350E1351E1352E1353E1354E1355E1356E1357E1358E1359E1360E1361E1362E1363E1364E1365E1366E1367E1368E1369E1370E1371E1372E1373E1374E1375E1376E1377E1378E1379E1380E1381E1382E1383E1384E1385E1386E1387E1388E1389E1390E1391E1392E1393E1394E1395E1396E1397E1398E1399E1400E1401E1402E1403E1404E1405E1406E1407E1408E1409E1410E1411E1412E1413E1414E1415E1416E1417E1418E1419E1420E1421E1422E1423E1424E1425E1426E1427E1428E1429E1430E1431E1432E1433E1434E1435E1436E1437E1438E1439E1440E1441E1442E1443E1444E1445E1446E1447E1448E1449E1450E1451E1452E1453E1454E1455E1456E1457E1458E1459E1460E1461E1462E1463E1464E1465E1466E1467E1468E1469E1470E1471E1472E1473E1474E1475E1476E1477E1478E1479E1480E1481E1482E1483E1484E1485E1486E1487E1488E1489E1490E1491E1492E1493E1494E1495E1496E1497E1498E1499E1500E1501E1502E1503E1504E1505E1506E1507E1508E1509E1510E1511E1512E1513E1514E1515E1516E1517E1518E1519E1520E1521E1522E1523E1524E1525E1526E1527E1528E1529E1530E1531E1532E1533E1534E1535E1536E1537E1538E1539E1540E1541E1542E1543E1544E1545E1546E1547E1548E1549E1550E1551E1552E1553E1554E1555E1556E1557E1558E1559E1560E1561E1562E1563E1564E1565E1566E1567E1568E1569E1570E1571E1572E1573E1574E1575E1576E1577E1578E1579E1580E1581E1582E1583E1584E1585E1586E1587E1588E1589E1590E1591E1592E1593E1594E1595E1596E1597E1598E1599E1600E1601E1602E1603E1604E1605E1606E1607E1608E1609E1610E1611E1612E1613E1614E1615E1616E1617E1618E1619E1620E1621E1622E1623E1624E1625E1626E1627E1628E1629E1630E1631E1632E1633E1634E1635E1636E1637E1638E1639E1640E1641E1642E1643E1644E1645E1646E1647E1648E1649E1650E1651E1652E1653E1654E1655E1656E1657E1658E1659E1660E1661E1662E1663E1664E1665E1666E1667E1668E1669E1670E1671E1672E1673E1674E1675E1676E1677E1678E1679E1680E1681E1682E1683E1684E1685E1686E1687E1688E1689E1690E1691E1692E1693E1694E1695E1696E1697E1698E1699E1700E1701E1702E1703E1704E1705E1

To : HUNAN EEXI TECHNOLOGY & SERVICE CO LTD  
Attention : 徐晓峰 Date : Jun 09, 2020

Re : Report Revision Notification

Labtest Report Number GZHT02294303 date JUN 03, 2020

Please be informed that all the content recorded in the above captioned report will be void. This captioned report is now superseded by a revised Labtest Report, Number GZHT02294303-S1, issued on Jun 09, 2020.

Thank you for your attention

Approved by:

*Jana*  
Sr. Manager

Erica Yu  
Assistant Survey

**北京大恒利源科技发展有限公司** 广州分公司  
Room 02, 1-8/F, Room 01, E101E201E301E401E501E601E701E801E901E1001E1101E1201E1301E1401E1501E1601E1701E1801E1901E2001E2101E2201E2301E2401E2501E2601E2701E2801E2901E3001E3101E3201E3301E3401E3501E3601E3701E3801E3901E4001E4101E4201E4301E4401E4501E4601E4701E4801E4901E5001E5101E5201E5301E5401E5501E5601E5701E5801E5901E6001E6101E6201E6301E6401E6501E6601E6701E6801E6901E7001E7101E7201E7301E7401E7501E7601E7701E7801E7901E8001E8101E8201E8301E8401E8501E8601E8701E8801E8901E9001E9101E9201E9301E9401E9501E9601E9701E9801E9901E10001E1001E1002E1003E1004E1005E1006E1007E1008E1009E1010E1011E1012E1013E1014E1015E1016E1017E1018E1019E1020E1021E1022E1023E1024E1025E1026E1027E1028E1029E1030E1031E1032E1033E1034E1035E1036E1037E1038E1039E1040E1041E1042E1043E1044E1045E1046E1047E1048E1049E1050E1051E1052E1053E1054E1055E1056E1057E1058E1059E1060E1061E1062E1063E1064E1065E1066E1067E1068E1069E1070E1071E1072E1073E1074E1075E1076E1077E1078E1079E1080E1081E1082E1083E1084E1085E1086E1087E1088E1089E1090E1091E1092E1093E1094E1095E1096E1097E1098E1099E1100E1101E1102E1103E1104E1105E1106E1107E1108E1109E1110E1111E1112E1113E1114E1115E1116E1117E1118E1119E1120E1121E1122E1123E1124E1125E1126E1127E1128E1129E1130E1131E1132E1133E1134E1135E1136E1137E1138E1139E1140E1141E1142E1143E1144E1145E1146E1147E1148E1149E1150E1151E1152E1153E1154E1155E1156E1157E1158E1159E1160E1161E1162E1163E1164E1165E1166E1167E1168E1169E1170E1171E1172E1173E1174E1175E1176E1177E1178E1179E1180E1181E1182E1183E1184E1185E1186E1187E1188E1189E1190E1191E1192E1193E1194E1195E1196E1197E1198E1199E1200E1201E1202E1203E1204E1205E1206E1207E1208E1209E1210E1211E1212E1213E1214E1215E1216E1217E1218E1219E1220E1221E1222E1223E1224E1225E1226E1227E1228E1229E1230E1231E1232E1233E1234E1235E1236E1237E1238E1239E1240E1241E1242E1243E1244E1245E1246E1247E1248E1249E1250E1251E1252E1253E1254E1255E1256E1257E1258E1259E1260E1261E1262E1263E1264E1265E1266E1267E1268E1269E1270E1271E1272E1273E1274E1275E1276E1277E1278E1279E1280E1281E1282E1283E1284E1285E1286E1287E1288E1289E1290E1291E1292E1293E1294E1295E1296E1297E1298E1299E1300E1301E1302E1303E1304E1305E1306E1307E1308E1309E1310E1311E1312E1313E1314E1315E1316E1317E1318E1319E1320E1321E1322E1323E1324E1325E1326E1327E1328E1329E1330E1331E1332E1333E1334E1335E1336E1337E1338E1339E1340E1341E1342E1343E1344E1345E1346E1347E1348E1349E1350E1351E1352E1353E1354E1355E1356E1357E1358E1359E1360E1361E1362E1363E1364E1365E1366E1367E1368E1369E1370E1371E1372E1373E1374E1375E1376E1377E1378E1379E1380E1381E1382E1383E1384E1385E1386E1387E1388E1389E1390E1391E1392E1393E1394E1395E1396E1397E1398E1399E1400E1401E1402E1403E1404E1405E1406E1407E1408E1409E1410E1411E1412E1413E1414E1415E1416E1417E1418E1419E1420E1421E1422E1423E1424E1425E1426E1427E1428E1429E1430E1431E1432E1433E1434E1435E1436E1437E1438E1439E1440E1441E1442E1443E1444E1445E1446E1447E1448E1449E1450E1451E1452E1453E1454E1455E1456E1457E1458E1459E1460E1461E1462E1463E1464E1465E1466E1467E1468E1469E1470E1471E1472E1473E1474E1475E1476E1477E1478E1479E1480E1481E1482E1483E1484E1485E1486E1487E1488E1489E1490E1491E1492E1493E1494E1495E1496E1497E1498E1499E1500E1501E1502E1503E1504E1505E1506E1507E1508E1509E1510E1511E1512E1513E1514E1515E1516E1517E1518E1519E1520E1521E1522E1523E1524E1525E1526E1527E1528E1529E1530E1531E1532E1533E1534E1535E1536E1537E1538E1539E1540E1541E1542E1543E1544E1545E1546E1547E1548E1549E1550E1551E1552E1553E1554E1555E1556E1557E1558E1559E1560E1561E1562E1563E1564E1565E1566E1567E1568E1569E1570E1571E1572E1573E1574E1575E1576E1577E1578E1579E1580E1581E1582E1583E1584E1585E1586E1587E1588E1589E1590E1591E1592E1593E1594E1595E1596E1597E1598E1599E1600E1601E1602E1603E1604E1605E1606E1607E1608E1609E1610E1611E1612E1613E1614E1615E1616E1617E1618E1619E1620E1621E1622E1623E1624E1625E1626E1627E1628E1629E1630E1631E1632E1633E1634E1635E1636E1637E1638E1639E1640E1641E1642E1643E1644E1645E1646E1647E1648E1649E1650E1651E1652E1653E1654E1655E1656E1657E1658E1659E1660E1661E1662E1663E1664E1665E1666E1667E1668E1669E1670E1671E1672E1673E1674E1675E1676E1677E1678E1679E1680E1681E1682E1683E1684E1685E1686E1687E1688E1689E1690E1691E1692E1693E1694E1695E1696E1697E1698E1699E1700E1701E1702E1703E1704E1705E1706E1707E1708E1709E1710E1711E1712E1713E1714



DE/000040627  
Formularnummer 00000000

Medizinprodukte - Informationssystem

Übersicht | Medizinprodukte | In-vitro-Diagnostika | Klinische Prüfungen | Adressen | Firmeninfos | Nutzereinstellungen | Kontakt

Suche | Suchergebnisse | Dokumentenübergabe | Hersteller (0)

Anzeigen Medizinprodukte (MPA) | 0 | 1

1 von 1 GEWID: MP Anzeigen (MPA) | 0 | 000000

Dokumentnummer: 00000000

Anzeige

Registrierungsdatum: 2020-05-09  
Registrierungsnummer: DE/CA65MP-238321-2357-00  
Typ der Anzeige: Bestandsanzeige  
Anzeigender nach § 25 MPG: Bestandsanzeiger  
Postleitzahl: 20537  
Straße, Haus-Nr.: Schillerstrasse 80  
Land: Hamburg  
Telefon: +49-40-2513175  
Telefax: +49-40-255726  
E-Mail: shholding@hotmail.com

Zuständige Behörde:

Code: DE/CA65  
Bezeichnung: Behörde für Gesundheit und Verbraucherschutz  
Stadt: Hamburg  
Land: Hamburg  
Ort: Hamburg  
Postleitzahl: 20537  
Straße, Haus-Nr.: Schillerstrasse 80  
Telefon: +49-40-438200  
Telefax: +49-40-47310017  
E-Mail: medienprodukte@hamburg.de  
Beauftragter: Frau Silvia Priesner  
Beauftragter Telefon: 0049-40-438200

Hersteller:

Bezeichnung: Hunan EEXI Technology&Service Co., Ltd.  
Land: China  
Ort: Hunan  
Postleitzahl: 410133  
Straße, Haus-Nr.: No. 6, North of Pingtuo road, Luyang Hi-tech Industrial Development Zone,  
Telefon: +86-731-83371666  
E-Mail: overseas01@idore.com.cn

Medizinprodukt

Produkttyp: Nichtinvasives Medizinprodukt  
Kategorie: 1  
App Software auf mobilen Geräten: Nein  
Medizinprodukt: EE01  
Allgemeine Produktbeschreibung: Disposable surgical mask  
Nomenklaturcode: Mask, 50 pieces  
Nomenklaturbeschreibung: Disposable surgical mask  
Kategorie: 10  
Kurzbeschreibung in Englisch: Disposable surgical mask is intended to be used for clinical staff and operation room doctors

Suche | Suchergebnisse | Dokumentenübergabe | Hersteller (0)

Serviceinfos:

- Hersteller
- Anleitung für Anwender und Anwender Prüfungen

Anzeige 1  
(zu § 4 Abs. 1 Nr. 1 DVGW)  
Formularnummer 00000000

Allgemeine Anzeigepflicht nach §§ 25 und 30 Abs. 2 MPG  
General Obligation to Notify pursuant to §§ 25 and 30 (2) Medical Devices Act, MPG

Formblatt für Medizinprodukte, außer In-vitro-Diagnostika  
Form for Medical Devices except In Vitro Diagnostic Medical Devices

Zuständige Behörde / Competent authority

Code: DE/CA65  
Bezeichnung / Name: Behörde für Gesundheit und Verbraucherschutz, Referat V43  
Land / Federal state: Deutschland  
Ort / City: Hamburg  
Postleitzahl / Postal code: 20539  
Straße, Haus-Nr. / Street, house no.: Billstraße 80  
Telefon / Phone: +49-40-438200  
Telefax / Fax: +49-40-47310017  
E-Mail / E-mail: medizinprodukte@bvg.hamburg.de

Anzeige / Notification

Registrierdatum bei der zuständigen Behörde  
Registration date at competent authority: 09.04.2020  
Registrierungsnummer / Registration number: DE/CA65MP-238321-2357-00

Typ der Anzeige / Notification type  
S: Erstanzeige / Initial notification  
E: Änderungsanzeige / Notification of change  
W: Widerrufsanzeige / Notification of withdrawal

Frühere Registrierungsnummer bei Änderungs- und Widerrufsanzeige  
Previous registration number if notification has been changed or withdrawn

Anzeigender nach § 25 MPG / Reporter pursuant to § 25 Medical Devices Act, MPG

E: Hersteller / Manufacturer  
S: Bevollmächtigter / Authorised Representative  
E: Einführer / Importer  
E: Verantwortlicher für das Zusammenstellen von Systemen oder Behandlungseinheiten nach § 10 Abs. 1 und 2 MPG / Assembler of systems or procedure packs pursuant to § 10 (1) and (2) Medical Devices Act, MPG  
E: Betrieb oder Einrichtung (aufbereiten) nach § 25 Abs. 1 MPG i. V. m. § 4 Abs. 2 MPBetreibV  
E: Betrieb oder Einrichtung (sterilisieren) nach § 25 Abs. 2 i. V. m. § 10 Abs. 3 MPG  
E: Institution (sterilizing) pursuant to § 25 (2) in connection with § 10 (3) Medical Devices Act, MPG

Anzeige 1  
(zu § 4 Abs. 1 Nr. 1 DVGW)  
Formularnummer 00000000

Anzeigender / Reporting organisation (person)

Code: DE/000040627  
Bezeichnung / Name: Shanghai International Holding Corporation GmbH (Europe)  
Land / Federal state: Deutschland  
Ort / City: Hamburg  
Postleitzahl / Postal code: 20537  
Straße, Haus-Nr. / Street, house no.: Effestraße 80  
Telefon / Phone: +49-40-2513175  
Telefax / Fax: +49-40-255726  
E-Mail / E-mail: shholding@hotmail.com

Hersteller / Manufacturer

Bezeichnung / Name: Hunan EEXI Technology&Service Co., Ltd.  
Land / Federal state: China  
Ort / City: Hunan  
Postleitzahl / Postal code: 410133  
Straße, Haus-Nr. / Street, house no.: No. 6, North of Pingtuo road, Luyang Hi-tech Industrial Development Zone,  
Telefon / Phone: +86-731-83371666  
Telefax / Fax: +86-731-83371666  
E-Mail / E-mail: overseas01@idore.com.cn

Sicherheitsbeauftragter für Medizinprodukte nach § 30 Abs. 2 MPG (9)  
Safety officer for medical devices pursuant to § 30 (2) Medical Devices Act, MPG

Bezeichnung / Name: Liang Jin  
Land / Federal state: Deutschland  
Ort / City: Hamburg  
Postleitzahl / Postal code: 20537  
Straße, Haus-Nr. / Street, house no.: Effestraße 80  
Telefon / Phone: +49-40-2513175  
Telefax / Fax: +49-40-255726  
E-Mail / E-mail: shholding@hotmail.com

P10

Ort	Hamburg	Datum	2020-03-29
City		Date	
		Name	Min Fang

Unterschrift  
Signature



## Skin Sensitization Test Guinea Pig Maximization Final Report

Article Name: Disposable medical face mask

Report Number: CSTBB20030201

Method Standard: ISO 10993-10: 2010

### Sponsor

Hunan FEXI Technology&Service  
Co., Ltd.

No.6 north of Pingtiao road, Liuyang Hi-tech  
industrial development zone, Hunan, China

### Test Facility

CCIC Huatongwei international inspection  
(Suzhou) Co., Ltd

Room 101, Building G, Raoshui Road 388, Suzhou,  
Jiangsu, China

CCIC Huatongwei international inspection (Suzhou) Co., Ltd

Address: Room 101, Building G, Raoshui Road 388, Suzhou, Jiangsu, China, 512123 Tel: 0512-87657288 Fax: 0512-87657288  
Page 1 of 11

Report No.: CSTBB20030201

### Notices

1. Please apply for rechecking within 15 days of receiving the report if there is any objection.
2. Any erasure or without special testing seal renders the report null and void.
3. The report is only valid when signed by the persons who edited, checked and approved it.
4. The report is only responsible for the test results of the tested samples.
5. The report shall not be reproduced except in full without the written approval of the company.

Page 3 of 11

Report No.: CSTBB20030201

### CONTENTS

Notices.....	3
Abstract.....	4
Study Verification and Signature.....	5
1.0 Purpose.....	6
2.0 Reference.....	6
3.0 Test and control articles.....	6
4.0 Identification of test system.....	7
5.0 Animal Management.....	7
6.0 Equipment and reagents.....	7
7.0 Experiment design.....	8
8.0 The results observed.....	9
9.0 Evaluation criteria.....	9
10.0 Results of the test.....	9
11.0 Conclusion.....	9
12.0 Record.....	9
13.0 Confidentiality Agreement.....	9

Page 2 of 11

Report No.: CSTBB20030201

### Abstract

In this study, we took guinea pigs to observe the skin sensitization of the test article according to ISO 10993-10: 2010.

The test article were extracted in Constant Temperature Vibrator at 50 °C, 60 rpm for 72 h by 0.9 % Sodium Chloride Injection and Sesame Oil. Mix 50:50 (by volume) stable emulsion of Freund's complete adjuvant with selected solvent. Intradermal induction and topical induction were operated in the clipped intrascapular region of each animal. After the topical induction phase was completed on day 14, all test and control animals were challenged with the test sample. The erythema and edema of the challenge site were observed to test the sensitization response of the test article. According to the Magnusson and Klignman scales, the response to erythema and edema at each application site of the skin was described and scored 24 hours and 48 hours after the challenge phase.


The results showed that the guinea pigs in the negative control group (0.9 % Sodium Chloride Injection, Sesame Oil) retained a normal appearance throughout the test and showed no skin irritants. A severe skin reactions for erythema and oedema were shown in the positive control group (DNCEB). While in test article group, the response of skin on testing side did not exceed that on the control side. The skin reactions for erythema and oedema were not observed in test article group. The data of each group met the acceptance criteria, and the results of this test were considered valid.

Based on the above results, it can be concluded that under the experimental conditions, the test article Disposable medical face mask has no potential skin sensitization on guinea pigs in the extraction method.

Page 4 of 11

P12

## Study Verification and Signature



Protocol Number SST20030603BB  
 Protocol Effective Date 2020-03-13  
 Technical Initiation Date 2020-03-20  
 Technical Completion Date 2020-04-17  
 Final Report Completion Date 2020-04-18

Personnel Date Completed 2020-04-18  
 Approved Date Completed 2020-04-18  
 Supervisory Date Completed 2020-04-18  
 Test Facility Manager  
 Huatongwei International Inspection (Shanghai) Co., Ltd.

## 1.0 Purpose

The test was designed to evaluate the potential of a test article to cause skin sensitization. The test is used as a procedure for screening of contact allergens in guinea pigs and extrapolating the results to humans, but it does not establish the actual risk of sensitization.

## 2.0 Reference

Biological evaluation of medical devices Part 10: Tests for irritation and skin sensitization (ISO 10993-10:2010)  
 Biological evaluation of medical devices-Part 12: Sample preparation and reference materials (ISO 10993-12:2012)  
 Biological evaluation of medical devices-Part 2: Animal welfare requirements (ISO 10993-2:2006)

## 3.0 Test and control articles

Groups	Test article	Negative Control Article(Polar)	Negative Control Article(Non-Polar)	Positive Control
Name	Disposable medical face mask	Sodium Chloride Injection (SC)	Sesame Oil (SO)	2, 4-Dinitrochlorobenzene (DNCB)
Manufacture	Hunan EEXI Technology&Service Co.,Ltd.	Shijiazhuang No.4 Pharmaceutical	Jiangxi xinren natural vegetable oil co., Ltd.	TOKYO CHEMICAL INDUSTRY CO., LTD
Sterilization state	No	/	/	/
Size	17.5cm*9.5cm	500 ml	25 kg	25 g
Model	YX001	/	/	/
Lot Batch#	Not provided	1912121907	181120	H2UKD-DM
Test Article Material	PP non-woven, Melblown filtration layer, earloop and nose clip	/	/	/
Physical State	Solid	Liquid	Liquid	Solid
Color	Not provided	Colorless	Light yellow	Light yellow
Package material	PE poly bag and paper box	/	/	/
Concentration	/	0.9 %	/	Induction Concentration: 0.5 % Challenge Concentration: 1.0 % Dissolved in ethanol
Total	Not provided	/	/	/

Surface/Weight				
Storage Condition	Room Temp.	Room Temp.	Room Temp.	Room Temp.

The information about the test article was supplied by the sponsor wherever applicable.

## 4.0 Identification of test system

## 4.1 Test animal

Species: Hartley Guinea Pig (Cavia Porcellus)  
 Number: 30 (20 Test +10 Control)  
 Sex: 15 ♀, 15 ♂  
 Initial body weight: 301.5-313.5 g  
 Health status: Healthy, not previously used in other experimental procedures  
 Animal identification: Ear tag  
 Cages: Plastic cage  
 Acclimation Period: 7 days under the same conditions as for the actual test

## 4.2 Justification of test system

The albino guinea pig has been used historically for sensitization studies (Magnusson and Kligman, 1970). The guinea pig is believed to be the most sensitive animal model for this type of study. DNCB is the positive control article recommended in the test instructions. To ensure the sensitivity of the experimental system, the positive control article should be verified every three months.

## 5.0 Animal Management

Animal purchase: Wuxi hengtai experimental animal breeding co. LTD SCXK (SU) 2015-0004  
 Bedding: Corncob Jiangsu Xietong Pharmaceutical Bio-engineering Co., Ltd.  
 Feed: Guinea pigs were fed with full-price pellets Jiangsu Xietong Pharmaceutical Bio-engineering Co., Ltd.  
 Water: Drinking water met the Standards for Drinking Water Quality GB 5749-2006  
 Animal room temperature: 18-26 °C  
 Animal room relative humidity: 30 %-70 %  
 Lights: 12 hours light/dark cycle, full-spectrum lighting  
 Personnel: Associates involved were appropriately qualified and trained  
 Selection: Only healthy, previously unused animals were selected  
 There were no known contaminants present in the feed, water, or bedding expected to interfere with the test data.

## 6.0 Equipment and reagents

## 6.1 Instruments

Constant Temperature Vibrator (SHB007, calibration data: 2020/3/16), Autoclave (SHB026, calibration data: 2020/3/16), Electronic scale (SHB017, calibration data: 2020/3/16)

## 6.2 Reagents

Freund's adjuvant Complete liquid (SIGMA, Lot No: SLBR3877V), Sodium dodecyl sulfate (SDS SIGMA, Lot No: SLBH2304V)

## 7.0 Experiment design

## 7.1 Sample preparation

The extracts of test article will be prepared according to the following steps:

Aspheric Sampling			Extraction in sterile vessels			
Sampling Manner	Actually sampling	Ratio	Reagent	Temperature	Time	pH
Whole	570.2 cm <sup>2</sup>	6cm <sup>2</sup> : 1 ml	SC 95.0 ml SO 1.95.0 ml	50 °C	72 h	5.5

Both inducements and excitations were prepared by the number of times. The state of the leaching solution did not change visually after the leaching was advanced. After extraction, the samples were stored at room temperature for no more than 24 h. The extraction solution is clear, and the pH value has not been adjusted, filtered, centrifuged, diluted and other processes. The control solution was prepared under the same conditions.

## 7.2 Test method

## 7.2.1 Intradermal induction phase

A pair of 0.1ml intradermal injections was made for each of the following, into each animal, at the injection sites (A, B and C) as shown in Figure 1 in the clipped intrascapular region.

Site A: A 50:50 (volume ratio) stable emulsion of Freund's complete adjuvant mixed with the chosen solvent.  
 Site B: The test sample (undiluted extract); the control animals were injected with the solvent alone.

Site C: The test sample at the concentration used at site B, emulsified in a 50:50 volume ratio stable emulsion of Freund's complete adjuvant and the solvent (50 %); the control animals were injected with an emulsion of the blank liquid with adjuvant.




Figure 1 Location of intradermal injection sites

## 7.2.2 Topical induction phase

The maximum concentration that can be achieved in Intradermal induction phase I did not produce irritation, animals are pretreated with 10% sodium dodecyl sulfate 24(±2) hours before the topical induction application.

At 7 d after completion of the intradermal induction phase, administer test article extract by topical application to the intrascapular region of each animal, using a patch of area approximately 8 cm<sup>2</sup> (absorbent gauze), so as to cover the intradermal injection sites. Secure the patches with an occlusive dressing. Remove the dressings and patches after (48±2) h.

Treat the control animals similarly, using the blank liquid alone.

## 7.2.3 Challenge phase



At 14d after completion of the topical induction phase, challenge all test and control animals with the test sample. Absorbent gauzes (2.5 cmx2.5 cm) were soaked respectively with test article and control article. Apply the test article extract and control article topically to two sites that were not treated during the induction stage. Secure with an occlusive dressing. Remove the dressings and patches after (24±2) h.

#### 8.0 The results observed

The day after challenge exposure, the patch will be removed and the area cleaned gently with gauze if necessary. The site will be wiped gently with a 0.9 % saline soaked gauze sponge prior to each scoring period. The challenge sites will be observed for signs of irritation and sensitization reactions, as indicated by erythema and edema. If necessary, the fur will be shaved or clipped in advance for the convenience of dermal score.

Daily challenge observation scores will be recorded approximately 24, and 48 hours after patch removal in accordance with the following classification system for skin reactions:

Table 1 Magnusson and Kilgman scale

Patch test reaction	Grading scale
No visible change	0
Discrete or patchy erythema	1
Moderate and confluent erythema	2
Intense erythema and/or swelling	3

#### 9.0 Evaluation criteria

Magnusson and Kilgman grades of 1 or greater in the test group generally indicate sensitization, provided grades of less than 1 are seen in control animals.

If grades of 1 or greater are noted in control animals, then the reactions of test animals which exceed the most severe reaction in control animals are presumed to be due to sensitization.

If the response is equivocal, rechallenge is recommended to confirm the results from the first challenge.

The outcome of the test is presented as the frequency of positive challenge results in test and control animals.

#### 10.0 Results of the test

All animals were survived and no abnormal signs were observed during the study. Individual results of dermal scoring for the challenge appear in Table 2.

#### 11.0 Conclusion

The test article showed no evidence of causing delayed dermal contact sensitization in the guinea pig. Results and conclusions apply only to the test article tested. Any extrapolation of these data to other articles is the sponsor's responsibility.

#### 12.0 Record

All raw data pertaining to this study and a copy of the final report are retained in designated Huatongwei archive.

#### 13.0 Confidentiality Agreement

Statements of confidentiality were as agreed upon prior to study initiation.

Table 2 Guinea pig Sensitization Dermal Reactions

Group	No.	Pretest weight(g)	Finished weight(g)	The Challenge patch was removed 24h later		The Challenge patch was removed 48h later		Positive rate	
				Erythema	Swelling	Erythema	Swelling		
SC	Test	1	312.8	350.1	0	0	0	0	0%
		2	310.8	349.8	0	0	0	0	
		3	311.7	346.8	0	0	0	0	
		4	311.8	346.5	0	0	0	0	
		5	313.5	350.2	0	0	0	0	
		6	308.2	349.3	0	0	0	0	
		7	310.3	352.1	0	0	0	0	
		8	312.2	350.6	0	0	0	0	
		9	307.9	346.8	0	0	0	0	
		10	311.9	347.9	0	0	0	0	
	Control	11	306.8	346.4	0	0	0	0	—
		12	311.6	351.3	0	0	0	0	
		13	307.1	347.9	0	0	0	0	
		14	311.5	356.1	0	0	0	0	
		15	305.9	349.3	0	0	0	0	
SO	Test	16	301.5	346.6	0	0	0	0	0%
		17	311.9	351.6	0	0	0	0	
		18	307.3	348.6	0	0	0	0	
		19	306.0	344.4	0	0	0	0	
		20	302.2	340.5	0	0	0	0	
		21	307.3	348.1	0	0	0	0	
		22	301.9	344.9	0	0	0	0	
		23	306.8	346.1	0	0	0	0	
		24	312.0	356.1	0	0	0	0	
		25	307.3	348.3	0	0	0	0	
	Control	26	312.3	355.7	0	0	0	0	—
		27	311.6	354.5	0	0	0	0	
		28	309.8	354.0	0	0	0	0	
		29	306.3	347.9	0	0	0	0	
		30	307.4	345.9	0	0	0	0	

Group	No.	Pretest weight(g)	Finished weight(g)	The Challenge patch was removed 24 h later		The Challenge patch was removed 48 h later		Positive rate
				Erythema	Swelling	Erythema	Swelling	
Test	1	311.2	351.7	2	0	2	0	100%
	2	315.6	360.2	1	0	1	0	
	3	317.2	352.7	0	0	1	0	
	4	312.8	353.9	1	0	1	0	
	5	306.9	341.1	2	0	2	0	
	6	312.2	350.6	0	0	1	0	
	7	317.1	352.2	1	0	2	0	
	8	306.8	350.0	1	0	1	0	
	9	316.5	348.7	1	0	2	0	
	10	317.6	350.2	2	0	2	0	
Control	11	320.5	364.5	0	0	0	0	—
	12	307.6	350.2	0	0	0	0	
	13	306.9	345.1	0	0	0	0	
	14	310.0	352.3	0	0	0	0	
	15	315.8	346.6	0	0	0	0	

Note: The positive control was CSTBB20010001P1(Finish date: 2020-02-07).



## Skin Irritation Test

### Extraction Method

### Final Report

Article Name: Disposable medical face mask

Report Number: CSTBB20030203

Method Standard: ISO 10993-10: 2010

#### Sponsor

Hunan EEXI Technology&Service Co.,Ltd.  
No.6 north of Pingtong road, Liuyang Hi-tech industrial development zone, Hunan, China

#### Test Facility

CCIC Huatongwei international inspection (Suzhou) Co., Ltd.  
Room 101, Building G, Raoxhai Road 388, Suzhou, Jiangsu, China

## CONTENTS

Notices.....	3
Abstract.....	4
Study Verification and Signature.....	5
1.0 Purpose.....	6
2.0 Reference.....	6
3.0 Test and control articles.....	6
4.0 Identification of test system.....	6
5.0 Animal Management.....	7
6.0 Equipment and reagents.....	7
7.0 Experiment design.....	7
8.0 The results observed.....	8
9.0 Evaluation criteria.....	9
10.0 Results of the test.....	9
11.0 Conclusion.....	9
12.0 Record.....	9
13.0 Confidentiality Agreement.....	9

## Abstract

In this study, we took New Zealand white Rabbits to observe the skin irritation of the test article according to ISO10993-10:2010.

The test article were extracted in Constant Temperature Vibrator at 50 °C, 60 rpm for 72 h by 0.9 % Sodium Chloride Injection and Sesame Oil. Apply 0.5 ml extracts of test article or control to 2.5 cm×2.5 cm absorbent gauze patches, and then apply the patch soaked with the extract of test article or control directly to the skin on each side of each rabbit, and then wrap the application sites with a bandage for a minimum of 4 h. At the end of the contact time, remove the dressing. The describe and score the skin reaction for erythema and oedema for each application site at each time interval. Record the appearance of each application site at (1±0.1) h, (24±2) h, (48±2) h and (72±2) h following removal of the patches.


The results showed that the rabbits in the negative control group (0.9 % Sodium Chloride Injection, Sesame Oil) retained a normal appearance throughout the test and showed no skin irritants. A severe skin reactions for erythema and oedema were shown in the positive control group (SDS). While in test article group, the response of skin on testing side did not exceed that on the control side. The skin reactions for erythema and oedema were not observed in test article group. The data of each group met the acceptance criteria, and the results of this test were considered valid.

Based on the above results, it can be concluded that under the experimental conditions, the test article Disposable medical face mask has no potential skin irritation on rabbit in the extraction method.

## Notices

1. Please apply for rechecking within 15 days of receiving the report if there is any objection.
2. Any erasure or without special testing seal renders the report null and void.
3. The report is only valid when signed by the persons who edited, checked and approved it.
4. The report is only responsible for the test results of the tested samples.
5. The report shall not be reproduced except in full without the written approval of the company.

## Study Verification and Signature



Protocol Number	SST2003006601BB
Protocol Effective Date	2020-03-13
Technical Initiation Date	2020-03-20
Technical Completion Date	2020-03-27
Final Report Completion Date	2020-04-18

Personnel		
		Date Completed

Approved		
	Study Director	Date Completed

Supervisory		
	Test Facility Manager	Date Completed

Huatongyi International Inspection (Shenzhen) Co., Ltd.

P15

**1.0 Purpose**

To evaluate the potential skin irritation caused by test article contact with the skin surface of rabbits and extrapolating the results to humans, but it does not establish the actual risk of irritation.

**2.0 Reference**

Biological evaluation of medical devices Part 10: Tests for irritation and skin sensitization (ISO 10993-10: 2010)

Biological evaluation of medical devices-Part 12: Sample preparation and reference materials (ISO 10993-12:2012)

Biological evaluation of medical devices-Part 2: Animal welfare requirements (ISO 10993-2:2006)

**3.0 Test and control articles**

Groups	Test article	Negative Control Article(Polar)	Negative Control Article(Non-Polar)	Positive Control
Name	Disposable medical face mask	Sodium Chloride Injection (SC)	Sesame Oil (SO)	10 % sodium dodecyl sulfate (SDS)
Manufacture	Hunan EEXI Technology&Service Co.,Ltd.	Shijiazhuang No.4 Pharmaceutical	Jiangxi xinsum natural vegetable oil co., Ltd.	SIGMA
Sterilization state	No	/	/	/
Size	17.5cm*9.5cm	500 ml	25 kg	25 g
Model	YX001	/	/	/
Lot Batch#	Not provided	1912121907	181120	SLBL2304V
Test Article Material	PP non-woven, Melblown filtration layer, earloop and nose clip	/	/	/
Physical State	Solid	Liquid	Liquid	Solid
Color	Not provided	Colorless	Light yellow	Colorless
Package material	PE poly bag and paper box	/	/	/
Concentration	/	0.9 %	/	10 %
Total Surface/Weight	Not provided	/	/	/
Storage Condition	Room Temp.	Room Temp.	Room Temp.	Room Temp.

The information about the test article was supplied by the sponsor wherever applicable.

**4.0 Identification of test system****4.1 Test animal**

Species: New Zealand white Rabbit

Number: 6

Sex: 3 ♀, 3 ♂

Weight: 2.04-2.17 kg

Health status: Healthy, not previously used in other experimental procedures

Animal identification: Ear tattoo

Cages: Stainless steel cage

Acclimation Period: 7 days under the same conditions as for the actual test

**4.2 Justification of test system**

The rabbit is specified as an appropriate animal model for evaluating potential skin irritants by the current testing standards. Positive control 10% sodium dodecyl sulfate has been substantiated at ITW with this method.

**5.0 Animal Management**

Animal purchase: Wuxi hengtai experimental animal breeding co. LTD SCXK (SU) 2015-0004

Breeding: /

Feed: Experimental rabbits were fed a maintenance diet, Wuxi hengtai experimental animal breeding co. LTD

Water: Drinking water met the Standards for Drinking Water Quality GB 5749-2006

Animal room temperature: 18-26 °C

Animal room relative humidity: 30 %-70 %

Light: 12 hours light/dark cycle, full-spectrum lighting

Personnel: Associates involved were appropriately qualified and trained

Selection: Only healthy, previously unused animals were selected

There were no known contaminants present in the feed, water, or bedding expected to interfere with the test data

**6.0 Equipment and reagents****6.1 Instruments**

Constant Temperature Vibrator (SHB007, calibration data: 2020/3/16), Autoclave (SHB026, calibration data: 2020/3/16), Electronic scale (SHB017, calibration data: 2020/3/16)

**7.0 Experiment design****7.1 Sample preparation**

The extracts of test article will be prepared according to the following steps:


Aseptic Sampling			Extraction in sterile vessels			
Sampling Manner	Actually sampling	Ratio	Reagent	Temperature	Time	pH
Whole	570.2 cm <sup>3</sup>	6cm <sup>3</sup> : 1 ml	SC 95.0 ml SO 95.0 ml	50 °C	72 h	5.5

The state of the leaching solution did not change visually after the leaching was advanced. The excretions were clear, and the pH value has not been adjusted, filtered, centrifuged, diluted and other processes, before doing stored at room temperature no more than 24 h. The control solution was prepared under the same conditions

**7.2 Test method**

Use the rabbits with healthy intact skin. Fur was generally clipped within 24 h period before testing on the backs of the rabbits, a sufficient distance on both sides of the spine for application and observation of all test sites (approximately 10\*15 cm).

Apply 0.5 ml extract (s) of test article or control to 2.5 cm\*2.5 cm absorbent gauze patches, and then apply the patch soaked with the extract of test article or control directly to the skin on each side of each rabbit as shown in Figure 1, and then wrap the application sites with a bandage (semi-occlusive or occlusive) for a minimum of 4h. At the end of the contact time, remove the dressing.



1- Cranial end, 2- Test site, 3- Control site, 4- Clipped dorsal region, 5- Caudal end

Figure1 Location of skin application sites

**8.0 The results observed**

The Describe and score the skin reaction for erythema and oedema according to the scoring system given in Table 1 for each application site at each time interval. Record the appearance of each application site at (1±0.1) h, (24±2) h, (48±2) h and (72±2) h following removal of the patches.

Table 1 Classification System for Skin Reaction

Erythema and Eschar Formation:	Numerical Grading
No erythema	0
Very slight erythema (barely perceptible)	1
Well-defined erythema	2
Moderate erythema	3
Severe erythema (beet redness) to eschar formation preventing grading of erythema	4
Edema Formation:	
No edema	0
Very slight edema (barely perceptible)	1
Well-defined edema (edges of area well-defined by definite raising)	2
Moderate edema (raised approximately 1mm)	3
Severe edema (raised more than 1mm and extending beyond exposure area)	4
Maximal possible score for irritation	8

Irritation Response Categories in the Rabbit

Response Category	Mean score
Negligible	0 to 0.4
Slight	0.5 to 1.9
Moderate	2 to 4.9
Severe	5 to 8

**9.0 Evaluation criteria**

Use only (24±2) h, (48±2) h and (72±2) h observations for calculation.

After the 72 h grading, all erythema grades plus oedema grades (24±2) h, (48±2) h and (72±2) h were totalled separately for each test article and blank for each animal. The primary irritation score for an animal was calculated by dividing the sum of all the scores by 6 (two test/observation sites, three time points).

To obtain the primary irritation index for the test article, add all the primary irritation scores of the individual animals and divide by the number of animals.

When blank or negative control was used, calculate the primary irritation score for the controls and subtract that score from the score using the test material to obtain the primary irritation score.

**10.0 Results of the test**

All animals were survived and no abnormal signs were observed during the study. According to what observed, the response of skin on testing side did not exceed that on the control side. Thus, the primary irritation index for the test article was calculated to be 0. See table 2.

**11.0 Conclusion**

The test result showed that the response of the test article extract was categorized as negligible under the test condition.

**12.0 Record**

All raw data pertaining to this study and a copy of the final report are retained in designated Huatongwei archive.

**13.0 Confidentiality Agreement**

Statements of confidentiality were as agreed upon prior to study initiation.



Table 2 Skin irritation response observation

Reagent	Rabbit No	Pretest weight(g)	Finished weight(g)	Group	Reaction	Interval (hours): score-left/right			
						1h	24h	48h	72h
SC	1	2.10	2.18	Test Article	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
	2	2.09	2.16	Test Article	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
	3	2.17	2.24	Test Article	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
Primary irritation index						0			
SO	4	2.04	2.14	Test Article	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
	5	2.11	2.20	Test Article	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
	6	2.13	2.21	Test Article	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
				Negative Control	Erythema	0/0	0/0	0/0	0/0
					Oedema	0/0	0/0	0/0	0/0
Primary irritation index						0			

Table 3 Positive control

Rabbit No	Group	Reaction	Interval (hours): score-left site/right site			
			1h	24h	48h	72h
1	Positive control	Erythema	0/0	1/2	2/3	3/3
		Oedema	0/0	2/1	2/2	3/3
	Negative Control	Erythema	0/0	0/0	0/0	0/0
		Oedema	0/0	0/0	0/0	0/0
2	Positive control	Erythema	0/1	2/1	3/3	4/5
		Oedema	1/0	2/2	3/3	3/4
	Negative Control	Erythema	0/0	0/0	0/0	0/0
		Oedema	0/0	0/0	0/0	0/0
3	Positive control	Erythema	1/0	1/2	3/3	4/5
		Oedema	0/1	2/1	3/4	3/4
	Negative Control	Erythema	0/0	0/0	0/0	0/0
		Oedema	0/0	0/0	0/0	0/0
Primary irritation index			5.2			

Positive control performed once every six months see CSTBB2003001P3(Finish date: 2020-02-21)



## In Vitro Cytotoxicity Test

### MTT Method

#### Final Report

Article Name: Disposable medical face mask

Report Number: CSTBB20030204

Method Standard: ISO 10993-5: 2009

#### Sponsor

Human EEXI Technology&amp;Service Co.,Ltd.

No.6 north of Pingtuo road, Luyang Hi-tech industrial development zone, Hunan, China

#### Test Facility

CCIC Huatongwei international inspection (Suzhou) Co., Ltd

Room 101, Building G, Raoshui Road 388, Suzhou, Jiangsu, China

CCIC Huatongwei international inspection (Suzhou) Co., Ltd

Address: Room 101, Building G, Raoshui Road 388, Suzhou, Jiangsu, China, 512123 Tel: 0512-87657288 Fax: 0512-87657288

## CONTENTS

Notices.....	3
Abstract.....	4
Study Verification and Signature.....	5
1.0 Purpose.....	6
2.0 Reference.....	6
3.0 Test and control articles.....	6
4.0 Identification and justification of test system.....	6
5.0 Equipment and reagents.....	7
6.0 Experiment design and dose.....	7
7.0 Statistical method.....	8
8.0 Evaluation criteria.....	8
9.0 Results of the test.....	8
10.0 Conclusion.....	9
11.0 Record.....	9
12.0 Confidentiality Agreement.....	9

## Notices

1. Please apply for rechecking within 15 days of receiving the report if there is any objection.
2. Any erasure or without special testing seal renders the report null and void.
3. The report is only valid when signed by the persons who edited, checked and approved it.
4. The report is only responsible for the test results of the tested samples.
5. The report shall not be reproduced except in full without the written approval of the company.

## Abstract

In this study, mammalian L-929 cells were cultured in vitro according to ISO 10993-5:2009 to test the potential cytotoxicity of the test article.

The test articles and the control material were separately placed in MEM medium containing 10% fetal bovine serum, and extracted in a 37 °C incubator for 24 hours. After the end of the extraction, the cell culture medium in the 96-well plate (10<sup>4</sup> cells/well) cultured for 24 hours was removed and replaced with the corresponding extract, cultured in 37 °C, 5% CO<sub>2</sub>, >90% humidity for 24 hours. After the culture, the morphology and cell lysis of the cells were observed under the microscope, and the cytotoxicity of the test samples was determined by MTT assay.

The results showed that the cells in the blank control group and the negative control group (high density polyethylene) were well-formed throughout the experiment and showed no cytotoxic reaction. A severe cytotoxic response was shown in the positive control group (ZDEC). The 100% concentration of the test extract retained a normal appearance after 24 hours of incubation, and the cell viability was 78.1%. The data of each group met the acceptance criteria, and the results of this test were valid.

Based on the above results, it can be concluded that under the experimental conditions, the test article Disposable medical face mask have no potential toxicity to L-929 in the MTT method.

## Study Verification and Signature



Protocol Number SST2003006602BB  
Protocol Effective Date 2020-03-23  
Technical Initiation Date 2020-03-23  
Technical Completion Date 2020-03-25  
Final Report Completion Date 2020-04-18

Personnel  2020-04-18  
Date Completed

Approved  2020-04-18  
Study Director Date Completed

Supervisory  2020-04-18  
Test Facility Manager Date Completed

Huatongwei International Inspection Co., Ltd.

## 1.0 Purpose

The purpose of the test is to determine the potential cytotoxicity toxicity of a mammalian cell culture (mouse fibroblast L-929 cells) in response to the test article.

## 2.0 Reference

Biological evaluation of medical devices-Part 5: Tests for In Vitro Cytotoxicity (ISO 10993-5: 2009)

Biological evaluation of medical devices-Part 12: Sample preparation and reference materials (ISO 10993-12: 2012)

## 3.0 Test and control articles

Groups	Test article	Negative Control Article	Positive Control Article	Blank Control
Name	Disposable medical face mask	High Density Polyethylene Film	ZDEC	MEM medium, with addition 10% FBS
Manufacture	Huano EEXI Technology&Service Co.,Ltd.	Hatano Research Institute, FDSC	Sigma-Aldrich	Hyclone
Size	17.5cm*9.5cm	3 cm*10 cm (5 sheets)	25 g	500 ml
Model	YX001	/	/	/
Lot Batch#	Not provided	C-161	BCB06847V	AE29441978
Test Article Material	PP non-woven, Melblown filtration layer, earloop and nose clip	/	/	/
Physical State	Solid	Solid	Solid	Liquid
Color	Not provided	White	White	Pink
Packaging Material	PE poly bag and paper box	/	/	/
Sterilized or Not	No	No	No	Yes
Concentration	/	/	0.1%	/
Total Surface	Not provided	/	/	/
Storage Condition	Room Temp.	Room Temp.	Room Temp.	4°C

Note: The information about the test article was supplied by the sponsor wherever applicable.

## 4.0 Identification and justification of test system

L-929 mouse fibroblast cells obtained from American Type Culture Collection (ATCC).

L-929 cells have been used for cytotoxicity studies because they demonstrate sensitivity to extractable cytotoxic articles. Also, the test article is extracted and administered in vitro to mouse fibroblast L929 cells through a solvent compatible with the test system, which is the optimal route of administration available in this test system as recommended in ISO 10993-5.

## 5.0 Equipment and reagents

### 5.1 Instruments

Vertical pressure steam sterilizer (SHB026), CO<sub>2</sub> Incubator (SHB002), Steel Straight Scale (SHB076), Electronic Balance (SHB016), Clean bench (SHB014), Multiskan Spectrum Microplate Spectrophotometer (SHB003), Bench type low speed centrifuge (SHB022), Inverted microscope (SHB005)

### 5.2 Reagents

MEM (Hyclone, AE29441978), FBS (Clark, JC55116), Penicillin-Streptomycin (Gibco, 2145453), Trypsin (Gibco, 2040880), PBS (Hyclone, AE29451445), MTT (3-(4,5-Dimethylthiazol-2-yl)-2,5-diphenyltetrazolium bromide) (Sigma, MKBG2038V), Isopropyl alcohol (Macklin, C10394667)

## 6.0 Experiment design and dose

### 6.1 Sample preparation

According to the table below, aseptic extraction of the test article scaled and incubated in MEM medium (10% FBS) at 37 °C, 5% CO<sub>2</sub> and 60 rpm for 24 hours.

Groups	Sampling		Sterilization	Aseptic Extraction In Inert Container				Final Extract
	Sampling Method	Actually sampling		Ratio	Extracts	Condition	pH	
Test article	Whole	570.2 cm <sup>3</sup>	EO	6 cm <sup>3</sup> : 1 ml	95.0 ml	37 °C 24 h	7.4	Clear
Negative Control	Random	60 cm <sup>3</sup>	UV	3 cm <sup>3</sup> : 1 ml	20.0 ml	37 °C 24 h	7.4	Clear
Positive Control	Random	0.02 g	Filter	0.1 g: 100 ml	20.0 ml	37 °C 24 h	7.4	Clear
Blank Control	/	/	/	/	10.0 ml	37 °C 24 h	7.4	Clear

The changes of the leaching solution was observed after extraction. No particulates or color changes were observed in pre- and post-extraction, and immediately be used in the follow-up experiment. The color and pH of the extraction solution did not change before and after use, and the pH value was 7.4 after leaching.

### 6.2 Test method

Aseptic procedures were used for handling cell cultures. L-929 cells were cultured in MEM medium (10% FBS, 1% Penicillin-Streptomycin solution) at 37 °C in a humidified atmosphere of 5% CO<sub>2</sub>, then digested by 0.25% trypsin containing EDTA to get single cell suspension.  $1 \times 10^5$  cells/ml suspension were obtained by centrifuging (1000 rpm, 5 min) and re-dispersing in MEM medium.

The suspended cells were dispensed at 100 µl per well in 96-well plate, and cultured in cell incubator (5% CO<sub>2</sub>, 37 °C, >90% humidity). Cell morphology was evaluated to verify that the monolayer was satisfactory.

After the cells grew to about 70% and form a monolayer, original culture medium was discarded. The 96-well plates were then treated with 100 µl of extract of test article (100%, 75%, 50%, 25%), control article, negative article and positive article respectively. The 96-well plate was incubated at 37 °C in cell incubator of 5% CO<sub>2</sub> for 24 h. Six replicates of each test were tested.

25% Test article extract			cell lysis and slight growth inhibition.
			The cells showed a round shape and a change in cell morphology occasionally, and there were particles in the cytoplasm, occasionally cell lysis and slight growth inhibition.

### 9.2 Results of the cell vitality

Table2 Results of the cell vitality

Group	OD value								Viab. (%)
	1	2	3	4	5	6	$\bar{x}$	s	
Blank control	0.601	0.663	0.638	0.655	0.606	0.629	0.632	0.025	100.0
Negative control	0.598	0.582	0.598	0.599	0.597	0.593	0.595	0.006	94.1
Positive control	0.054	0.062	0.061	0.059	0.068	0.064	0.061	0.005	9.7
100% test article extract	0.489	0.497	0.486	0.495	0.492	0.502	0.494	0.006	78.1
75% test article extract	0.523	0.525	0.535	0.524	0.519	0.508	0.522	0.009	82.6
50% test article extract	0.539	0.538	0.547	0.536	0.535	0.537	0.539	0.004	85.2
25% test article extract	0.549	0.555	0.567	0.589	0.585	0.565	0.568	0.016	89.9

## 10.0 Conclusion

Under the conditions of this study, the test article have no potential toxicity to L-929 cells.

## 11.0 Record

All raw data pertaining to this study and a copy of the final report are to be stored in the designated archive files at Huestongwei.

## 12.0 Confidentiality Agreement

Statements of confidentiality were as agreed upon prior to study initiation.

After incubation, observe the cell morphology first and then discard the culture medium. Add 50 µl MTT (1mg/ml) to each well and then incubated at 37 °C in a humidified atmosphere of 5% CO<sub>2</sub> for 2 hours. The liquid in each well was tipped out and 100 µl isopropyl alcohol was added to each well to suspend the cell layer.

Evaluate the suspension above with a dual-wavelength spectrophotometer with the measurement wavelength at 570 nm.

## 7.0 Statistical method

Mean±standard deviation ( $\bar{x} \pm s$ )

The cell cytotoxicity ratio = OD<sub>750</sub> of test (or positive or negative) article group/ OD<sub>750</sub> of blank control group×100%.

## 8.0 Evaluation criteria

8.1 The 50% extract of the test article should have at least the same or a higher viability than the 100% extract. Otherwise the test should be repeated.

8.2 The lower the Viab.% value, the higher the cytotoxic potential of the test article is.

8.3 If viability is reduced to <70% of the blank, it has a cytotoxic potential.

8.4 The Viab.% of the 100% extract of the test article is the final result.

## 9.0 Results of the test

### 9.1 Results of the cell morphology

Table 1 Observation of the cell morphology

Group	Before inoculation	Before treated with extract	24 h after treatment
Blank control			Discrete intracytoplasmic granules, no cell lysis, no reduction of cell growth.
Negative control			Discrete intracytoplasmic granules, no cell lysis, no reduction of cell growth.
Positive control			Nearly complete or complete destruction of the cell layers.
100% Test article extract	Discrete intracytoplasmic granules, no cell lysis, no reduction of cell growth.	Discrete intracytoplasmic granules, no cell lysis, no reduction of cell growth.	The cells showed a round shape and a change in cell morphology occasionally, and there were particles in the cytoplasm, occasionally cell lysis and slight growth inhibition.
75% Test article extract			The cells showed a round shape and a change in cell morphology occasionally, and there were particles in the cytoplasm, occasionally cell lysis and slight growth inhibition.
50% Test article extract			The cells showed a round shape and a change in cell morphology occasionally, and there were particles in the cytoplasm, occasionally



中华人民共和国医疗器械注册证

注册证编号：湘械注准20202140297

注册人名称	湖南一喜科技发展有限公司
注册人住所	湖南浏阳高新技术产业开发区坪头北路6号3栋
生产地址	湖南浏阳高新技术产业开发区坪头北路6号3栋
代理人名称	不适用
代理人住所	不适用
产品名称	一次性使用医用口罩（非无菌型）
型号、规格	产品按照面罩形状分为平面形（PM），按照佩戴方式可以分为耳挂式（G）、带带式（B）或头带式（D），按照尺寸可以分为A、B、C、D、E、F、G七种规格。
结构及组成	本产品由罩布、鼻夹、口罩带组成。罩布内外层由无纺布，中间层由熔喷布制成。口罩带由无纺布或松紧带制成，鼻夹由可弯折的聚丙烯材料制成。该产品非无菌供应。
适用范围	适用于佩戴者在无体液和飞溅风险的普通医疗环境下的卫生护理。
附件	产品技术要求
其他内容	
备注	1. 该产品为应急审批注册，有效期为六个月；2. 该产品在延续/变更注册时应按医疗器械注册管理要求完善相关资料。

审批部门：湖南省药品监督管理局

批准日期：2020年03月19日

有效期至：2020年11月02日

中华人民共和国  
PEOPLE'S REPUBLIC OF CHINA  
医疗器械产品出口销售证明  
CERTIFICATE FOR EXPORTATION OF MEDICAL PRODUCTS

证书编号：湘药监械出 20200033 号  
Certificate NO.: HNMPA 20200033

产品名称：见附件  
Product(s): Attachment

规格型号：见附件  
Model: Attachment

产品注册或备案凭证号：见附件  
Registration certificate(s): Attachment

生产企业：湖南一喜科技发展有限公司  
Manufacturer: Hunan EEXI Technology & Service Co., Ltd.

生产企业住所：湖南浏阳高新技术产业开发区坪头北路6号3栋  
Address of manufacturer: No. 6, North of Pingtou road, Liuyang Hi-tech industrial development zone, Hunan, China

生产许可或备案凭证号：湘食药监械生产许 20200023 号  
Manufacturing License(s): Hunan CFDA Production Permit No. 20200023

兹证明上述产品已准许在中国生产和销售。  
This is to certify that the above products have been registered to be manufactured and sold in China.

证明有效期至：2022 年 3 月 19 日  
This certification valid until: Mar. 19, 2022

湖南省药品监督管理局  
HUNAN MEDICAL PRODUCTS ADMINISTRATION

2020 年 3 月 20 日  
Mar. 20, 2020



附件 (Attachment)			
序号 No.	产品名称 Product (s)	规格型号 Model	注册证编号 Registration certificate(s)
1	一次性使用医用口罩 (非无菌) Disposable medical mask (non-sterile)	规格型号: 平面挂耳型, A 11.5cm×7cm, B 12.5cm× 9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm×9.5cm, G 14.5cm×9cm Model: Flat mounting ear model, A 11.5cm× 7cm, B 12.5cm×9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm× 9.5cm, G 14.5cm×9cm	器械注册: 20202140297 Registration certificate(s): Hunan Machinery Registration: 20202140297
2	一次性使用医用口罩 (无菌型) Disposable medical face mask (Sterilized)	规格型号: 平面挂耳型, A 11.5cm×7cm, B 12.5cm× 9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm×9.5cm, G 14.5cm×9cm Model: Flat mounting ear model, A 11.5cm× 7cm, B 12.5cm×9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm× 9.5cm, G 14.5cm×9cm	器械注册: 20202140289 Registration certificate(s): Hunan Machinery Registration: 20202140289
3	医用外科口罩 (非无菌型) Surgical mask (Sterilized)	规格型号: 平面挂耳型, A 11.5cm×7cm, B 12.5cm× 9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm×9.5cm, G 14.5cm×9cm Model: Flat mounting ear model, A 11.5cm× 7cm, B 12.5cm×9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm× 9.5cm, G 14.5cm×9cm	器械注册: 20202140298 Registration certificate(s): Hunan Machinery Registration: 20202140298
4	医用外科口罩 (非无菌) Surgical mask (non-sterile)	规格型号: 平面挂耳型, A 11.5cm×7cm, B 12.5cm× 9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm×9.5cm, G 14.5cm×9cm Model: Flat mounting ear model, A 11.5cm× 7cm, B 12.5cm×9cm, C 17cm×9.5cm, D 17.5cm×9.5cm, E 18cm×9cm, F 18cm× 9.5cm, G 14.5cm×9cm	器械注册: 20202140288 Registration certificate(s): Hunan Machinery Registration: 20202140288

