

- 1. Table of Contents (Page 1)**
- 2. Basic Information (Page 2)**
- 3. Packaging & Pictures (Page 3-5)**
- 4. European Test & Technical Documentation Review Report
(Page 6-18)**
- 5. CE Declaration of Conformity (Page 19)**
- 6. ISO Production Standards Report (Page 20)**

Civil Protective Mask Eexilnherent YX153 With Valve

Brand: Eexilnherent

Type No.: YX153

**EU Standard: EN149 2001 + A1:2009, Regulation (EU) 2016/425,
tested by SGS (0598) Finland**

**Classifications: FFP2, Protects against Bacterial, particulates and
virus, Filtration efficiency $\geq 94\%$**

Inspection: The Netherlands Ministry of Labor and the Tax

Investigation Bureau provide the passed proof

**Features: Against COVID-19, purchased by British and German
Federal Ministry of Health**

**Material: Total 4 layers, PP nonwoven(outer layer), Melt-blown
fabric(2 layers), PP nonwoven(Inner layer)**

Mask Specifications: Universal, 15.50x10.50cm

**Earloop Design: Breathable and comfortable for prolonged
wearing**

Date of manufacture: From October 2020

Expiration date: 3 years

**Packing specifications: 5 pcs./bag, 60 bags/carton, 20
cartons/euro pallet**

P2



P3



P4

SGS

Certificate FI20/966321

**Hunan EEXI Technology &
Service Co., Ltd.**

No.6, North of Pingtong Road,
Luyang Hi-Tech Industrial Development Zone,
Hunan,
China

It is certified that the manufacturer's technical file and the PPE product detailed on
page 2 have been assessed and found to be in accordance with

Regulation (EU) 2016/425
Module B, EU type-examination

This certificate is valid from 11 August 2020 until 11 August 2025
2. Certified since 6 August 2020

Authorised by

FINAS
Finnish Accreditation Service
S003 (EN ISO/IEC 17065)

Tom Tom

SGS FIMKO OY, Notified Body 0598
Takomatie 8, FI-00380, Helsinki, Finland
t +358 9 696 361 f +358 9 692 5474 www.sgs.com

Page 1 of 2



This document is issued by the Company subject to its General Conditions of
Certification Services accessible at www.sgs.com/terms_and_conditions.htm.
Attention is drawn to the limitations of liability, indemnification and jurisdictional
issues established therein. The authenticity of this document may be verified at
<http://www.sgs.com/en/certified-clients-and-products/certified-client-directory>.
Any unauthorised alteration, forgery or falsification of the content or appearance
of this document is unlawful and offenders may be prosecuted to the fullest
extent of the law.



Certificate FI20/966321, continued

Hunan EEXI Technology & Service Co., Ltd.

Regulation (EU) 2016/425

Module B, EU type-examination

Issue 2

PPE Product

- a) Exinherent (logo) YX152 particle filtering half mask, consisting of a white five layer (polypropylene) disposable face mask, with nose clip, and polyester ear loops with polypropylene hook;
- b) Exinherent (logo) YX153 particle filtering half mask, consisting of a white five layer (polypropylene) disposable face mask, with polypropylene valve, nose clip, and polyester ear loops with polypropylene hook.

It is certified that the manufacturer's technical file and the above mentioned PPE have been assessed and found to meet the applicable Essential Health and Safety Requirements in Annex II of Regulation (EU) 2016/425 Personal Protective Equipment

The following have been applied:

EN 149:2001+A1:2009 (Respiratory protective devices - filtering half masks to protect against particles) device classification: FFP2 NR.



This certificate is issued on the strict condition that appropriate checks on manufactured PPE, as detailed in Article 19 (c) of the Regulation are implemented and maintained while the model is in production

Certification is based on technical file reference:

YX152/YX153, Rev 1, dated: 2020-08-07.

SGS Reference Number UK/CRS 241545.

This certificate remains the property of SGS Fimko Oy to whom it must be returned on request.



Certificate CN20/42194

The management system of

Hunan EEXI Technology & Service Co., Ltd.

No.6, North of Pingtuo Road, Liuyang Hi-Tech Industrial Development Zone, Hunan, 410300, P.R. China

has been assessed and certified as meeting the requirements of

Regulation (EU) 2016/425 Module C2

For the following activities

Manufacture of EexInherent (logo) YX152/YX153 Particle Filtering Half Mask.

(Note: All products marked CE0598 must have a valid EU type-examination certificate issued under Module B or a valid EC type-examination certificate issued under Article 10 of Directive 89/686/EEC.)

This certificate is valid from 14 August 2020
and remains valid subject to satisfactory surveillance audits.
Issue 1. Certified since 14 August 2020



Authorised by

Don Don

SGS FIMKO OY, Notified Body 0598

Takomitie 8, FI-00380 Helsinki, Finland
t +358 9 696 361 f +358 9 692 5474 www.sgs.com



Page 1 of 1

This document is issued by the Company subject to its General Conditions of Certification Services accessible at www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitations of liability, indemnification and jurisdictional issues established therein. The authenticity of this document may be verified at <http://www.sgs.com/notified-clients-and-products/notified-client-directory>. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.



中国认可
国际互认
检测
TESTING
CNAS L0599

Test Report

SL52035273350201TX

Date: July 29, 2020

Page 1 of 10

HUNAN EEXI TECHNOLOGY&SERVICE CO.,LTD
NO.6, NORTH OF PINGTOU ROAD, LIUYANG HI-TECH INDUSTRIAL DEVELOPMENT ZONE, HUNAN,
CHINA

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A) Particle filtering half mask

Claimed : FFP2

Style No. : YX153

Composition : (A) non-woven fabric, melt-blown fabric, bridge of nose, ear belt, breathing valve

Sample Color : (A) WHITE

Manufacturer : HUNAN EEXI TECHNOLOGY&SERVICE CO.,LTD

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Jul 16, 2020

Testing Period : Jul 16, 2020 - Jul 29, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, for further details, please refer to the following page(s).

Conclusion:

Sample No.	Recommendation Level
(A)	FFP2 NR

Signed for and on behalf of
SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo

Sara Guo (Account Executive)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing inspection report & certificate, please contact us at telephone: (86-755) 8367 1443, or email: CN1.doccheck@sgs.com

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd. 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
Testing Center, 3rd Floor, Yishan Road, Xuhui District Shanghai, China e sgs.china@sgs.com
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763

Member of the SGS Group (SGS SA)



Test Report

SL52035273350201TX

Date: July 29, 2020

Page 2 of 10

Test Result

Personal Protective Equipment - Respiratory Protective Devices- Filtering Half Masks to Protect against Particles- Requirements, Testing, Marking

EN 149:2001+A1:2009

Clause 7.4 Packaging

(EN 149:2001+A1:2009 Clause 8.2)

Test Requirement	Results	Comment
Particle filtering half masks shall be offered for sale packaged in such a way that they are protected against mechanical damage and contamination before use.	Comply	Pass

Clause 7.5 Material

(EN 149:2001+A1:2009, Clause 8.2 & 8.3.1 & 8.3.2)

Test Requirement	Results	Comment
Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used.	Comply	Pass
After undergoing the conditioning described in 8.3.1 none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps.	Comply	
When conditioned in accordance with 8.3.1 and 8.3.2 the particle filtering half mask shall not collapse.	Comply	
Any material from the filter media released by the air flow through the filter shall not constitute a hazard or nuisance for the wearer.	Comply	

Clause 7.6 Cleaning and Disinfecting

(EN 149:2001+A1:2009, Clause 8.4 & 8.5 & 8.11)

Test Requirement	Results	Comment
If the particle filtering half mask is designed to be re-usable, the materials used shall withstand the cleaning and disinfecting agents and procedures to be specified by the manufacturer. With reference to 7.9.2, after cleaning and disinfecting the re-usable particle filtering half mask shall satisfy the penetration requirement of the relevant class.	Not applicable (Not designed to be re-usable)	N.A.

Clause 7.7 Practical Performance

(EN 149:2001+A1:2009, Clause 8.4)

Test Requirement	Results	Comment
The particle filtering half mask shall undergo practical performance tests under realistic conditions. These general tests serve the purpose of checking the equipment for imperfections that cannot be determined by the tests described elsewhere in this standard.	No imperfections	Pass



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8397 1443, or email: CN_Doccheck@sgs.com

SGS CSTC (Shanghai) Technical Services (Shanghai) Co., Ltd. 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com

Member of the SGS Group (SGS SA)

P10



Test Report

SL52035273350201TX

Date: July 29, 2020

Page 3 of 10

Clause 7.8 Finish of Parts

(EN 149:2001+A1:2009, Clause 8.2)

Test Requirement	Results	Comment
Parts of the device likely to come into contact with the wearer shall have no sharp edges or burrs.	No sharp edges or burrs	Pass

Clause 7.9.1 Total Inward Leakage

(EN 149:2001+A1:2009, Clause 8.5)

Test Requirement	Results	Comment
The total inward leakage consists of three components: face seal leakage, exhalation value leakage (if exhalation value fitted) and filter penetration. For particle filtering half masks fitted in accordance with the manufacturer's information, at least 46 out of the 50 individual exercise results (i.e. 10 subjects x 5 exercises) for total inward leakage shall be not greater than: 25% for FFP1, 11% for FFP2, 5% for FFP3 and, in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than: 22% for FFP1, 8% for FFP2, 2% for FFP3	Detail refer to Appendix 1	Pass

Appendix 1: Summarization of Test Data**Inward Leakage Test Data**

Subject	Sample No.	Condition	Walk(%)	Head Side/side(%)	Head up/down(%)	Talk(%)	Walk(%)	Mean(%)
Zhou	1	A.R.	5.76	6.43	5.61	6.55	5.82	6.03
Luo	2	A.R.	6.89	7.59	7.49	7.00	6.19	7.03
Lu	3	A.R.	4.75	6.13	6.12	6.79	6.85	6.13
Wang	4	A.R.	5.21	4.32	6.64	4.21	4.87	5.05
Bao	5	A.R.	6.26	6.64	6.73	7.81	7.66	7.02
Ding	6	T.C.	4.85	5.74	5.09	5.40	5.51	5.32
Li	7	T.C.	8.24	7.34	6.82	6.03	8.70	7.43
Chen	8	T.C.	6.49	4.22	6.70	5.82	4.82	5.61
Song	9	T.C.	7.33	6.75	6.33	7.76	5.64	6.76
Ye	10	T.C.	8.07	8.80	7.31	6.62	7.53	7.67

Facial Dimension(mm)

Subject	Face length	Face Width	Face Depth	Mouth Width
Chen	125	150	120	58
Lu	115	132	107	48
Zhou	115	135	106	52
Li	125	130	107	46
Luo	125	136	100	43
Zheng	128	140	112	55
Wang	120	147	103	48
Song	120	140	100	50



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

SGS-CTC (Shanghai) Technical Services (Shanghai) Co., Ltd. 3rd Building No. 888, Yishan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
中国·上海·徐汇区宜山路888号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com

Member of the SGS Group (SGS SA)



Test Report SL52035273350201TX Date: July 29, 2020 Page 4 of 10

Bao	130	134	104	50
Ding	134	150	110	52
Liu	120	135	117	50
Ye	126	137	105	52

Clause 7.9.2 Penetration of Filter Material
(EN 149:2001+A1:2009, Clause 8.11 & EN 13274-7:2019)

Test Requirement			Results	Comment
The penetration of the filter of the particle filtering half mask shall meet the requirements of the following table.			Detail refer to Appendix 2	Pass
Classification	Maximum penetration of test aerosol			
	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min		
	%	%		
	max.	max.		
FFP1	20	20		
FFP2	6	6		
FFP3	1	1		

Appendix 2: Summarization of Test Data

Penetration of filter material

Aerosol	Condition	Sample No.	Penetration (%)
Sodium chloride test	As received	1	0.132
		2	0.110
		3	0.149
	Simulated wearing treatment	4	0.122
		5	0.141
		6	0.137
	Mechanical strength +Temperature conditioned	7	0.231
		8	0.272
		9	0.245
Paraffin oil test	As received	10	0.235
		11	0.347
		12	0.296
	Simulated wearing treatment	13	0.302
		14	0.346
		15	0.374
	Mechanical strength +Temperature conditioned	16	0.334
		17	0.574
		18	0.472
Flow conditioning : Single filter: 95.0 L/min			



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

SGS China Technical Services (Shanghai) Co., Ltd.
Testing Center: 300, Yehou Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958783 www.sgs.com.cn
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958783 e sgs.china@sgs.com

Member of the SGS Group (SGS SA)



Test Report SL52035273350201TX
Clause 7.10 Compatibility with Skin
(EN 149:2001+A1:2009, Clause 8.4 & 8.5)

Date: July 29, 2020

Page 5 of 10

Test Requirement	Results	Comment
Materials that may come into contact with the wearer's skin shall not be known to be likely to cause irritation or any other adverse effect to health.	No irritation or any other adverse effect to health	Pass

Clause 7.11 Flammability

(EN 149:2001+A1:2009, Clause 8.6)

Test Requirement	Results	Comment
The material used shall not present a danger for the wearer and shall not be of highly flammable nature When tested, the particle filtering half mask shall not burn or not to continue to burn for more than 5 s after removal from the flame.	Detail refer to Appendix 3	Pass

Appendix 3: Summarization of Test Data

Flammability

Condition	Sample No.	Result
As received	1	NIL
	2	NIL
Temperature conditioned	3	NIL
	4	NIL

Clause 7.12 Carbon Dioxide Content of The Inhalation Air

(EN 149:2001+A1:2009, Clause 8.7)

Test Requirement	Results	Comment
The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1,0 % (by volume)	Detail refer to Appendix 4	Pass

Appendix 4: Summarization of Test Data

Carbon Dioxide Content of The Inhalation Air

Condition	Sample No.	Result(%)
As received	1	0.4810
	2	0.4831
	3	0.4824
		Mean value: 0.48



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
SGS CSTC (Shanghai) Technical Services (Shanghai) Co., Ltd. 3rd Building, No. 689, Yixian Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64956763 www.sgs.com.cn
中国·上海·徐汇区宜山路689号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64956763 e sgs.china@sgs.com

Member of the SGS Group (SGS SA)



Test Report SL52035273350201TX
Clause 7.13 Head Harness
 (EN 149:2001+A1:2009, Clause 8.4 & 8.5)

Date: July 29, 2020

Page 6 of 10

Test Requirement	Results	Comment
The head harness shall be designed so that the particle filtering half mask can be donned and removed easily.	Comply	Pass
The head harness shall be adjustable or self-adjusting and shall be sufficiently robust to hold the particle filtering half mask firmly in position and be capable of maintaining total inward leakage requirements for the device.	Comply	

Clause 7.14 Field of Vision
 (EN 149:2001+A1:2009, Clause 8.4)

Test Requirement	Results	Comment
The field of vision is acceptable if determined so in practical performance tests.	Comply	Pass

Clause 7.15 Exhalation Valve(s)
 (EN 149:2001+A1:2009, Clause 8.2 & 8.9.1 & 8.3.4 & 8.8)

Test Requirement	Results	Comment
(a) A particle filtering half mask may have one or more exhalation valve(s), which shall function correctly in all orientations.	Comply	Pass
(b) If an exhalation valve is provided it shall be protected against or be resistant to dirt and mechanical damage and may be shrouded or may include any other device that may be necessary for the particle filtering half mask to comply with 7.9.	Comply	
(c) Exhalation valve(s), if fitted, shall continue to operate correctly after a continuous exhalation flow of 300 l/min over a period of 30 s.	Comply	
(d) When the exhalation valve housing is attached to the faceblank, it shall withstand axially a tensile force of 10N applied for 10 s.	Comply	



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
 Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

SGS-CTI (Shanghai) Technical Services (Shanghai) Co., Ltd. 3rd Building, No. 689, Yehon Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
 Testing Center for Service Support 中国·上海·徐汇区宜山路689号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com

Member of the SGS Group (SGS SA)



Test Report

SL52035273350201TX

Date: July 29, 2020

Page 7 of 10

Clause 7.16 Breathing Resistance

(EN 149:2001+A1:2009, Clause 8.9)

Test Requirement				Results	Comment				
The penetration of the filter of the particle filtering half mask shall meet the requirements of the following table.				Detail refer to Appendix 5	Pass				
						Classification	Maximum permitted resistance (mbar)		
							Inhalation		Exhalation
							30 l/min	95 l/min	160 l/min
						FFP1	0.6	2.1	3.0
						FFP2	0.7	2.4	3.0
						FFP3	1.0	3.0	3.0

Appendix 5: Summarization of Test Data**Breathing resistance (mbar)**

As received	Flow rate(l/min)		1					2					3				
			A	B	C	D	E	A	B	C	D	E	A	B	C	D	E
	Inhalation	30	0.5	0.6	0.5	0.6	0.5	0.6	0.5	0.6	0.5	0.5	0.6	0.6	0.5	0.6	0.6
Simulated wearing treatment	Inhalation	95	2.1	2.1	2.2	2.1	2.2	2.1	2.1	2.2	2.1	2.1	2.2	2.1	2.1	2.2	2.2
		160	2.3	2.4	2.3	2.3	2.3	2.4	2.4	2.3	2.3	2.3	2.3	2.4	2.3	2.4	2.3
	Exhalation	160	2.3	2.4	2.3	2.3	2.3	2.4	2.4	2.3	2.3	2.3	2.4	2.3	2.3	2.4	2.4
Temperature conditioned	Flow rate(l/min)		4					5					6				
			A	B	C	D	E	A	B	C	D	E	A	B	C	D	E
	Inhalation	30	0.5	0.6	0.5	0.6	0.6	0.5	0.6	0.5	0.5	0.6	0.6	0.5	0.6	0.5	0.6
Flow conditioned	Inhalation	95	2.2	2.2	2.2	2.1	2.1	2.2	2.2	2.1	2.1	2.2	2.2	2.1	2.2	2.1	2.2
		160	2.2	2.3	2.3	2.2	2.3	2.3	2.3	2.2	2.3	2.2	2.3	2.3	2.2	2.3	2.3
	Exhalation	160	2.2	2.3	2.3	2.2	2.3	2.3	2.3	2.2	2.3	2.2	2.3	2.2	2.2	2.2	2.3
Flow conditioned	Flow rate(l/min)		7					8					9				
			A	B	C	D	E	A	B	C	D	E	A	B	C	D	E
	Inhalation	30	0.5	0.6	0.5	0.6	0.6	0.5	0.5	0.6	0.6	0.5	0.6	0.5	0.6	0.6	0.5
Flow conditioned	Inhalation	95	2.1	2.2	2.1	2.1	2.2	2.1	2.2	2.1	2.2	2.2	2.1	2.1	2.1	2.2	2.2
		160	2.2	2.3	2.3	2.2	2.3	2.3	2.3	2.2	2.3	2.2	2.3	2.3	2.2	2.3	2.3
	Exhalation	160	2.2	2.3	2.2	2.2	2.3	2.3	2.3	2.2	2.3	2.3	2.3	2.2	2.2	2.2	2.3

A: facing directly ahead; B: facing vertically upwards; C: facing vertically downwards; D: lying on the left side; E: lying on the right side



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

SGS-CSL (Shanghai) Technical Services Co., Ltd. 200233 t: (86-21) 61402666 f: (86-21) 64958763 www.sgs.com.cn
3rd Building No. 889, Yehon Road, Xuhui District Shanghai, China
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t: (86-21) 61402666 f: (86-21) 64958763 e: sgs.china@sgs.com

Member of the SGS Group (SGS SA)



Test Report

SL52035273350201TX

Date: July 29, 2020

Page 8 of 10

Clause 7.17 Clogging

(EN 149:2001+A1:2009, Clause 8.9 & 8.10)

Test Requirement	Results	Comment																				
Clause 7.17.2 Breathing resistance Valved particle filtering half masks: After clogging the inhalation resistances shall not exceed: FFP1: 4 mbar, FFP2: 5 mbar, FFP3: 7 mbar at 95L/min continuous flow The exhalation resistance shall not exceed 3 mbar at 160 L/min continuous flow. Valveless particle filtering half masks: After clogging the inhalation and exhalation resistances shall not exceed: FFP1: 3 mbar, FFP2: 4 mbar, FFP3: 5 mbar at 95L/min continuous flow	Optional for single shift device only	N.A.																				
Clause 7.17.3 Penetration of filter material All types (valved and valveless) of particle filtering half masks claimed to meet the clogging requirement shall also meet the requirements. <table><tr><th rowspan="2">Classification n</th><th colspan="2">Maximum penetration of test aerosol</th></tr><tr><th>Sodium chloride test 95 l/min</th><th>Paraffin oil test 95 l/min</th></tr><tr><td></td><td>%</td><td>%</td></tr><tr><td></td><td>max.</td><td>max.</td></tr><tr><td>FFP1</td><td>20</td><td>20</td></tr><tr><td>FFP2</td><td>6</td><td>6</td></tr><tr><td>FFP3</td><td>1</td><td>1</td></tr></table>	Classification n	Maximum penetration of test aerosol		Sodium chloride test 95 l/min	Paraffin oil test 95 l/min		%	%		max.	max.	FFP1	20	20	FFP2	6	6	FFP3	1	1	Optional for single shift device only	N.A.
Classification n		Maximum penetration of test aerosol																				
	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min																				
	%	%																				
	max.	max.																				
FFP1	20	20																				
FFP2	6	6																				
FFP3	1	1																				

Clause 7.18 Demountable Parts

(EN 149:2001+A1:2009, Clause 8.2)

Test Requirement	Results	Comment
All demountable parts (if fitted) shall be readily connected and secured, where possible by hand	Comply	Pass

Test	Uncertainty
Total inward leakage	3.4%
Penetration of filter material	4.8%
Carbon dioxide content of the inhalation air	3.9%
Breathing resistance (30L/min)	5.9%
Breathing resistance (95L/min)	4.9%
Breathing resistance (160L/min)	4.3%




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

SGS-CSH, Shanghai Technical Services (Shanghai) Co., Ltd. 3rd Building No. 888, Yishan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64968763 www.sgs.com.cn
中国·上海·徐汇区宜山路888号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64968763 e sgs.china@sgs.com

Member of the SGS Group (SGS SA)

Sample Photo



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

SGS-CSL (Shanghai) Technical Services (Shanghai) Co., Ltd.	3 rd Building No.889, Yishan Road, Xuhui District Shanghai, China	200233	t (86-21) 61402666	f (86-21) 64958763	www.sgsgroup.com.cn
Testing Center	中国·上海·徐汇区宜山路889号3号楼	邮编: 200233	t (86-21) 61402666	f (86-21) 64958763	e sgs.china@sgs.com

Member of the SGS Group (SGS SA)



End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
 3rd Building No. 888, Yishan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763 www.sgs.com.cn
 中国·上海·徐汇区宜山路888号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com

Member of the SGS Group (SGS SA)

EU Declaration of Conformity

Manufacturer: Hunan EEXI Technology & Service Co., Ltd.
Address: No.6, North of Pingtuo road, Liuyang Hi-Tech industrial
development zone, Hunan, China

declares that the new PPE described hereafter

Product name	Particle filtering half mask
Model	YX152, YX153
Standard	EN 149 : 2001+A1 : 2009
Classification	FFP2 NR

is in conformity with the Regulation (EU) 2016/425 and with harmonized standard
EN 149:2001 + A1:2009

performed the EU type-examination (Module B) and issued the EU
type-examination certificate

FI20/966321

performed the EU type-examination (Module C2) and issued the EU
type-examination certificate

CN20/42194

Notified Body information

Module B	Module C2
SGS Fimko Oy Takomotie 8, FI-00380 Helsinki, Finland Notified Body No. 0598	SGS Fimko Oy Takomotie 8, FI-00380 Helsinki, Finland Notified Body No. 0598

Signed for on behalf of Company

Name: Robin Liu

Position: CEO & Owner

Date: 2020-08-18

Signature:



Certificate CN20/42091

The management system of

**Hunan EEXI
Technology & Service Co., Ltd.**

No. 6, North of Pingtuo Road, Liuyang Hi-tech Industrial Development Zone,
Hunan Province, 410300, P.R. China

has been assessed and certified as meeting the requirements of

**ISO 13485:2016
EN ISO 13485:2016**

For the following activities

**Design and manufacture of non-sterile surgical face mask,
non-sterile single use medical face mask and
non-sterile medical protective face mask**

This certificate is valid from 16 June 2020 until 15 June 2023
and remains valid subject to satisfactory surveillance audits.

Re certification audit due before 5 June 2023
Issue 1. Certified since 16 June 2020

Authorised by



SGS United Kingdom Ltd
Rossmore Business Park, Ellesmere Port, Cheshire, CH65 3EN, UK
t +44 (0)151 350-6666 f +44 (0)151 350-6600 www.sgs.com

HC SGS 13485 2016 0118

Page 1 of 1



