



INTCO

英科医疗



医用丁腈检查手套

蓝色

MEDICAL NITRILE EXAMINATION GLOVES

产品描述：采用丁腈胶乳制造。有足够的强度和阻隔性能。

非无菌提供，一次性使用。

预期用途：用于戴在医生手上对患者病情进行检查或触检。

山东英科医疗制品有限公司

山东省青州市狷山工业园齐王路

Tel:86-536-6136888



www.intcomedical.com.cn

医疗器械生产备案凭证编号：鲁潍食药监械生产备20160008号

医疗器械备案凭证编号：鲁潍械备20180030号

医疗器械技术要求编号：鲁潍械备20180030号

其他内容详见产品使用说明书



请将本产品置于阴凉干燥处，避免阳光直射



生产日期：20200205



生产批次：20200205



有效期：五年



6 970245 756090



医用丁腈检查手套

MEDICAL NITRILE EXAMINATION GLOVES
INMBN-350 左右手通用 非灭菌

无粉 指麻 蓝色

包装: 100支/盒 10盒/箱
净重: 3.5 KG
毛重: 4.5 KG
型号: 中号
批号: 20181102
生产日期: 201811
有效期: 五年

**INTCO**
英科医疗

股票代码 300677



中号

生产商:

山东英科医疗制品有限公司
山东省青州市玛山工业园齐王路
Tel: 86-536-6136888

医疗器械生产备案凭证编号: 鲁潍食药监械生产备20160008号
医疗器械备案凭证编号: 鲁潍械备20180030号
医疗器械技术要求编号: 鲁潍械备20180030号

医用丁腈检查手套

MEDICAL NITRILE EXAMINATION GLOVES
INMBN-350 左右手通用 非灭菌

无粉 指麻 蓝色




请将本产品置于阴凉干燥处, 避免阳光直射

Nitrile gloves

Nitrile Examination Gloves

Comfortable, super soft flexible powder free nitrile gloves provide added safety in many applications. Designed with a special nitrile formulation they feel and fit like latex and allow full range of motion and excellent flexibility to minimize stress and fatigue. They do not contain natural rubber latex and are an excellent alternative for those suffering from Type I allergies.



Features

- 100% latex free – made from durable, puncture resistant, protein and powder-free nitrile, eliminates the Type I allergic reaction associated with natural rubber latex
- Highly elastic and super soft
- Rolled rim to facilitate easy donning
- Designed to give a natural rubber-like feel
- Textured in finger tips for enhanced grip in all situations
- Food safe, suitable for handling fatty foods incl. cooking, olive and cod liver oil (Special Product)
- Protects against a wide variety of chemicals
- AQL 1.5, 2.5, 4.0 Available



Quality standards

- Complies with EN 455 and EN 374
- Complies with ASTM D6319 (USA Related Product)

Technical information

Type	Powder-free, unsterile
Material	Nitrile butadiene rubber (NBR)
Colour	Blue, White, Purple, Pink, Black, Red
Model	Ambidextrous, with rolled rim
Storage	Protect from heat, humidity, strong light and ozone
Size/overall length as per EN 455-2	XS, S, M, L and XL 240 mm (≥230mm for USA)
Impermeability as per EN 455-1	AQL 1.5, 2.5, 4.0 Available
Durability, in original package if stored as per din 7716, ISO 2230	Min. 5 years

Powder-free

REF	Size	Dispenser	Transport carton
NG-141	XS	100	1000
NG-141	S	100	1000
NG-141	M	100	1000
NG-141	L	100	1000
NG-141	XL	100	1000

Color coding in sizes

Glove boxes color coded for easy size identification



Loading Methods

Loading	Dispenser	Transport carton	PCs/CTN
Method 1	100 PCs	10 Boxes	1,000 PCs
Method 2	150 PCs	10 Boxes	1,500 PCs
Method 3	200 PCs	10 Boxes	2,000 PCs

- Method for loading Nitrile, Vinyl and Latex series.
- Loading more than 2,000 pieces per carton is much heavy for carrying and delivery.



Technical Data

Sizes	XS	S	M	L	XL
Length (mm)	235±5mm	235±5mm	235±5mm	235±5mm	235±5mm
Palm Width	75±5mm	85±5mm	95±5mm	110±5mm	120±5mm
Thumb L.	52±2mm	54±2mm	57±2mm	59±2mm	65±2mm
Forefinger L.	66±2mm	68±2mm	72±2mm	74±2mm	80±2mm
Middle finger L.	77±2mm	79±2mm	82±2mm	85±2mm	92±2mm
Ring finger L.	68±2mm	71±2mm	75±2mm	77±2mm	82±2mm
Little finger L.	52±2mm	54±2mm	56±2mm	60±2mm	64±2mm
Physical Specification		Vinyl Premium	Synthetic Premium	Nitrile Premium	Latex Premium
	Palm Thickness	Min. 0.08mm	Min. 0.08mm	Min. 0.05mm	Min. 0.08mm
	Finger Thickness	Min. 0.05mm	Min. 0.05mm	Min. 0.05mm	Min. 0.08mm
	Wrist Thickness	Min. 0.06mm	Min. 0.06mm	Min. 0.05mm	Min. 0.08mm
	Tensile Strength Before Aging	Min. 11MPa Avg. 12MPa	Min. 11MPa Avg. 16MPa	Min. 14MPa Avg. 18MPa	Min. 18MPa Avg. 20MPa
	Tensile Strength After Aging	Min. 11MPa Avg. 11MPa	Min. 11MPa Avg. 15MPa	Min. 14MPa Avg. 16MPa	Min. 14MPa Avg. 15MPa
	Elongation Rate Before Aging	Min. 300% Avg. 420%	Min. 300% Avg. 460%	Min. 500% Avg. 580%	Min. 650% Avg. 680%
	Elongation Rate After Aging	Min. 300% Avg. 360%	Min. 300% Avg. 410%	Min. 400% Avg. 470%	Min. 500% Avg. 600%
Packaging Dimensions (mm) & Loading Capabilities (Each product may differ from others)	L x W x H	Method 1 (1000 PCs/CTN)	Method 2 (1500 PCs/CTN)	Method 3 (2000 PCs/CTN)	
	Dispenser	230x125x60mm	240x125x70mm	230x130x78mm	
	Trans. CTN.	315x258x250mm	385x260x255mm	408x268x240mm	
	Capability for 20 GP	1500 CTNs	1170 CTNs	1150 CTNs	
	Capability for 40 GP	3100 CTNs	2400 CTNs	2380 CTNs	
	Capability for 40 HQ	3450 CTNs	2700 CTNs	2600 CTNs	

The information above are for reference only. Specific order shall be confirmed according to regional regulations and customers' requirements.

Qualification

MDD Product Qualification



FDA 510(K)

EN 455

ASTM D5250

ASTM D6319

Quality Control System



GMP

SGS EN ISO 13485

SGS EN ISO 13485

SGS EN ISO 9001

Other Qualification



日本食品衛生安全法検査成績書
Japanese Food Safety Code 370

BRC

CE-PPE-PVC

CE-PPE-NBR



Document Number : INTCO-CE-DC-NBR-001

Version: A/1

EU DECLARATION OF CONFORMITY

Manufacturer

Authorized Representative

Name: Shandong Intco Medical Products Co., Ltd.

Name: Lotus NL B.V.

Address: Qiwang Road, Naoshan Industrial Park. Qingzhou, Shandong, China

Address: Koningin Julianaplein 10, le Verd, 2595AA, The Netherlands

Declares that the MDR described hereafter

Product name and model:

Disposable Nitrile (NBR) Gloves

UMDNS code: 11882

UDI-DI: (XS, S, M, L, XL, XXL)

Blue Gloves: 6970245756014 / 6970245756021 / 6970245756038 / 6970245756045 / 6970245756052 / 6970245756069

White Gloves: 6970245756410 / 6970245756427 / 6970245756434 / 6970245756441 / 6970245756458 / 6970245756465

Black Gloves: 6970245756519 / 6970245756526 / 6970245756533 / 6970245756540 / 6970245756557 / 6970245756564

Violet Gloves: 6970245756311 / 6970245756328 / 6970245756335 / 6970245756342 / 6970245756359 / 6970245756366

Meet the provisions of the Council Regulation EU 2017/745 which apply to them.

The medical device has been assigned to class **I** according to Annex *VIII* of the Regulation EU 2017/745. It bears the mark



CONFORMITY ASSESSMENT ROUTE: *EU 2017/745, Annex I & VII*

This Declaration of conformity is valid in connection with the release document for the respective batch of produced devices.

The above mentioned declaration of conformity is exclusively under the responsibility of

Company: Shandong Intco Medical Products Co., Ltd.

Address: Qiwang Road, Naoshan Industrial Park. Qingzhou, Shandong, China

Shandong 2019-05-06

Place, date

Chi Yongtao Plant manager

Legally binding signature, Function

EU Type-Examination Certificate

Certificate number: 2777/11804-01/E00-00

This EU Type-Examination Certificate covers the following product group(s) supported by testing to the relevant standards/technical specifications and examination of the technical file documentation:

Following the EU Type-Examination this product group has been shown to satisfy the applicable essential health and safety requirements of Annex II of the PPE Regulation (EU) 2016/425 as a Category III product.

Product reference:

Description:

697024575

Five fingered disposable nitrile non-sterile gloves.

Blue 697024575 601-605

Violet 697024575 631-635

White 697024575 641-645

Black 697024575 651-655

Sizes:

6/XS, 6.5/S, 7/M, 8/L, 9/XL

Classification:

EN ISO 374-1:2016+A1:2018 /Type B

40% Sodium Hydroxide (K)

30% Hydrogen peroxide (P)

37% Formaldehyde (T)

EN ISO 374-5: 2016

Protection against Bacteria and fungi

Protection against viruses

Level

6

2

3

Pass

Pass

EN 374-4: 2013

-11.5 %

-9.5%

7.4 %

Standards/Technical specifications applied:

EN 420: 2003+A1: 2009; EN ISO 374-1:2016+A1:2018; EN ISO 374-5:2016

Technical reports/Approval documents:

SATRA: CHT0278438/1848

SGS: QDHL1806013113OT, CH:TX:6420074520, CH:TX: 9420020333, CH:TX: 9420029243 CH:TX: 9420026599-1, CH:TX:

9420014953-1, CH:TX: 9420026316-1, CH:TX: 9420614959

TUV: 721642857-2

Signed on behalf of SATRA:



Tara Saunders



Geoff Graham

Date first issued: 30/01/2019

Date of issue: 30/01/2019

Expiry date: 30/01/2024

TERMS AND CONDITIONS

The following conditions apply in addition to SATRA's standard terms and conditions of business and those given in the current certification agreement.

The certificate holder is licensed to mark the products detailed within this certificate in accordance with Annex V (Module B) of the Regulation (EU) 2016/425 of the European Parliament and of the council of 9th March 2016 on personal protective equipment once you have drawn up an EU declaration of product conformity. Please note:

1. Where the product is classified as category III then CE Marking of production is reliant on current compliance with Regulation 2016/425 module C2 or Module D. (Except that specifically produced to fit an individual user).
2. Full details of the certification and product are contained within the manufacturer's technical documentation.
3. Where a translation of this certificate exists, the English language version shall be considered as the authoritative text.
4. Certification is limited to production undertaken at the sites listed in the manufacturers technical documentation.
5. Ongoing manufactured product shall be consistent with the product(s) certified and listed on this certificate.
6. The Manufacturer shall inform SATRA of any changes to the certified product or technical documentation.
7. This certificate shall be kept together with the relevant technical documentation in a safe place by the client named on this certificate. Production of this certificate and other documentation may be required by a representative of the EC member state government.
8. This certificate relates only to the condition of the testable items at the time of the certification procedure and is subject to the expiry date shown.
9. SATRA Technology reserves the right to withdraw this certificate if it is found that a condition of manufacture, design, materials or packaging have been changed and therefore no longer comply with the requirements of Regulation 2016/425.

Test Report No. 7191205302-EEC19-WBH
dated 01 Mar 2019



PSB Singapore

**Add value.
Inspire trust.**

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

SUBJECT:

Testing of Disposable Nitrile Glove submitted by
Shandong Intco Medical Products Co., Ltd. on 18 Feb 2019.

TESTED FOR:

Shandong Intco Medical Products Co., Ltd
No. 9888 Qiwang Road
Naoshan Industry Park, Qingzhou, Shandong, China

TEST DATE:

25 Feb 2019

DESCRIPTION OF SAMPLES:

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Disposable Nitrile Glove	Blue	/	M	217	Shandong Intco Medical Products Co., Ltd

Lot size as specified by client: 35,001 to 150,000 pieces

METHOD OF TEST:

EN 455-1:2000 Medical gloves for single use
Part 1: Requirements and testing for freedom from holes



Laboratory:
TÜV SÜD PSB Pte. Ltd.
No.1 Science Park Drive
Singapore 118221

Phone : +65-6885 1333
Fax : +65-6776 8670
E-mail: enquiries@tuv-sud-psb.sg
www.tuv-sud-psb.sg
Co. Reg : 199002667R

Regional Head Office:
TÜV SÜD Asia Pacific Pte. Ltd.
1 Science Park Drive, #02-01
Singapore 118221
TUV®

Test Report No. 7191205302-EEC19-WBH
dated 01 Mar 2019



PSB Singapore

RESULTS:

Sample: Disposable Nitrile Glove, Size M

Table: Results for EN 455-1:2000

Clause	Tests	Requirements	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4 5	Freedom from holes	Shall not leak	7	200	0	Passed

REMARKS:

1. The manufacturing lot no. was not provided by the client.



Yeo Poh Kwang
Higher Associate Engineer



Wong Bee Hui
Product Manager
Medical Health Services (NAM)

APPENDIX:



Photo : Disposable Nitrile Glove, Size M

Test Report No. 7191205302-EEC19-WBH
dated 01 Mar 2019



Please note that this Report is issued under the following terms :

1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
2. The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, No.1 Science Park Drive Singapore 118221.

July 2011



Test Report

No.: QDHL1909015461OT

Date: SEP.25,2019

Page: 1 of 3

SHANDONG INTCO MEDICAL PRODUCTS CO., LTD
NO.9888, QIWANG ROAD, NAOSHAN INDUSTRY PARK, QINGZHOU, SHANDONG, CHINA

The following sample(s) was/were submitted and identified by the client as:

Sample Description : METRO/MAKRO PROFESSIONAL NITRILE GLOVES, NON-
POWDERED, BLUE

Sample Receiving Date : SEP.12,2019

Testing Period : SEP.12,2019 TO SEP.25,2019

Test Performed : SELECTED TEST(S) AS REQUESTED BY APPLICANT

Test Requested : EN 455-2-2015 MEDICAL GLOVES FOR SINGLE USE – PART 2:
REQUIREMENTS AND TESTING FOR PHYSICAL PROPERTIES

Test Result(s) : PLEASE REFER TO THE FOLLOWING PAGE(S)

Conclusion : THE SUBMITTED SAMPLE MET THE TEST REQUIREMENT.

Remark: Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

SGS-CSTC Standards
Technical Services (Qingdao)
Co., Ltd.



Zhou Xinkuan, SK
Lab Manager



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101
t (86-532) 68999888 f (86-532) 80991955

www.sgs.com
sgs.china@sgs.com

Test Report

No.: QDHL1909015461OT

Date: SEP.25,2019

Page: 2 of 3

Test Conducted:

EN 455-2-2015 Medical gloves for single use – part 2: Requirements and testing for physical properties

Number of test sample	:	26 Pieces
The type of gloves	:	examination/procedure gloves c)
Manufacturing batch code	:	/
Size	:	Examination/procedure gloves: M
Defects observed before testing	:	No defects

Clause	Test Items	Result	Note
5	Strength	---	---
5.2	Force at break	Pass	# 1
5.3	Force at break after challenge testing	Pass	# 1

Notes : #1 See result 1

Test Result:

1. Strength

Sample Quantity: 13pcs

Size	M												
Force at break(N)	7.8	8.5	8.0	9.0	9.4	8.9	6.8	7.1	8.2	8.9	8.3	8.6	8.4
Force at break after challenge testing(N)	7.8	7.6	8.3	7.6	6.5	6.1	8.4	7.4	6.8	6.8	8.5	7.2	6.0

Median value:

Force at break during shelf life (N): 8.4

Force at break after challenge testing (N): 7.4



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101
t (86-532) 68999888 f (86-532) 80991955

www.sgs.com
sgs.china@sgs.com

Requirements: see table 3

Table 3 — Median values of force at break

	Force at break in Newton		
	Surgical gloves a)	Examination/procedure gloves	
		b)	c)
Throughout shelf life tested according to 5.2 and within 12 months of manufacture tested according to 5.3	≥ 9,0	≥ 6,0	≥ 3,6
a) Requirements for all surgical gloves. b) Requirements for all examination gloves, except gloves made from thermoplastic materials (e.g. polyvinylchloride, polyethylene). c) Requirements for gloves made from thermoplastic materials (e.g. polyvinylchloride, polyethylene).			

Sample Photo:

Received sample



SGS authenticate the photo on original report only

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.
Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com



Material and Engineering Laboratory-Kaohsiung

Test Report



Report No. : KV-18-11251
Page No. : 1 OF 2
Date of Report : Dec. 25, 2018

Shandong Intco Medical Products Co., Ltd.

No.9888, Qiwang Road, Naoshan Industry Park, Qingzhou, Shandong, China

<u>Product Name</u>	Disposable Nitrile Glove (QDHL1811025521OT)
<u>Date of Sample Received</u>	Dec. 10, 2018
<u>Date of Testing</u>	Dec. 10, 2018~Dec. 25, 2018
<u>Remark</u>	The information mentioned in the above section is provided by Client(Exclude Date of Sample Received and Date of Testing)

The laboratory tests according to the test requests and samples provided by client, and the results are as follows:

Test Request : Aqueous Extractable Protein

Test Method : Refer to BS EN 455-3:2015 Medical gloves for single use —
Part 3 : Requirements and testing for biological evaluation

Test Result : Please see attached pages

----- 1 -----

The required specification(s) offered in this test report is/are for reference only.
The conformity judgment is at the Applicant's final verdict.


Signed for and on behalf of
SGS Taiwan Ltd.

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

Test Report



Report No. : KV-18-11251

Page No. : 2 OF 2

Date of Report : Dec. 25, 2018

Test Equipment :

Name	Brand	Model
UV-VISIBLE Spectrophotometer	SHIMADZU	UV-1700

Lab. Environmental Conditions:

Ambient Temperature : (25 ± 2) °C

Relative humidity : (50 ± 10) %


Test Result :

INSPECTION ITEM	TEST RESULT
Aqueous Extractable Protein (ppm)	n.d.

Note: 1. n.d. = not detected.

2. MDL (METHOD DETECTION LIMIT):0.2ppm.

Sample Photo :



----- oOo -----

The required specification(s) offered in this test report is/are for reference only.
The conformity judgment is at the Applicant's final verdict.

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.