


Measurement of UV radiation from exhaust and suction SRSmedilux: TW220



Opening to the UVC chamber after screwing out one screw.


The subject of the measurement is harmful UV-C radiation invisible to the human eye (blue radiation visible to the eye is not harmful to humans and living organisms).

The measurement was performed by using of the calibrated spectrometer **ASEQ INSTRUMENTS LR1-B V1.0** at first in the opening of the UVC chamber, then in the exhaust / suction of TW220.

Chamber UVC measurement:



Probe aimed at the opening of the UVC chamber.




Measured peak of UVC radiation.

Exhaust / suction UVC measurement:



Probe aimed at the exhaust / suction.



UVC radiation was not detected.

Bratislava, 18.3.2021

SRS Group s.r.o.
Rybnicna 36/D, 831 07 Bratislava, Slovakia
Phone: +421948590709, email: sales@srs-group.com
IČO: 44520433, IČ-DPH/VAT No.: SK2022726629