



CERTIFICATE OF CONFORMITY

Personal Protective Equipment (PPE) (EU) 2016/425


Holder: 
Address: 
Manufacturer: 
Address: 
Product: PE/TPE Gloves
Trade Mark: /
Model/Type: S, M, L, XL
Test Standard: EN ISO 374-1:2016/A1:2018 & EN ISO 374-5:2016
Test Report: 

This certificate of conformity is based on the product mentioned above. It is applied in connection with the corresponding Test Report. Technical data and documentation are at the manufacturer's disposal.

This is to certify that the product tested is in conformity with the above EC Directive(s). The CE mark as shown below can be used, under the responsibility of the manufacturer or the importer, after compliance with all relevant EC Directive(s).



Authorized Signatory:


Daniel Hwang
Mar. 19th, 2020

IDIC TESTING AND CERTIFICATION LIMITED

Add: 10/F., San Toi Building, 137-139 Connaught Road Central, Hong Kong

Website: www.idictc.com

Email: ce@idictc.com



TEST REPORT

Sample Name: TPE Gloves

Model / type: S, M, L, XL

Date (s) of tests: 2020-03-02 ~ 2020-03-19

- The test result refers exclusively to the test presented test model / sample.
- Without written approval of *IDIC Testing and Certification Limited*, the test report shall not be reproduced except in full.
- This test report is only applicable to European Community.



IDIC TESTING AND CERTIFICATION LIMITED

10/F., San Toi Building, 137-139 Connaught Road Central, Hong Kong

Website: www.idictc.com

Email: ce@idictc.com



Description of Sample(s): Model No.: S, M, L, XL
Brand: Hybrids

Date Sample(s) Received: 2020-03-02

Test Standard(s): EN ISO 374-1:2016/A1:2018 & EN ISO 374-5:2016

Test Result: Please see next pages in detail.

Conclusion: The submitted product **COMPLIED** with the requirements of Regulation (EU) 2016/425.

Date of issue: 2020-03-19

Tested by: Frank Wu

Checked by: B. Li Wang

Approved by: Kevin



1.0 Description

1.1 EN ISO 374-1 Marking of gloves type

EN ISO 374-1 / Type B



J K P T

1.2 EN ISO 374-5 Marking of gloves protecting against viruses, bacteria and fungi

EN ISO 374-5:2016



VIRUS

2.0 Test Results

2.1 EN ISO 374-1:2016 Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks

Method	Description	Result	Class
EN 374-2:2019 Protective gloves against dangerous chemicals and microorganisms. Determination of resistance to penetration	Penetration	PASS	/
EN 16523-1:2015+A1:2018 Determination of material resistance to permeation by chemicals. Permeation by potentially hazardous liquid chemicals under conditions of continuous contact	Permeation J	PASS	Level 6
	Permeation K	PASS	Level 6
	Permeation P	PASS	Level 2
	Permeation T	PASS	Level 6
EN ISO 374-4:2019 Protective gloves against dangerous chemicals and micro-organisms. Determination of resistance to degradation by chemicals	Degradation J	PASS	/
	Degradation K	PASS	/
	Degradation P	PASS	/
	Degradation T	PASS	/
Type of glove		PASS	Type B

2.2 EN ISO 374-5:2016 Protective gloves against dangerous chemicals and micro-organisms - Part 5: Terminology and performance requirements for micro-organisms risk

Method	Description	Result	Class
ISO 16604:2004 Clothing for protection against contact with blood and body fluids. Determination of resistance of protective clothing materials to penetration by blood-borne pathogens. Test method using Phi-X174 Bacteriophage. Proc. B	Protection against bacteria and fungi	PASS	/
EN 374-2:2019 Protective gloves against dangerous chemicals and microorganisms. Determination of resistance to penetration	Penetration	PASS	

3.0 Photographs of samples



Fig. appearance

******* End of Test Report *******