

EC Declaration of Conformity

Manufacturer: *DeVilbiss Healthcare LLC
100 DeVilbiss Drive
Somerset, PA 15501, USA*

EC Authorized Representative: *DeVilbiss Healthcare GmbH
Kamenzerstraße 3, 68309
Mannheim, Germany*

1. Oxygen Concentrators (UMDNS 12-873)

Description: *iGo2 Portable Oxygen Concentrator*
Catalogue nos.: *model # 125K, 125K-XB*
Classification (MDD Annex IX): *Ila (Rule 11)*
Conformity Assessment Procedure: *MDD 93/42/EEC, Annex II excluding Section 4*

2. Accessories:

Product Description (Catalogue no.):

<i>Spare Battery</i>	<i>125D-613</i>
<i>Power Supply 120 Watt (AC/DC Adapter)</i>	<i>DV68-620</i>
<i>Carry Case</i>	<i>125D-670</i>
<i>AC Power Cord, USA</i>	<i>DV51D-606</i>
<i>AC Power Cord, EU</i>	<i>DV51D-607</i>
<i>AC Power Cord, UK</i>	<i>DV51D-608</i>
<i>AC Power Cord, Australia</i>	<i>DV51D-609</i>
<i>AC Power Cord, China</i>	<i>DV51D-614</i>
<i>DC Power Cord (Auto Adapter)</i>	<i>DV6X-619</i>
<i>External Battery Charger, USA</i>	<i>125CH-613</i>
<i>External Battery Charger, Continental Europe</i>	<i>125CH-614</i>
<i>External Battery Charger, UK</i>	<i>125CH-615</i>
<i>Sieve Bed Package</i>	<i>125D-619</i>
<i>Cabinet Screws, 6 pk.</i>	<i>125D-621</i>
<i>O2 Port & Bacteria Filter</i>	<i>125D-610</i>

Applied standards: All applicable harmonized standards as adopted under the EC Directive 93/42/EEC and published in the Official Journal of the European Communities (See attached listing).

This Declaration of Conformity is based on the EC Directive 93/42/EEC, Annex II.3 and supported by the TÜV NORD CERT GmbH (0044) certificate, with reference to articles 1 and 3 of the directive 93/42/EEC.

Notified Body: *TÜV NORD CERT GmbH
Lungemarckstrasse 20, 45141 Essen, Germany*
Identification No.: *0044*
EC Certificate No.: *44 232 117803*
Start of EC Marking: *2019-08-07*

We, DeVilbiss Healthcare LLC., hereby declare under our sole responsibility that the above-mentioned products comply with the essential requirements and provisions of Council Directive 93/42/EEC.

EC Declaration of Conformity

Validity of this Declaration:

2019-08-07 – 2024-05-26

Somerset, PA, _____

11/12/2021

Place

Date



Tejash Naik Senior Director, Regulatory Affairs and Quality Assurance
Name and Position

Applied Standards:

125K
AAMI / ANSI ES60601-1:2005/(R) 2012 Ed 3.1 and A1:2012, C1:2009/(R) 2012 and A2:2010/(R) 2012 (Consolidated Text) Medical Electrical Equipment–Part 1: General Requirements for Basic Safety and Essential Performance (IEC 60601-1:2005, Mod).
AAMI / ANSI / IEC 60601–1–2:2014, Ed. 4.0, Medical Electrical Equipment—Part 1-2: General Requirements for Basic Safety And Essential Performance – Collateral Standard; Electromagnetic Disturbances - Requirements and Tests (associated with IEC 60601-1 Ed. 3.0)
IEC 60601-1-6:2010 + AMD 1:2013 Ed. 3.1 Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability (associated with IEC 60601-1 Ed. 3.0)
IEC 60601-1-8 Ed. 2.1 b.2012 Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems (associated with IEC 60601-1 3rd Edition)
IEC 60601-1-11:2010 Medical electrical equipment - Part 1-11: General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment (associated with IEC 60601-1 3rd Edition)
AAMI / ANSI / IEC 62304:2006 + AMD 1:2015, Medical Device Software – Software Life Cycle Process (Software/Informatics)
IEC 62366:2007 Ed. 1.0 + AMD 1:2014 – Medical devices - Application of usability engineering to medical devices
ISO 80601-2-69:2014 Medical electrical equipment – Part 2-69: Particular requirements for basic safety and essential performance of oxygen concentrator equipment
ISO 80601-2-67:2014 Medical electrical equipment – Part2-67: Particular requirements for basic safety and essential performance of oxygen conserving equipment
ISTA Procedure 3A: Packaged-Products for Parcel Delivery System Shipments 70 kg (150 lb) or Less (standard)
RTCA/DO-160G section 21, category M - Environmental Conditions and Test Procedures for Airborne Equipment
RTCA/DO-160G section 20, Radio Frequency Susceptibility – Radiated only 100 MHz up to 18 GHZ